



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Johany Z Lawson</u>		
Location: <u>Bowman County Sheriff</u>		
Intoxilyzer® Serial Number: <u>80-004188</u>		
Gas Lot Number: <u>23411080A1</u>	Gas Cylinder Number: <u>69</u>	Expiration Date: <u>10/01/2013</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>050</u> %		
Test 2: 0. <u>050</u> %		
Test 3: 0. <u>079</u> %		
(ATTACH TEST RECORD)		

[Signature]
Chemical Test Operator Signature

1-15-2013
Date

Deb Keshner
Reviewed By

22 Jan 2013
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004188
Location = BOWM 8164.13.00 06/09
01/15/2013 16:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:41
02 Std. Gas	0.080	16:42
03 Room Air	0.000	16:42
04 Std. Gas	0.080	16:42
05 Room Air	0.000	16:43
06 Std. Gas	0.079	16:43
07 Room Air	0.000	16:44

Lot No = 23411080A1

Cyl No = 69

Exp Date = 10/01/2013

County = 06

Oper No. = 020814



Operator Signature
JOHNNY LAWSON

Remarks:

Form 106-I8000