

**INTOXILYZER® 8000 INITIAL INSPECTION**

Intoxilyzer® 8000 Serial Number: 80-006490

**1) GENERAL Setup and Checks:**

- a)  Power On
  - 5 min. countdown during initial warm-up
  - 1 min. countdown to begin test after standby mode
- b)  Display Intensity (Ctrl + ↑ or ↓)
- c) Password Security:
  - Level 1 (Operator)     Level 2 (Field Inspector)     Level 3 (Lab Personnel)
- d)  Printer Setup (Level 3,S,P)
- e)  Forms Load (CMI, Inc., CoBrA Software)

**2) LEVEL 1 Setup and Checks:**

- a)  Internal Printer Test and Paper Feed (Level 1,P). Attach record.
- b)  External Printer Test (Level 1,P). Attach record.
- c) Barcode Reader (Level 1,B):
  - Operator ID     ND License     R key: Refusal. Attach record.

**3) LEVEL 2 Setup and Checks:**

- a)  Continuous Air Blank (Level 2,A)
- b)  Diagnostic (Level 2,D). Attach record.
- c)  Prelim Data Entry (Level 2,E)
- d) Ser Number/Version Info (Level 2,V):
  - Inst Ser#: 80-006490    DSP Ver: 1362.14
  - Core Ver: 8164.13    CPLD Ver: 7

**4) LEVEL 3 Setup and Checks:**

- a) DVM Monitor (Level 3,D,D): 3: 12538    s: 1    C: 47.00  
9: 13481    s: 1    B: 45 0  
(RFI indicator with keyed radio): 1
- b) Function (F) Keys :  F5 – Tone                       F9 - Fan  
 F6 – Green LED                   F10 – Pump  
 F7 – Red LED                       F11 – Solenoid
- c) Atmospheric Monitor (Level 3,D,A): Atmo Sensor, P: 949    CF: 1.06
- d) Flow Monitor (Level 3,D,F): Flow Sensor, R: 197    F: 0.000 L/s

- e) Temperature Monitor (Level 3,D,T): C: 47.0 A: 48.0 I: 0.0  
B: 45.0 V: 0.00
- f) Tank Monitor (Level 3,D,G): P (display): 713 psi P (gauge): ~725 psi
- g)  General Setup (Level 3,S,G)
- h)  Configure Start Test (Level 3,S,T)
- i)  Network (Level 3,S,N)

**5) TESTS (attach all test records):**

**Configure wet bath simulator standards before running tests (Level 1,S):**

- a) Zero AC (Level 1,C):  
Sim ser #: MP3067 Lot #: MILLIQ H<sub>2</sub>O Exp date: NA  
Known value: 0.000  
Results: 0.000 AC 0.000 AC 0.000 AC
- b) Low AC (Level 1,C):  
Sim ser #: MP3064 Lot #: 14050 Exp date: 02/10/2016  
Known value: 0.030  
Results: 0.030 AC 0.030 AC 0.031 AC
- c) High AC (Level 1,C):  
Sim ser #: MP3060 Lot #: 201409D Exp date: 9/9/16  
Known value: 0.250  
Results: 0.248 AC 0.249 AC 0.248 AC

**Configure dry gas standard before running tests (Level 1,S):**

- d) Ethanol + 0.05% Acetone (Level 1,B):  
Sim ser #: DR5135 Lot #: ICS 4 Exp date: 9/9/2017  
Known value: 0.10 AC + 0.05% ACETONE  
 Display reads "Interferent Detect"
- e) Acetone (Level 1,B):  
Sim ser #: DR5143 Lot #: ACE 2 Exp date: 9/29/2017  
Known value: 0.05% ACETONE  
 Display reads "Interferent Detect"
- f) Methanol (Level 1,B):  
Sim ser #: DR3377 Lot #: MET 2 Exp date: 9/29/2017  
Known value: 0.10% METHANOL  
 Display reads "Interferent Detect"
- g) Isopropanol (Level 1,B):  
Sim ser #: DR3453 Lot #: IPA 2 Exp date: 9/29/2017  
Known value: 0.10% ISOPROPRANOL  
 Display reads "Interferent Detect"

- h)  Ambient Fail (Level 1,B): Display and test record read correct.
- i)  Invalid Sample (Level 1,B): Display and test record read correct.
- j)  Sequence Aborted (Level 1,B): Display and test record read correct.
- k)  ABA (Level 1,B): Test record reads correct.
- l)  Calibration Check Out of Tolerance (CMS mode): Display and test record read correct.
- m)  RFI Detected (B,C, or CMS mode): Display and test record read correct.
- n)  Tank Pressure Below Minimum (CMS mode): Display reads correct.
- o)  Deficient Samples (CMS mode): Display and test record read correct.
- p)  Improper Sample (CMS mode): Display and test record read correct.
- q)  Difference Too Great (CMS mode): Display and test record read correct.
- r)  CMS (CMS mode): Test record reads correct.
- s) Stability Test (Level 3,T,S):

Ave: 0.0811 Std Dev: 0.0008 Rel Std Dev: 0.94%

Maintenance/Repairs/Notes: \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

Instrument is acceptable to be used in the field.

Charles E. Edr  
Field Inspector's Signature

1/15/16  
Date

Deb Shanauer  
Reviewed by

15 Jan 16  
Date

Intoxilyzer Test Record and Checklist  
 NDAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer  
 North Dakota Model 8000 SN 80-006490  
 Location = TOXL 8164.13.00 06/09  
 01/15/2016 11:59

\*\*\*\*\* Printer Test \*\*\*\*\*

12 char/line  
 abcdefghijkl  
 ABCDEFGHIJKL

16 char per line  
 abcdefghijklmnop  
 ABCDEFGHIJKLMNOP

21 chars per line  
 abcdefghijklmnopqrstu  
 ABCDEFGHIJKLMNOPQRSTU

24 characters per line  
 abcdefghijklmnopqrstuvw  
 ABCDEFGHIJKLMNOPQRSTUWX

32 characters per line  
 abcdefghijklmnopqrstuvwxyz123456  
 ABCDEFGHIJKLMNOPQRSTUVWXYZ123456

42 characters per line  
 abcdefghijklmnopqrstuvwxyz1234567890abcd  
 ABCDEFGHIJKLMNOPQRSTUVWXYZ1234567890abcd


Data Entry Mode: Enabled  
 Start Test Sequence: DABACABA  
 Display Prelim Rslt? Yes  
 Display Third Digit? Yes  
 Inhib Printer(Y/N)? No  
 Display Volume? No  
 Disable On Memfull? No  
 # of Print Copies? 1  
 Select Std (D/W/I)? Dry  
 Standard Value? 0.080  
 Standard Lot #? 19615080A5  
 Standard Cyl #? 3  
 Standard Expiration? 09/05/2017  
 Oper No? 666666


Flow Cal. Date: 08/18/2015  
 Slope 675  
 Intercept -635799

IR Calibration Date: 08/18/2015

	3um	9um
0th Coef(*100):	-18073	-15729
1st Coef(*100):	265337	133887
2nd Coef(*100):	3965	1560
H2O adj(ng/l*10K):	767	457

\*\*\*\*\* Test End \*\*\*\*\*

  
 Operator Signature  
 CHARLES EDER

Remarks:  
  
 Form 106-18000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer          Alcohol Analyzer  
North Dakota Model 8000          SN 80-006490  
Location = TOXL                  8164.13.00 06/09  
01/15/2016                                  12:01

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^\*()\_+?

Current Instrument Setup


Data Entry Mode:                  Enabled  
Start Test Sequence:              DABACABA  
Display Prelim Rslt?              Yes  
Display Third Digit?              Yes  
Inhib Printer(Y/N)?              No  
Display Volume?                  No  
Disable On Memfull?              No  
# of Print Copies?                1  
Select Std (D/W/I)?              Dry  
Standard Value?                  0.080  
Standard Lot #?                  19615080A5  
Standard Cyl #?                  3  
Standard Expiration?              09/05/2017  
Oper No?                          666666

Flow Cal. Date:                  08/18/2015  
Slope                              675  
Intercept                        -635799

IR Calibration Date:              08/18/2015  
    3um                              9um

-----  
0th Coef(\*100):                  -18073                          -15729  
1st Coef(\*100):                  265337                        133887  
2nd Coef(\*100):                  3965                            1560  
H2O adj(mg/l\*10k):              767                              457

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks:

PRINT TEST


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006490  
Location = TOXL      8164.13.00 06/09  
01/15/2016      12:03

Test	AC	Time
01 Room Air	0.000	12:03
02 *Subject Test	REF*	12:03
03 Room Air	0.000	12:04

\*Subject Test Refused

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = REFUSAL  
Dr. Lic. = ND/DIS821456  
Lot No = 19615080A5  
Cyl No = 3  
Expiration Date = 09/05/2017  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

REFUSAL

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006490  
Location = TOXL      8164.13.00 06/09  
01/15/2016      12:15

DIAGNOSTIC

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temp Regulation Test	Pass

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

DIAGNOSTIC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006490  
Location = TOXL      8164.13.00 06/09  
01/15/2016      12:29

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:30
02 Std. Sol.	0.000	12:30
03 Room Air	0.000	12:31
04 Std. Sol.	0.000	12:32
05 Room Air	0.000	12:32
06 Std. Sol.	0.000	12:33
07 Room Air	0.000	12:33

08 Sim Temp = 34.0°C

Simul Ser No = MP3067  
Std Sol No = MILLIQ H2O  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

ZERO AC  
0.000AC

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006490  
Location = TOXL      8164.13.00 06/09  
01/15/2016      12:36

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:36
02 Std. Sol.	0.030	12:37
03 Room Air	0.000	12:38
04 Std. Sol.	0.030	12:38
05 Room Air	0.000	12:39
06 Std. Sol.	0.031	12:40
07 Room Air	0.000	12:40

08 Sim Temp = 34.0°C

Simul Ser No = MP3064  
Std Sol No = 14050  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Low AC*  
*0.030 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006490  
Location = TOXL      8164.13.00 06/09  
01/15/2016      12:43

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:43
02 Std. Sol.	0.248	12:44
03 Room Air	0.000	12:45
04 Std. Sol.	0.249	12:46
05 Room Air	0.000	12:46
06 Std. Sol.	0.248	12:47
07 Room Air	0.000	12:48


08 Sim Temp = 34.0°C

Simul Ser No = MP3060

Std Sol No = 201409D

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

HIGH AC  
0.250 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006490  
Location = TOXL      8164.13.00 06/09  
01/15/2016      12:58

Test	AC	Time
01 Room Air	0.000	12:59
02 *Subject Test	INT*	13:00
03 Room Air	0.000	13:00

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = 0.10AC+0.05ACE  
Dr. Lic. = ND/DIS821456  
Lot No = 19615080A5  
Cyl No = 3  
Expiration Date = 09/05/2017  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*INTERFERENT DETECT*  
*0.10AC + 0.05% ACETONE*

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006490  
Location = TOXL      8164.13.00 06/09  
01/15/2016      13:09

Test	AC	Time
01 Room Air	0.000	13:10
02 *Subject Test	INT*	13:10
03 Room Air	0.000	13:11

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = 0.05 ACETONE  
Dr. Lic. = ND/DIS821456  
Lot No = 19615080A5  
Cyl No = 3  
Expiration Date = 09/05/2017  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*INTERFERENT DETECT*  
*CEE*  
*~~0.10% METHANOL~~ CEE*  
*0.05% ACETONE*  
Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006490  
Location = TOXL      8164.13.00 06/09  
01/15/2016      13:12

Test	AC	Time
01 Room Air	0.000	13:13
02 *Subject Test	INT*	13:13
03 Room Air	0.000	13:14

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = 0.10 METHANOL  
Dr. Lic. = ND/DIS821456  
Lot No = 19615080A5  
Cyl No = 3  
Expiration Date = 09/05/2017  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *INTERFERENT DETECT*  
*0.10% METHANOL*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006490  
Location = TOXL      8164.13.00 06/09  
01/15/2016      13:15

Test	AC	Time
01 Room Air	0.000	13:16
02 *Subject Test	INT*	13:16
03 Room Air	0.000	13:17

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = 0.10 IPA  
Dr. Lic. = ND/DIS821456  
Lot No = 19615080A5  
Cyl No = 3  
Expiration Date = 09/05/2017  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *INTERFERENT DETECT*  
*0.10% ISOPROPANOL*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006490  
Location = TOXL                    8164.13.00 06/09  
01/15/2016                            13:36

Test	AC	Time
01 Room Air	AMB*	13:37
02 Room Air	0.000	13:37

\*Invalid Test  
Check Ambient Conditions

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = AMBIENT FAIL  
Dr. Lic. = ND/DIS821456  
Lot No = 19615080A5  
Cyl No = 3  
Expiration Date = 09/05/2017  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *AMBIENT FAIL*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006490  
Location = TOXL      8164.13.00 06/09  
01/15/2016      13:46

Test	AC	Time
01 Room Air	0.000	13:47
02 Invalid Sample	X.XXX	13:48
03 Room Air	0.000	13:48

\*Invalid Test - Mouth Alcohol

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = INVALID SAMPLE  
Dr. Lic. = ND/DIS821456  
Lot No = 19615080A5  
Cyl No = 3  
Expiration Date = 09/05/2017  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *INVALID SAMPLE*




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006490  
Location = TOXL      8164.13.00 06/09  
01/15/2016      13:56

Test	AC	Time
01 Room Air	0.000	13:57
02 *Subject Test	ABT*	13:57
03 Room Air	0.000	13:57

\*Sequence Aborted

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = SEQUENCE ABORT  
Dr. Lic. = ND/DIS821456  
Lot No = 19615080A5  
Cyl No = 3  
Expiration Date = 09/05/2017  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: SEQUENCE ABORTED

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006490  
Location = TOXL      8164.13.00 06/09  
01/15/2016      14:17

Test	AC	Time
01 Room Air	0.000	14:18
02 Subject Test 1	0.064	14:19
03 Room Air	0.000	14:20
04 Reported AC	0.064	14:19

No RFI Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = ABA TEST  
Dr. Lic. = ND/DIS821456  
Lot No = 19615080A5  
Cyl No = 3  
Expiration Date = 09/05/2017  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *ABA TEST*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006490  
Location = TOXL                    8164.13.00 06/09  
01/15/2016                            13:58

Test	AC	Time
01 Diagnostic	OK	13:59
02 Room Air	0.000	14:00
03 Subject Test 1	0.054	14:00
04 Room Air	0.000	14:02
05 Std. Gas	0.088*	14:03
06 Room Air	0.000	14:04

\*Cal Check Out of Tolerance

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = CAL OUT TOL  
Dr. Lic. = ND/DIS821456  
Lot No = 19615080A5  
Cyl No = 3  
Expiration Date = 09/05/2017  
County = 08                            Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
OUT OF TOLERANCE

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006490  
Location = TOXL      8164.13.00 06/09  
01/15/2016      14:05

Test	AC	Time
01 Diagnostic	OK	14:06
02 Room Air	0.000	14:06
03 Subject Test 1	0.007	14:07
04 Room Air	0.000	14:09
05 Std. Gas	0.081	14:11
06 Room Air	0.000	14:12
07 *Subject Test	RFI*	14:12
08 Room Air	0.000	14:13

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = RFI CHECK  
Dr. Lic. = ND/DIS821456  
Lot No = 19615080A5  
Cyl No = 3  
Expiration Date = 09/05/2017  
County = 08      Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006490  
Location = TOXL                    8164.13.00 06/09  
01/15/2016                            14:40

Test	AC	Time
01 Diagnostic	OK	14:41
02 Room Air	0.000	14:41
03 *Subject Test	0.000*	14:44
04 Room Air	0.000	14:46
05 Std. Gas	0.080	14:48
06 Room Air	0.000	14:49
07 *Subject Test	0.000*	14:52
08 Room Air	0.000	14:54

No RFI Detected

\*Deficient Sample - Value Printed was  
Highest Obtained

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = DEFICIENT  
Dr. Lic. = ND/DIS821456  
Lot No = 19615080A5  
Cyl No = 3  
Expiration Date = 09/05/2017  
County = 08                            Oper No. = 666666

20 minute waiting period ascertained? Y  
Clean Mouthpiece used and disposed? Y  
Clean Mouthpiece used and disposed? Y

I followed the Approved Method and the  
instructions displayed by the Intoxilyzer  
in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: DEFICIENT SAMPLES

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006490  
Location = TOXL      8164.13.00 06/09  
01/15/2016      14:55

Test	AC	Time
01 Diagnostic	OK	14:55
02 Room Air	IPS*	14:56
03 Room Air	0.000	14:56

\*Improper Sample

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = IMPROPER SAMPLE  
Dr. Lic. = ND/DIS821456  
Lot No = 19615080A5  
Cyl No = 3  
Expiration Date = 09/05/2017  
County = 08      Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: *IMPROPER SAMPLE*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006490  
Location = TOXL                    8164.13.00 06/09  
01/15/2016                            15:06

Test	AC	Time
01 Diagnostic	OK	15:06
02 Room Air	0.000	15:07
03 Subject Test 1	0.079*	15:08
04 Room Air	0.000	15:10
05 Std. Gas	0.081	15:12
06 Room Air	0.000	15:13
07 Subject Test 2	0.141*	15:14
08 Room Air	0.000	15:15

\*Difference Too Great

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = DIFF TOO GREAT  
Dr. Lic. = ND/DIS821456  
Lot No = 19615080A5  
Cyl No = 3  
Expiration Date = 09/05/2017  
County = 08                            Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: *DIFFERENCE TOO GREAT*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006490  
Location = TOXL      8164.13.00 06/09  
01/15/2016      15:16

Test	AC	Time
01 Diagnostic	OK	15:17
02 Room Air	0.000	15:17
03 Subject Test 1	0.141	15:18
04 Room Air	0.000	15:20
05 Std. Gas	0.080	15:21
06 Room Air	0.000	15:23
07 Subject Test 2	0.141	15:23
08 Room Air	0.000	15:24
09 Reported AC	0.141	15:23

Difference OK  
No RFI Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = CMS TEST  
Dr. Lic. = ND/DIS821456  
Lot No = 19615080A5  
Cyl No = 3  
Expiration Date = 09/05/2017  
County = 08      Oper No. = 666666

20 minute waiting period ascertained? Y  
Clean Mouthpiece used and disposed? Y  
Clean Mouthpiece used and disposed? Y

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:

CMS TEST



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006490  
Location = TOXL      8164.13.00 06/09  
01/15/2016      15:26

Test	AC	Time
01 Room Air	0.000	15:26
02 Std. Gas	0.080	15:27
03 Room Air	0.000	15:27
04 Std. Gas	0.080	15:28
05 Room Air	0.000	15:28
06 Std. Gas	0.081	15:28
07 Room Air	0.000	15:29
08 Std. Gas	0.080	15:29
09 Room Air	0.000	15:30
10 Std. Gas	0.081	15:30
11 Room Air	0.000	15:31
12 Std. Gas	0.080	15:31
13 Room Air	0.000	15:32
14 Std. Gas	0.080	15:32
15 Room Air	0.000	15:33
16 Std. Gas	0.081	15:33
17 Room Air	0.000	15:34
18 Std. Gas	0.082	15:34
19 Room Air	0.000	15:34
20 Std. Gas	0.081	15:35
21 Room Air	0.000	15:35
22 Std. Gas	0.081	15:36
23 Room Air	0.000	15:36
24 Std. Gas	0.081	15:37
25 Room Air	0.000	15:37
26 Std. Gas	0.082	15:37
27 Room Air	0.000	15:38
28 Std. Gas	0.081	15:38
29 Room Air	0.000	15:39
30 Std. Gas	0.081	15:39
31 Room Air	0.000	15:40
32 Std. Gas	0.081	15:40
33 Room Air	0.000	15:41
34 Std. Gas	0.082	15:41
35 Room Air	0.000	15:42
36 Std. Gas	0.082	15:42
37 Room Air	0.000	15:43
38 Std. Gas	0.082	15:43
39 Room Air	0.000	15:43
40 Std. Gas	0.082	15:44
41 Room Air	0.000	15:44
42 Cal Check Stats:		
43 Average	0.0811	
44 Std Dev	0.0008	
45 Rel Std Dev	0.94%	

STABILITY TEST  
*Charles E. Edr*

Operator Signature  
CHARLES EDER

Remarks: Lot: 19615080A5  
CYL: 003 Exp. DATE 9/5/2017

Form 106-I8000

STABILITY TEST

*Charles Eder*  
1/15/16