

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program

Form: BrF-105-G

ETHANOL BREATH STANDARD CYLINDER REPORT

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Shannon Liebold

Location: Mercer County Jail

Intoxilyzer Serial No.: 80-003060

Lot Number: 659357

Cylinder Number: 43

Expiration Date: 09/03/2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.081%

Test 2: 0.081%

Test 3: 0.081%

(ATTACH TEST RECORD)

Shannon Liebold
Chemical Test Operator Signature

1-17-2010
Date

Deb Keshue
Reviewed By

11 Mar 2010
Date

80-003060
7/7/10 JAM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003060
Location = STAT 8164.13.00 06/09
01/17/2010 04:45

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	04:46
02 Std. Gas	0.081	04:46
03 Room Air	0.000	04:47
04 Std. Gas	0.081	04:47
05 Room Air	0.000	04:48
06 Std. Gas	0.081	04:48
07 Room Air	0.000	04:49

Lot No = 659357
Cyl No = 43
Exp Date = 09/03/2011
County = 29 Oper No. = 020904



Operator Signature
SHANNON P. LIEBOLD

Remarks:

ACA test

Form 106-I8000

80-003060
75/75 KM