

INTOX
PBT
TOXITEST

DATE	1-19-10
RECEIVED VIA	FEDEX

CMI INC.

316 E. 9th St., Owensboro, KY 42303
270-685-6200 — Fax 270-685-6288
1-800-835-0690

SERVICE WORK ORDER

3227778

B
I
L
L
T
O

ATTN:

CITY	STATE	ZIP
CUSTOMER NO.	585013	
PHONE		

S
H
I
P
T
O

ATTN:	Marlene Pearson		
	ND Crime Lab.		
	2641 East Main Ave		
CITY	STATE	ZIP	
Bismarck	ND	58501	
PHONE			

MODEL #	S/N	DATE RCVD.	P.O. #	REQ.	BILL CODE	SERVICE LOCATION CODE
8000	80-004203				3	

PROD. CODE	EXTRA PARTS RCVD.
	Battery

DESCRIPTION OF PROBLEM	<input type="checkbox"/> QUOTE	WARRANTY
"Would not let coin out of Not Ready"		<input type="checkbox"/> SERVICE DEPT. <input type="checkbox"/> NEW

- PRELIMINARY TEST DISMANTLE, INSPECT, DIAGNOSTIC TEST
- POWER SUPPLY TEST, INSPECT REPLACE
- MAIN PWB's - REMOVE & CLEAN, REINSTALL, TEST
- PWB's DIAGNOSE, REPAIR REPLACE
- PWB's - SURFACE CLEAN IN PLACE
- FILTER WHEEL - REMOVE, CLEAN, REINSTALL, CHECK CELL INTERIOR
- CHOPPER WHEEL MOTOR TEST & INSPECT REPLACE
- CELL - REMOVE, CLEAN, REINSTALL, REPLACE LENSES/GASKETS REQ'D
- IR LIGHT SOURCE REMOVE, REALIGN REPLACE
- PRINTER ASSY. DISASSEMBLE CLEAN & LUBRICATE
- REMOVE & REPLACE BELT REALIGNMENT
- REPLACE TEST PRINTER FUNCTIONS
- BATTERIES - REPLACE CPU RAM
- REPLACE PARTS REASSEMBLE CALIBRATION FINAL TEST
- MISC.
- PBT PRELIMINARY TESTING - PRESSURE, BATTERY, ZERO SET
- DISASSEMBLE REPLACE PARTS REASSEMBLE
- TEST CALIBRATE FINAL TEST
- TOXITEST INCOMING LEAK TEST VISUAL/MECH. INSPECTION
- TEST THERMOMETER & THERMOSTAT DISASSEMBLE
- REPLACE PARTS REASSEMBLE OUTGOING & LEAK TEST
- FINAL TEST
- MISC.

PARTS USED

QTY	ID NO.	DESCRIPTION	CODE

COMMENTS:
The instrument did not display neither of the problems listed. I did open the instrument, did an inspection, and found nothing.
<i>DN</i>

HOURS:	2
TECH.	Michael Wood
DATE	1/25/10
CAL. TECH.	
DATE	

SHIPPING DATE/CHARGES
80-004203
P. 68-69
YCN
SPECIAL SHIPPING INSTRUCTIONS

Certificate of Calibration

Intoxilyzer® 8000

*This is to certify that the calibration of **Intoxilyzer**® serial number 80-00 4203, manufactured by CMI Inc., a subsidiary of MPD, Inc. of Owensboro, Kentucky, was tested and found to be in compliance with the National Highway Traffic Safety Administration Standard for Devices to Measure Breath Alcohol (F.R., No. 179 48705-48710 Sept. 17, 1993) and with International Recommendation OIML R126 Edition 1998 (E). Calibration solutions are traceable to NIST standard material 1828.*

Date 1/28/10 Signed Kristi Robinson
Technician

EMI INC.

80-004203
P. 69-69 *sm*

316 East 9th Street
Owensboro, KY 42303
USA

Part No. 650519 CMI 7/9/09