

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program

Form: BrF-105-G

**ETHANOL BREATH STANDARD CYLINDER REPORT**

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Hinton, Rick W

Location: Cass County Jail

Intoxilyzer Serial No.: 80-004205 DK  
80-004025

Lot Number: 659357

Cylinder Number: 40

Expiration Date: 9-3-11


Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 83 %

Test 2: 0. 82 %

Test 3: 0. 83 %

(ATTACH TEST RECORD)

  
Chemical Test Operator Signature.

1-19-2010  
Date

Deb Kashner  
Reviewed By

11 Mar 2010  
Date

80-004205  
P. 142  
YM


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004205  
Location = CASS      8164.13.00 06/09  
01/19/2010      04:14

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	04:16
02 Std. Gas	0.083	04:16
03 Room Air	0.000	04:17
04 Std. Gas	0.082	04:17
05 Room Air	0.000	04:18
06 Std. Gas	0.083	04:18
07 Room Air	0.000	04:18

Lot No = 659357  
Cyl No = 40  
Exp Date = 09/03/2011  
County = 09      Oper No. = 029920



Operator Signature  
RICK HINTON

Remarks:

Form 106-I8000

80-004205  
P. 2 of 2  
RH