



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Chad Harmon</u>		
Location: <u>FYPD</u>		
Intoxilyzer® Serial Number: <u>80-004954</u>		
Gas Lot Number: <u>154ND80.A1</u>	Gas Cylinder Number: <u>34</u>	Expiration Date: <u>07/01/12</u>
<input checked="" type="checkbox"/> New Cylinder Installed <input type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.079</u> %		
Test 2: <u>0.079</u> %		
Test 3: <u>0.080</u> %		
(ATTACH TEST RECORD)		

[Signature]
Chemical Test Operator Signature

01-20-11
Date

Deb Kashur
Reviewed By

28 Jan 2011
Date

80-004954
P. G. S. G. L. L. E.
FNU

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = FYPD 8164.13.00 06/09
01/20/2011 22:23

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	22:24
02 Std. Gas	0.079	22:25
03 Room Air	0.000	22:25
04 Std. Gas	0.079	22:25
05 Room Air	0.000	22:26
06 Std. Gas	0.080	22:26
07 Room Air	0.000	22:27


Lot No = 15410080A1

Cyl No = 34

Exp Date = 07/01/2012

County = 61

Oper No. = 010606



Operator Signature
CHAD L. HARMON

Remarks:

Form 106-I8000

80-004954
P. Coleaf Lole
KM