

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program

Form: BrF-105-G

### ETHANOL BREATH STANDARD CYLINDER REPORT

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Darcy W. Aberle

Location: Williston LEC

Intoxilyzer Serial No.: 80-003061

Lot Number: 659357

Cylinder Number: 45

Expiration Date: 09/03/2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.078 %

Test 2: 0.078 %

Test 3: 0.078 %

(ATTACH TEST RECORD)

Darcy Aberle  
Chemical Test Operator Signature

01/21/2010  
Date

Deb Kashner  
Reviewed By

23 Mar 2010  
Date

80-003061  
P.14/2 PM

PJ/SS

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003061  
Location = WILL      8164.13.00 06/09  
01/21/2010      13:10

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:11
02 Std. Gas	0.078	13:11
03 Room Air	0.000	13:12
04 Std. Gas	0.078	13:12
05 Room Air	0.000	13:13
06 Std. Gas	0.078	13:13
07 Room Air	0.000	13:14

Lot No = 659357  
Cyl No = 45  
Exp Date = 09/03/2011  
County = 53      Oper No. = 010101

*Darcy W Aberle*

Operator Signature  
DARCY W. ABERLE

Remarks:

Test 7

Form 106-I8000

80-003061  
P. 2 of 2 KM

88/88