

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004199
Location = ROLL 8164.13.00 06/09
01/21/2010 09:39

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:40
02 Std. Gas	0.082	09:40
03 Room Air	0.000	09:41
04 Std. Gas	0.082	09:41
05 Room Air	0.000	09:41
06 Std. Gas	0.082	09:42
07 Room Air	0.000	09:42

Lot No = 659358
Cyl No = 11
Exp Date = 09/03/2011
County = 40 Oper No. = 049904



Operator Signature
STEVEN R. FISCHER

Remarks:

Form 106-I8000

80-004199
P. 74 of 74
YH