

**INTOXILYZER® 8000 INITIAL INSPECTION**

Intoxilyzer® 8000 Serial Number: 80-006492

**A. GENERAL Setup and Checks:**

- Power On  
 5 min. countdown during initial warm-up  
 1 min. countdown to begin test after standby mode
- Display Intensity (Ctrl + ↑ or ↓)
- Password Security:  
 Level 1 (Operator)  Level 2 (Field Inspector)  Level 3 (Lab Personnel)
- Printer Setup (Level 3,S,P)
- Forms Load (CMI, Inc., CoBrA Software)

**B. LEVEL 1 Setup and Checks:**

- Internal Printer Test and Paper Feed (Level 1,P). Attach record.
- External Printer Test (Level 1,P). Attach record.
- Barcode Reader (Level 1,B):  
 Operator ID  ND License  R key: Refusal. Attach record.

**C. LEVEL 2 Setup and Checks:**

- Continuous Air Blank (Level 2,A)
- Diagnostic (Level 2,D). Attach record.
- Prelim Data Entry (Level 2,E)
- Ser Number/Version Info (Level 2,V):  
Inst Ser#: 80-006492 DSP Ver: 1362.14  
Core Ver: 8164.13 CPLD Ver: 7

**D. LEVEL 3 Setup and Checks:**

- DVM Monitor (Level 3,D,D): 3: 12625 s: 1 C: 47.06  
9: 13508 s: 0 B: 45 0  
(RFI indicator with keyed radio): 1
- Function (F) Keys: :  F5 - Tone  F9 - Fan  
 F6 - Green LED  F10 - Pump  
 F7 - Red LED  F11 - Solenoid
- Atmospheric Monitor (Level 3,D,A): Atmo Sensor, P: 966 CF: 1.04
- Flow Monitor (Level 3,D,F): Flow Sensor, R: 187 F: 0.000 L/s

5. Temperature Monitor (Level 3,D,T): C: 47.0 A: 49.0 I: 0.0  
B: 45.0 V: 0.0
6. Tank Monitor (Level 3,D,G): P (display): 648 psi P (gauge): ≈ 650 psi
7.  General Setup (Level 3,S,G)
8.  Configure Start Test (Level 3,S,T)
9.  Network (Level 3,S,N)

**E. TESTS (attach all test records):**

**Configure wet bath simulator standards before running tests (Level 1,S):**

1. Zero AC (Level 1,C):  
Sim ser #: DR 3375 Lot #: N/A Milli-Q Exp date: N/A  
Known value: 0.000 AC  
Results: 0.000 AC 0.000 AC 0.000 AC
2. Low AC (Level 1,C):  
Sim ser #: DR 3374 Lot #: 14240 Exp date: 10/20/16  
Known value: 0.020 AC  
Results: 0.019 AC 0.019 AC 0.019 AC
3. High AC (Level 1,C):  
Sim ser #: DR 7111 Lot #: 201408C Exp date: 08/12/16  
Known value: 0.300 AC  
Results: 0.299 AC 0.300 AC 0.300 AC

**Configure dry gas standard before running tests (Level 1,S):**

4. Ethanol + 0.05% Acetone (Level 1,B):  
Sim ser #: DR 5144 Lot #: ICS4 Exp date: 09/09/15<sup>DS</sup> 17  
Known value: 0.10 AC + 0.05% acetone  
 Display reads "Interferent Detect"
5. Acetone (Level 1,B):  
Sim ser #: DR 3455 Lot #: Acet Exp date: 09/09/17  
Known value: 0.05% acetone  
 Display reads "Interferent Detect"
6. Methanol (Level 1,B):  
Sim ser #: DR 5189 Lot #: Met 2 Exp date: 09/09/17  
Known value: 0.10% methanol  
 Display reads "Interferent Detect"
7. Isopropanol (Level 1,B):  
Sim ser #: DR 3454 Lot #: Ipad 2 Exp date: 09/09/2017  
Known value: 0.10% Isopropanol  
 Display reads "Interferent Detect"

- 8.  Ambient Fail (Level 1,B): Display and test record read correct.
- 9.  Invalid Sample (Level 1,B): Display and test record read correct.
- 10.  Sequence Aborted (Level 1,B): Display and test record read correct.
- 11.  ABA (Level 1,B): Test record reads correct.
- 12.  Calibration Check Out of Tolerance (CMS mode): Display and test record read correct.
- 13.  RFI Detected (B,C, or CMS mode): Display and test record read correct.
- 14.  Tank Pressure Below Minimum (CMS mode): Display reads correct.
- 15.  Deficient Samples (CMS mode): Display and test record read correct.
- 16.  Improper Sample (CMS mode): Display and test record read correct.
- 17.  Difference Too Great (CMS mode): Display and test record read correct.
- 18.  CMS (CMS mode): Test record reads correct.
- 19. Stability Test (Level 3,T,S):

Ave: 0.0811 Std Dev: 0.0005 Rel Std Dev: 0.63%

Maintenance/Repairs/Notes: Re-aligned internal printer  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

Instrument is acceptable to be used in the field.

Deb Shanaver  
Field Inspector's Signature

21 Jan 16  
Date

Charles E. Edr  
Reviewed by

21 JAN 16  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer  
North Dakota Model 8000 SN 80-006492  
Location = TOXL 8164.13.00 06/09  
01/20/2016 16:15

\*\*\*\*\* Printer Test \*\*\*\*\*

12 char/line  
abcdefghijkl  
ABCDEFGHIJKL

16 char per line  
abcdefghijklmnop  
ABCDEFGHIJKLMN

21 chars per line  
abcdefghijklmnopqrstu  
ABCDEFGHIJKLMNQRSTU

24 characters per line  
abcdefghijklmnopqrstuwx  
ABCDEFGHIJKLMNQRSTUWX

32 characters per line  
abcdefghijklmnopqrstuwxz123456  
ABCDEFGHIJKLMNQRSTUWXYZ123456

42 characters per line  
abcdefghijklmnopqrstuwxz1234567890abcdef  
ABCDEFGHIJKLMNQRSTUWXYZ1234567890ABCDEF

Current Instrument Setup

Data Entry Mode: Enabled  
Start Test Sequence: DABACABA  
Display Prelim Rslt? Yes  
Display Third Digit? Yes  
Inhib Printer(Y/N)? No  
Display Volume? No  
Disable On Memfull? No  
# of Print Copies? 1  
Select Std (D/W/I)? Dry  
Standard Value? 0.080  
Standard Lot #? 1234567  
Standard Cyl #? 1234567  
Standard Expiration? 11/11/2020  
Oper No? 777777

Flow Cal. Date: 08/18/2015  
Slope 669  
Intercept -631515

IR Calibration Date: 08/18/2015  
3um 9um

0th Coef(\*100): -41685 -23145

1st Coef(\*100): 269226 142194

2nd Coef(\*100): 2671 1239

H2O adj(mg/l\*10K): 872 437

\*\*\*\*\* Test End \*\*\*\*\*



Operator Signature  
DEB SHANAVER

Remarks:

*Print test*

Form 106-18000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006492  
Location = TOXL      8164.13.00 06/09  
01/21/2016      09:15

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

Current Instrument Setup

Data Entry Mode:      Enabled  
Start Test Sequence:      DABACABA  
Display Prelim Rslt?      Yes  
Display Third Digit?      Yes  
Inhib Printer(Y/N)?      No  
Display Volume?      No  
Disable On Memfull?      No  
# of Print Copies?      1  
Select Std (D/W/I)?      Dry  
Standard Value?      0.080  
Standard Lot #?      1234567  
Standard Cyl #?      1234567  
Standard Expiration?      11/11/2020  
Oper No?      777777

Flow Cal. Date:      08/18/2015  
Slope      669  
Intercept      -631515

IR Calibration Date:      08/18/2015

	3um	9um
0th Coef(*100):	-41685	-23145
1st Coef(*100):	269226	142194
2nd Coef(*100):	2671	1239
H2O adj(mg/l*10k):	872	437

\*\*\*\*\* Printer Test End \*\*\*\*\*

Deb Shanaver  
Operator Signature  
DEB SHANAVER

Remarks:

Print Test


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006492  
Location = TOXL      8164.13.00 06/09  
01/21/2016      09:19

Test	AC	Time
01 Room Air	0.000	09:20
02 *Subject Test	REF*	09:20
03 Room Air	0.000	09:21

\*Subject Test Refused

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 1234567  
Cyl No = 1234567  
Expiration Date = 11/11/2020  
County = 08      Oper No. = 777777



Operator Signature  
DEB SHANAVER

Remarks:

*Refusal*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006492  
Location = TOXL      8164.13.00 06/09  
01/21/2016      09:22

DIAGNOSTIC

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temp Regulation Test	Pass

Oper No. = 777777

  
\_\_\_\_\_

Operator Signature  
DEB SHANAVER

Remarks:

*Diagnostic test*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006492  
Location = TOXL      8164.13.00 06/09  
01/21/2016      09:53

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:54
02 Std. Sol.	0.000	09:54
03 Room Air	0.000	09:55
04 Std. Sol.	0.000	09:56
05 Room Air	0.000	09:56
06 Std. Sol.	0.000	09:57
07 Room Air	0.000	09:57


08 Sim Temp = 34.0°C

Simul Ser No = DR3375

Std Sol No = MILLIQ

County = 08

Oper No. = 777777

  
\_\_\_\_\_  
Operator Signature  
DEB SHANAVER

Remarks:

zero alcohol    0.000 AC



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006492  
Location = TOXL      8164.13.00 06/09  
01/21/2016      09:59

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:00
02 Std. Sol.	0.019	10:01
03 Room Air	0.000	10:01
04 Std. Sol.	0.019	10:02
05 Room Air	0.000	10:02
06 Std. Sol.	0.019	10:03
07 Room Air	0.000	10:04


08 Sim Temp = 34.0°C

Simul Ser No = DR3374

Std Sol No = 14240

County = 08

Oper No. = 777777

  
\_\_\_\_\_  
Operator Signature  
DEB SHANAVER

Remarks:

Low AC      0.020AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006492  
Location = TOXL      8164.13.00 06/09  
01/21/2016      10:06

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:07
02 Std. Sol.	0.299	10:08
03 Room Air	0.000	10:08
04 Std. Sol.	0.300	10:09
05 Room Air	0.000	10:10
06 Std. Sol.	0.300	10:10
07 Room Air	0.000	10:11

08 Sim Temp = 34.0°C

Simul Ser No = DR7111

Std Sol No = 201408C

County = 08

Oper No. = 777777

Deb Shanaver

Operator Signature  
DEB SHANAVER

Remarks:

High AC 0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006492  
Location = TOXL      8164.13.00 06/09  
01/21/2016      13:22

Test	AC	Time
01 Diagnostic	OK	13:23
02 Room Air	0.000	13:23
03 *Subject Test	INT*	13:23
04 Room Air	0.000	13:24

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 29614080A2  
Cyl No = 25  
Expiration Date = 12/05/2016  
County = 08      Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

*Deb Shanaver*

Operator Signature  
DEB SHANAVER

Remarks:

*Interferent detected 0.10AC + 0.05% acetone*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006492  
Location = TOXL                    8164.13.00 06/09  
01/21/2016                            13:27

Test	AC	Time
01 Room Air	0.000	13:28
02 *Subject Test	INT*	13:28
03 Room Air	0.000	13:29

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 29614080A2  
Cyl No = 25  
Expiration Date = 12/05/2016  
County = 08                          Oper No. = 777777

*Deb Shanaver*

Operator Signature  
DEB SHANAVER

Remarks:

*Interferent detected 0.05% acetone*

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006492  
Location = TOXL                    8164.13.00 06/09  
01/21/2016                            13:30

Test	AC	Time
01 Room Air	0.000	13:31
02 *Subject Test	INT*	13:31
03 Room Air	0.000	13:32

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 29614080A2  
Cyl No = 25  
Expiration Date = 12/05/2016  
County = 08                            Oper No. = 777777

  
\_\_\_\_\_  
Operator Signature  
DEB SHANAVER

Remarks:

*Interferent detected 0.10% Methanol*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006492  
Location = TOXL                8164.13.00 06/09  
01/21/2016                        13:36

Test	AC	Time
01 Room Air	0.000	13:36
02 *Subject Test	INT*	13:37
03 Room Air	0.000	13:38

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                Weight = 150  
Test = DUI                        Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 29614080A2  
Cyl No = 25  
Expiration Date = 12/05/2016  
County = 08                        Oper No. = 777777

Deb Shanaver  
Operator Signature  
DEB SHANAVER

Remarks:  
*Interferent detected 0.10% isopropanal*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006492  
Location = TOXL      8164.13.00 06/09  
01/21/2016      13:19


Test	AC	Time
01 Room Air	AMB*	13:20
02 Room Air	0.000	13:21

\*Invalid Test  
Check Ambient Conditions

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 29614080A2  
Cyl No = 25  
Expiration Date = 12/05/2016  
County = 08      Oper No. = 777777

  
Operator Signature  
DEB SHANAVER

Remarks:

  
Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006492  
Location = TOXL                 8164.13.00 06/09  
01/21/2016                        11:41

Test	AC	Time
01 Room Air	0.000	11:42
02 Invalid Sample	X.XXX	11:42
03 Room Air	0.000	11:43

\*Invalid Test - Mouth Alcohol

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                 Weight = 150  
Test = DUI                         Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 29614080A2  
Cyl No = 25  
Expiration Date = 12/05/2016  
County = 08                        Oper No. = 777777



Operator Signature  
DEB SHANAVER

Remarks:

*Invalid sample - mouth alcohol*

Form 106-I8000




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006492  
Location = TOXL                    8164.13.00 06/09  
01/21/2016                            11:39

Test	AC	Time
01 Room Air	0.000	11:39
02 *Subject Test	ABT*	11:40
03 Room Air	0.000	11:40

\*Sequence Aborted

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 29614080A2  
Cyl No = 25  
Expiration Date = 12/05/2016  
County = 08                            Oper No. = 777777

  
\_\_\_\_\_  
Operator Signature  
DEB SHANAVER

Remarks:

*Sequence aborted*  
Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006492  
Location = TOXL      8164.13.00 06/09  
01/21/2016      11:35

Test	AC	Time
01 Room Air	0.000	11:36
02 Subject Test 1	0.009	11:36
03 Room Air	0.000	11:36
04 Reported AC	0.009	11:36

No RFI Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 29614080A2  
Cyl No = 25  
Expiration Date = 12/05/2016  
County = 08      Oper No. = 777777

  
\_\_\_\_\_  
Operator Signature  
DEB SHANAVER

Remarks:

ABA test

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006492  
Location = TOXL                    8164.13.00 06/09  
01/21/2016                            09:38

Test	AC	Time
01 Diagnostic	OK	09:39
02 Room Air	0.000	09:39
03 Subject Test 1	0.000	09:39
04 Room Air	0.000	09:41
05 Std. Gas	0.099*	09:43
06 Room Air	0.000	09:43

\*Cal Check Out of Tolerance

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 29614080A2  
Cyl No = 25  
Expiration Date = 12/05/2016  
County = 08                          Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Shanaver  
Operator Signature  
DEB SHANAVER

Remarks:

*Calibration check out of tolerance*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006492  
Location = TOXL      8164.13.00 06/09  
01/21/2016      11:31

Test	AC	Time
01 Room Air	RFI*	11:32
02 Room Air	0.000	11:32

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 29614080A2  
Cyl No = 25  
Expiration Date = 12/05/2016  
County = 08      Oper No. = 777777

  
\_\_\_\_\_  
Operator Signature  
DEB SHANAVER

Remarks:

RFI

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006492  
Location = TOXL                    8164.13.00 06/09  
01/21/2016                            11:13

Test	AC	Time
01 Diagnostic	OK	11:14
02 Room Air	0.000	11:14
03 *Subject Test	0.000*	11:18
04 Room Air	0.000	11:20
05 Std. Gas	0.081	11:21
06 Room Air	0.000	11:22
07 *Subject Test	0.000*	11:25
08 Room Air	0.000	11:26


No RFI Detected

\*Deficient Sample - Value Printed was  
Highest Obtained

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 29614080A2  
Cyl No = 25  
Expiration Date = 12/05/2016  
County = 08                          Oper No. = 777777

20 minute waiting period ascertained? Y  
Clean Mouthpiece used and disposed? Y  
Clean Mouthpiece used and disposed? Y

I followed the Approved Method and the  
instructions displayed by the Intoxilyzer  
in conducting this test.



Operator Signature  
DEB SHANAVER

Remarks:

*Deficient samples*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006492  
Location = TOXL      8164.13.00 06/09  
01/21/2016      11:11

Test	AC	Time
01 Diagnostic	OK	11:12
02 Room Air	IPS*	11:12
03 Room Air	0.000	11:13

\*Improper Sample

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 29614080A2  
Cyl No = 25  
Expiration Date = 12/05/2016  
County = 08      Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

  
\_\_\_\_\_  
Operator Signature  
DEB SHANAVER

Remarks:

*Improper Sample*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006492  
Location = TOXL                    8164.13.00 06/09  
01/21/2016                            11:00

Test	AC	Time
01 Diagnostic	OK	11:00
02 Room Air	0.000	11:01
03 Subject Test 1	0.000*	11:01
04 Room Air	0.000	11:03
05 Std. Gas	0.081	11:04
06 Room Air	0.000	11:06
07 Subject Test 2	0.043*	11:06
08 Room Air	0.000	11:07

\*Difference Too Great

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 29614080A2  
Cyl No = 25  
Expiration Date = 12/05/2016  
County = 08                          Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

  
\_\_\_\_\_  
Operator Signature  
DEB SHANAVER

Remarks:

*Difference too great*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006492  
Location = TOXL                    8164.13.00 06/09  
01/21/2016                            10:50

Test	AC	Time
01 Diagnostic	OK	10:51
02 Room Air	0.000	10:52
03 Subject Test 1	0.000	10:52
04 Room Air	0.000	10:54
05 Std. Gas	0.081	10:55
06 Room Air	0.000	10:57
07 Subject Test 2	0.000	10:57
08 Room Air	0.000	10:58
09 Reported AC	0.000	10:52

Difference OK  
No RFI Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 29614080A2  
Cyl No = 25  
Expiration Date = 12/05/2016  
County = 08                          Oper No. = 777777

20 minute waiting period ascertained? Y  
Clean Mouthpiece used and disposed? Y  
Clean Mouthpiece used and disposed? Y

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Shanaver  
Operator Signature  
DEB SHANAVER

Remarks:

CMS



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006492  
Location = TOXL                    8164.13.00 06/09  
01/21/2016                            10:13

Test	AC	Time
01 Room Air	0.000	10:14
02 Std. Gas	0.081	10:14
03 Room Air	0.000	10:15
04 Std. Gas	0.081	10:15
05 Room Air	0.000	10:16
06 Std. Gas	0.081	10:16
07 Room Air	0.000	10:17
08 Std. Gas	0.081	10:17
09 Room Air	0.000	10:18
10 Std. Gas	0.081	10:18
11 Room Air	0.000	10:18
12 Std. Gas	0.080	10:19
13 Room Air	0.000	10:19
14 Std. Gas	0.081	10:20
15 Room Air	0.000	10:20
16 Std. Gas	0.081	10:21
17 Room Air	0.000	10:21
18 Std. Gas	0.080	10:22
19 Room Air	0.000	10:22
20 Std. Gas	0.081	10:22
21 Room Air	0.000	10:23
22 Std. Gas	0.081	10:23
23 Room Air	0.000	10:24
24 Std. Gas	0.081	10:24
25 Room Air	0.000	10:25
26 Std. Gas	0.081	10:25
27 Room Air	0.000	10:26
28 Std. Gas	0.081	10:26
29 Room Air	0.000	10:27
30 Std. Gas	0.081	10:27
31 Room Air	0.000	10:27
32 Std. Gas	0.082	10:28
33 Room Air	0.000	10:28
34 Std. Gas	0.081	10:29
35 Room Air	0.000	10:29
36 Std. Gas	0.081	10:30
37 Room Air	0.000	10:30
38 Std. Gas	0.082	10:30
39 Room Air	0.000	10:31
40 Std. Gas	0.082	10:31
41 Room Air	0.000	10:32
42 Cal Check Stats:		
43 Average	0.0811	
44 Std Dev	0.0005	
45 Rel Std Dev	0.63%	

*Stability test*  
*Deb Shanaver*

Operator Signature  
DEB SHANAVER

Remarks:

lot: 29614080A2 Cyl: 25 Exp: 12/05/16

Form 106-I8000

Stability Test  
Deb Shanaver