



ETHANOL BREATH STANDARD CYLINDER REPORT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Darcy W Aberle</u>		
Location: <u>Williams Co LEC</u>		
Intoxilyzer® Serial Number: <u>80-004189</u>		
Gas Lot Number: <u>23411080A1</u>	Gas Cylinder Number: <u>23</u>	Expiration Date: <u>10/01/2013</u>
<input checked="" type="checkbox"/> New Cylinder Installed	<input type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>080</u> %		
Test 2: 0. <u>080</u> %		
Test 3: 0. <u>080</u> %		
(ATTACH TEST RECORD)		

Darcy Aberle
Chemical Test Operator Signature

01-22-12
Date

Deb Kashner
Reviewed By

02 Aprilth 2012
Date


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = WILL 8164.13.00 06/09
01/22/2012 13:31

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:32
02 Std. Gas	0.080	13:32
03 Room Air	0.000	13:32
04 Std. Gas	0.080	13:33
05 Room Air	0.000	13:33
06 Std. Gas	0.080	13:34
07 Room Air	0.000	13:34

Lot No = 23411080A1
Cyl No = 23
Exp Date = 10/01/2013
County = 53 Oper No. = 010101



Operator Signature
DARCY W. ABERLE

Remarks: