



ETHANOL GAS STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <i>Scott Nordquist</i>		
Location <i>Emmons County S.O.</i>	Intoxilyzer® Serial Number <i>80-004946 DK</i>	
Gas Lot Number <i>16512080A2</i>	Gas Cylinder Number <i>16</i>	Gas Expiration Date <i>08/01/14</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
A. *0.082* AC
B. *0.082* AC
C. *0.082* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Scott Nordquist</i>	Date <i>01/23/14</i> ✓
Reviewed By (Crime Laboratory Use Only) <i>Deb Kashner</i>	Date <i>27 Jan 2014</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = LINT 8164.13.00 06/09
01/23/2014 14:22

DRY CAL CHECK


Test	AC	Time
01 Room Air	0.000	14:23
02 Std. Gas	0.082	14:24
03 Room Air	0.000	14:24
04 Std. Gas	0.082	14:24
05 Room Air	0.000	14:25
06 Std. Gas	0.082	14:25
07 Room Air	0.000	14:26

Lot No = 16512080A2
Cyl No = 016
Exp Date = 08/01/2014
County = 15 Oper No. = 130380


Operator Signature
SCOTT NORDQUIST

Remarks:

Form 106-I8000

I certify that this is a true and correct
copy of the Intoxilyzer test record.
Dated this 23 day of January,
2014.
Scott Nordquist 4274
Print name Badge #

Signature