



# ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (5-2011)

|  |  |  |
|--|--|--|
| Chemical Test Operator Name (Print)<br><i>Jeremy Grohs</i> |  |  |
| Location<br><i>Burke County Sheriffs Office</i>            | Intoxilyzer® Serial Number<br><i>80-004940</i> |  |
| Gas Lot Number<br><i>23411080A1</i>                        | Gas Cylinder Number<br><i>31</i>               | Gas Expiration Date<br><i>10/01/2013</i> |

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
  - A. *0.079* AC
  - B. *0.079* AC
  - C. *0.079* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Breath Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

|   |                            |
|---|----------------------------|
| Chemical Test Operator Signature<br><i>Jeremy Grohs</i>       | Date<br><i>1-25-2013</i>   |
| Reviewed By (Crime Laboratory Use Only)<br><i>Deb Kashner</i> | Date<br><i>31 Jan 2013</i> |

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004940  
Location = BURK                    8164.13.00 06/09  
01/25/2013                            17:34

DRY CAL CHECK

| Test        | AC    | Time  |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 17:35 |
| 02 Std. Gas | 0.079 | 17:35 |
| 03 Room Air | 0.000 | 17:35 |
| 04 Std. Gas | 0.079 | 17:36 |
| 05 Room Air | 0.000 | 17:36 |
| 06 Std. Gas | 0.079 | 17:37 |
| 07 Room Air | 0.000 | 17:37 |

Lot No = 23411080A1  
Cyl No = 31  
Exp Date = 10/01/2013  
County = 07                            Oper No. = 130230



Operator Signature  
JEREMY GROHS

Remarks:

GOOD TEST  
Calibration Check

Form 106-I8000