



# ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name:	Shannon Liebhold				
Location:	Mercer County Sheriff's Dept.				
Intoxilyzer® Serial Number:	<del>80-003061</del> 80-003060 DK				
Gas Lot Number:	659357	Gas Cylinder Number:	43	Expiration Date:	9-3-11
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed		-		
Instrument Test Results (Report to 3 Digits; Example, 0.081%)					
Test 1: 0.081 %					
Test 2: 0.080 %					
Test 3: 0.080 %					
(ATTACH TEST RECORD)					

Shannon Liebhold  
Chemical Test Operator Signature

1-26-11  
Date

Deb Kestner  
Reviewed By

28 Jan 2011  
Date

80-003060  
P. 98/99  
JK

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003060  
Location = STAT      8164.13.00 06/09  
01/26/2011      03:23

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	03:24
02 Std. Gas	0.081	03:25
03 Room Air	0.000	03:25
04 Std. Gas	0.080	03:25
05 Room Air	0.000	03:26
06 Std. Gas	0.080	03:26
07 Room Air	0.000	03:27

Lot No = 659357  
Cyl No = 43  
Exp Date = 09/03/2011  
County = 29      Oper No. = 020904



Operator Signature  
SHANNON P. LIEBOLD

Remarks:

Form 106-I8000

80-003060  
P. 99 of 99  
AM