

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program

Form: BrF-105-G

ETHANOL BREATH STANDARD CYLINDER REPORT

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Officer Pete Nielsen

Location: West Fargo PD

Intoxilyzer Serial No.: 80-003059

Lot Number: 659357

Cylinder Number: 10

Expiration Date: 9-3-2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 082 %

Test 2: 0. 081 %

Test 3: 0. 081 %

(ATTACH TEST RECORD)


Chemical Test Operator Signature

1-28-10
Date

Deb Keshner
Reviewed By

02 Feb 2010
Date

80-003059
P. 1 of 2


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003059
Location = WFPD 8164.13.00 06/09
01/28/2010 09:52

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:53
02 Std. Gas	0.082	09:53
03 Room Air	0.000	09:54
04 Std. Gas	0.081	09:54
05 Room Air	0.000	09:54
06 Std. Gas	0.081	09:55
07 Room Air	0.000	09:55

Lot No = 659357
Cyl No = 10
Exp Date = 09/03/2011
County = 09 Oper No. = 039617


Operator Signature
PETER NIELSEN

Remarks:

Form 106-I8000

80-003059
P-2/2/11