

CMI INC.

316 E. 9th St., Owensboro, KY 42303
 270-685-6200 — Fax 270-685-6288
 1-800-835-0690

SERVICE WORK ORDER

322777

PBT
 TOXITEST

DATE 1-19-10
RECEIVED VIA FEDEX

B
I
L
L
T
O

ATTN:		
CITY	STATE	ZIP
CUSTOMER NO.	3585013	
PHONE		

S
H
I
P
T
O

ATTN: I. Job Kashim
AD Crime Lab
2041 East Main Ave
CITY Bismarck, ND STATE ND ZIP 58501
PHONE

MODEL # 8000	S/N 80-004191	DATE RCVD.	P.O. #	REQ.	BILL CODE	SERVICE LOCATION CODE
-----------------	------------------	------------	--------	------	-----------	-----------------------

PROD. CODE	EXTRA PARTS RCVD. air, gas shelf
------------	-------------------------------------

DESCRIPTION OF PROBLEM	<input type="checkbox"/> QUOTE	WARRANTY <input type="checkbox"/> SERVICE DEPT. <input type="checkbox"/> NEW
------------------------	--------------------------------	---

- PRELIMINARY TEST DISMANTLE, INSPECT, DIAGNOSTIC TEST
- POWER SUPPLY TEST, INSPECT REPLACE
- MAIN PWB's - REMOVE & CLEAN, REINSTALL, TEST
- PWB's DIAGNOSE, REPAIR REPLACE
- PWB's - SURFACE CLEAN IN PLACE
- FILTER WHEEL - REMOVE, CLEAN, REINSTALL, CHECK CELL INTERIOR
- CHOPPER WHEEL MOTOR TEST & INSPECT REPLACE
- CELL - REMOVE, CLEAN, REINSTALL, REPLACE LENSES/GASKETS REQ'D
- IR LIGHT SOURCE REMOVE, REALIGN REPLACE
- PRINTER ASSY. DISASSEMBLE CLEAN & LUBRICATE
- REMOVE & REPLACE BELT REALIGNMENT
- REPLACE TEST PRINTER FUNCTIONS
- BATTERIES - REPLACE CPU RAM
- REPLACE PARTS REASSEMBLE CALIBRATION FINAL TEST
- MISC.
- PBT PRELIMINARY TESTING - PRESSURE, BATTERY, ZERO SET
- DISASSEMBLE REPLACE PARTS REASSEMBLE
- TEST CALIBRATE FINAL TEST
- TOXITEST INCOMING LEAK TEST VISUAL/MECH. INSPECTION
- TEST THERMOMETER & THERMOSTAT DISASSEMBLE
- REPLACE PARTS REASSEMBLE OUTGOING & LEAK TEST
- FINAL TEST
- MISC.

PARTS USED			
QTY	ID NO.	DESCRIPTION	CODE

COMMENTS:
 instrument was recalibrated. The unit passed calibration with no problems.

WJM

HOURS: 2	DATE 1/19/10	DATE 1/24/10
TECH. Michael	CAL. TECH. V	DATE

SHIPPING DATE/CHARGES	80-004191 P. 49 of 50 FK
-----------------------	--------------------------------

Certificate of Calibration

Intoxilyzer® 8000

*This is to certify that the calibration of **Intoxilyzer**® serial number 80-004191, manufactured by CMI Inc., a subsidiary of MPD, Inc. of Owensboro, Kentucky, was tested and found to be in compliance with the National Highway Traffic Safety Administration Standard for Devices to Measure Breath Alcohol (F.R., No. 179 48705-48710 Sept. 17, 1993) and with International Recommendation OIML R126 Edition 1998 (E). Calibration solutions are traceable to NIST standard material 1828.*

Date 1/28/10

Signed Kristi Robinson
Technician

CMI INC.

80-004191
P. 50 of 50
PM

316 East 9th Street
Owensboro, KY 42303
USA

Part No. 650519 CMI 7/9/09