

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program

Form: BrF-105-G

ETHANOL BREATH STANDARD CYLINDER REPORT

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Phil Wolf

Location: Cumtuer, ND

Intoxilyzer Serial No.: 80-002669

Lot Number: 659358

Cylinder Number: 34

Expiration Date: 9/3/11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 078 %

Test 2: 0. 078 %

Test 3: 0. 078 %

(ATTACH TEST RECORD)

Phil Wolf
Chemical Test Operator Signature

2/2/10 01/29/2010 ^{DK}
Date

Deb Kashur
Reviewed By

03 Feb 2010
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = CAVA 8164.13.00 06/09
01/29/2010 04:47

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 04:48 |
| 02 Std. Gas | 0.078 | 04:48 |
| 03 Room Air | 0.000 | 04:49 |
| 04 Std. Gas | 0.078 | 04:49 |
| 05 Room Air | 0.000 | 04:50 |
| 06 Std. Gas | 0.078 | 04:50 |
| 07 Room Air | 0.000 | 04:51 |


Lot No = 659358

Cyl No = 34

Exp Date = 09/03/2011

County = 34

Oper No. = 120547



Operator Signature
PHIL WOLF

Remarks:

Form 106-I8000

80-002669
P. Wolf