



ETHANOL BREATH STANDARD CYLINDER REPORT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>D. OLDENBURG</u>		
Location: <u>GRAFTON LEC</u>		
Intoxilyzer® Serial Number: <u>80-003072</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>16</u>	Expiration Date: <u>09/03/11</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) DK <u>082</u> % ^{18 Apr 2011} Test 1: <u>0.082</u> % Test 2: <u>0.082</u> % Test 3: <u>0.082</u> %		
(ATTACH TEST RECORD)		

D. Oldenburg 274
Chemical Test Operator Signature

01/29/10
Date

Deb Kashur
Reviewed By

31 Mar 2010
Date

80-003072
P. 21 of 24
80-003072
P. 134 of 135
FM
7477


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003072
Location = GRAF 8164.13.00 06/09
01/29/2010 21:18

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	21:18
02 Std. Gas	0.082	21:19
03 Room Air	0.000	21:19
04 Std. Gas	0.082	21:20
05 Room Air	0.000	21:20
06 Std. Gas	0.082	21:20
07 Room Air	0.000	21:21

Lot No = 659358
Cyl No = 16
Exp Date = 09/03/2011
County = 50 Oper No. = 048915


Operator Signature
DOLF H. OLDENBURG

Remarks:

Form 106-I8000

80-003072
P. 2 of 2 RM
80-003072
P. 135 of 135
RM
77/77