

**BREATH ALCOHOL INSTRUMENT CHAIN OF CUSTODY WORKSHEET**

Instrument: 8000 Serial Number: 80-003067

Transported Out of the Crime Laboratory:

Agency Being Assigned Instrument: \_\_\_\_\_

Name/Agency of Person Transporting Instrument:  
\_\_\_\_\_  
(Print Name) \_\_\_\_\_ (Print Agency)

Signature of Person Receiving Instrument: \_\_\_\_\_

Date Receiving Instrument: \_\_\_\_/\_\_\_\_/\_\_\_\_ From: \_\_\_\_\_

Or Shipped Via:  
USPS \_\_\_\_ UPS \_\_\_\_ FedEx \_\_\_\_ Other \_\_\_\_\_

Date Sending Instrument: \_\_\_\_/\_\_\_\_/\_\_\_\_ Air Bill No. \_\_\_\_\_

Complete Air Bill (Identify Instrument/Serial Number) and attach a copy to this worksheet.

Transported to the Crime Laboratory:

Agency Returning Instrument: CMI

Name/Agency of Person Returning Instrument:  
\_\_\_\_\_  
(Print Name) \_\_\_\_\_ (Print Agency)

Signature of Person Returning Instrument: \_\_\_\_\_

Date Returning Instrument: 30, Jan, 2013 Received By: Stacie Fleck-Merkel  
(DK)

Or Received Via:  
USPS \_\_\_\_ UPS X FedEx \_\_\_\_ Other \_\_\_\_\_

Date Receiving Instrument: \_\_\_\_/\_\_\_\_/\_\_\_\_ Air Bill No. \_\_\_\_\_

Attach shipping paperwork (if any) to this worksheet.

Instrument Transfer Information Entered Into the Equipment Database:

Date of Action 31, Jan, 2013 By: Deb Kashur



# SERVICE WORK ORDER

316 E. 9<sup>th</sup> St., Owensboro KY 42303  
Phone 866-835-0690  
Fax: 270-685-6268

330482

DATE	11-16-12
RCVD VIA	FEDX

<b>BILL TO:</b>		
ATTN:		
CITY	STATE	ZIP
CUSTOMER NO.	585013	
PHONE		

<b>SHIP TO:</b> Attorney Gen. Off. Crime Lab		
ATTN: Del Kasher		
2641 E. Main Ave.		
CITY	STATE	ZIP
Bismarck	ND	58501
PHONE	701-328-6185	
FAX	701-328-6159	

MODEL #	8000	SERIAL #	80-003067	BILL CODE	
EXTRA PARTS RCVD: BL, pin box, battery					
DESCRIPTION OF PROBLEM: RAM Error in field					

**Infrared / Fuel Cell**  
PRELIMINARY TEST INFORMATION  
BATTERY- EXT 12 V; INT 2.918 V  
Solenoid leaky, hole blocked in check valve, battery discharged fast  
VERIFIED PROBLEM? YES

**WORK PERFORMED:**  
Cleared Error  
Replaced parts listed  
Unit checks fine

PARTS USED		
QTY	PART NO.	DESCRIPTION
1	021492 FL	CHK Valve
1	013010	Battery
1	021357	Solenoid

**Miscellaneous**  
PRELIMINARY TEST INFORMATION  
**WORK PERFORMED:**  
MOT

REPAIR	CAL. ADJ.	CALIBRATION	FINAL
JMT	DR	DR	DR
1/2 HOURS	HOURS	HOURS	HOURS
4	1	.5	.5
DATE	DATE	DATE	DATE
JAN 23-2013	JAN 23-13	JAN 23-13	JAN 23-13

SHIPPING STICKER HERE

SPECIAL SHIPPING INSTRUCTIONS