



# ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>David Feland</u>		
Location: <u>Kidder S.O.</u>		
Intoxilyzer® Serial Number: <u>80-004186</u> <sup>DK</sup>		
Gas Lot Number: <u>659357</u>	Gas Cylinder Number: <u>137 37</u>	Expiration Date: <u>09/03/11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.083</u> %		
Test 2: <u>0.082</u> %		
Test 3: <u>0.082</u> %		
(ATTACH TEST RECORD)		

David Feland  
Chemical Test Operator Signature

31 Jan 11  
Date

Deb Kashur  
Reviewed By

16 Feb 2011  
Date

80-004186  
P. 97 of 98  
FM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004186  
Location = STEE      8164.13.00 06/09  
01/31/2011      13:43

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:43
02 Std. Gas	0.083	13:44
03 Room Air	0.000	13:44
04 Std. Gas	0.082	13:45
05 Room Air	0.000	13:45
06 Std. Gas	0.082	13:46
07 Room Air	0.000	13:46


Lot No = 659357

Cyl No = 37

Exp Date = 09/03/2011

County = 22

Oper No. = 119404



Operator Signature  
DAVID R. FELAND

Remarks:

Form 106-I8000

80-004186  
P. 98 of 98  
AM