

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program

Form: BrF-105-G

**ETHANOL BREATH STANDARD CYLINDER REPORT**

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: David M Wallace

Location: Dickinson IEC

Intoxilyzer Serial No.: 80 - 0030166

Lot Number: 659358

Cylinder Number: 23

Expiration Date: 9-3-11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.083%

Test 2: 0.082%

Test 3: 0.083%

(ATTACH TEST RECORD)

David M Wallace  
Chemical Test Operator Signature

2-7-2010  
Date

Deb Kashner  
Reviewed By

02-03-2010  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003066  
Location = DKPD      8164.13.00 06/09  
02/01/2010      10:59

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:59
02 Std. Gas	0.083	11:00
03 Room Air	0.000	11:00
04 Std. Gas	0.082	11:01
05 Room Air	0.000	11:01
06 Std. Gas	0.083	11:01
07 Room Air	0.000	11:02

Lot No = 659358

Cyl No = 23

Exp Date = 09/03/2011

County = 45

Oper No. = 068332



Operator Signature  
DAVID WALLACE

Remarks:

Form 106-I8000

80-003066  
P. 2 of 2