



OFFICE OF ATTORNEY GENERAL
 Crime Laboratory Division
 2641 East Main Avenue
 P.O. Box 937
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Patrick C Hudson

Location: Ward County Jail

Intoxilyzer Serial No.: 80-003058

Lot Number: 659358

Cylinder Number: 32

Expiration Date: 9/3/2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 082 %

Test 2: 0. 082 %

Test 3: 0. 082 %

(ATTACH TEST RECORD)

Patrick C Hudson

Chemical Test Operator Signature

Deb Kashner

2/2/2010

Date Cylinder Changed

04 Feb 2010

Form 105-G
 Toxicology Laboratory
 Crime Laboratory Division

80-003058
 P.142 PM

7/8/09

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003058
Location = WARD 8164.13.00 06/09
02/02/2010 16:17

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:18
02 Std. Gas	0.082	16:18
03 Room Air	0.000	16:19
04 Std. Gas	0.082	16:19
05 Room Air	0.000	16:20
06 Std. Gas	0.082	16:20
07 Room Air	0.000	16:21

Lot No = 659358
Cyl No = 32
Exp Date = 09/03/2011
County = 51 Oper No. = 049509



Operator Signature
PATRICK C. HUDSON

Remarks:

Form 106-I8000

80-003058
P. 2 of 2 pm
7/9/99