



# ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Aaron R. Hummel</u>		
Location: <u>CFSD</u>		
Intoxilyzer® Serial Number: <u>80-004205</u>		
Gas Lot Number: <u>15410080A1</u>	Gas Cylinder Number: <u>23</u>	Expiration Date: <u>7/1/12</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>081</u> %		
Test 2: 0. <u>082</u> %		
Test 3: 0. <u>081</u> %		
(ATTACH TEST RECORD)		

Aaron R. Hummel  
Chemical Test Operator Signature

2/3/11  
Date

Deb Keshur  
Reviewed By

15 Feb 2011  
Date

80-004205  
P 101 & 103  
FM

## Instructions for Calibration Checks on the Intoxilyzer® 8000

### **Purpose:**

To check the calibration of the Intoxilyzer® 8000. This should be done about every 45 days or 50 tests, whichever comes first. All chemical test operators are qualified to check the instrument calibration.

### **Complete:**

1. An ACA test.
2. The top portions of the Intoxilyzer® Record SFN50496 - Form 120-G (Pages 1 and 2) and place it by the Intoxilyzer® 8000.
3. The Ethanol Breath Standard Cylinder Report SFN59282 - Form 105-G.

### **Send in the Following to the Crime Laboratory:**

1. ACA Test Record Form 106-I8000.
2. Completed SFN59282 - Form 105-G.
3. Previously completed SFN50496 - Form 120-G forms with accompanying subject test records.

80-064205  
p. 102 of 103  
SM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004205  
Location = GFSO      8164.13.00 06/09  
02/03/2011      08:46

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:47
02 Std. Gas	0.081	08:47
03 Room Air	0.000	08:48
04 Std. Gas	0.082	08:48
05 Room Air	0.000	08:48
06 Std. Gas	0.081	08:49
07 Room Air	0.000	08:49

Lot No = 15410080A1  
Cyl No = 23  
Exp Date = 07/01/2012  
County = 18      Oper No. = 049907



Operator Signature  
AARON HUMMEL

Remarks:

ACA

Form 106-I8000

80-004205  
P.103 of 163  
FM