



ETHANOL BREATH STANDARD CYLINDER REPORT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Danielle m. Roscraven</u>		
Location: <u>CARR</u>		
Intoxilyzer® Serial Number: <u>80-004944</u>		
Gas Lot Number: <u>709352</u>	Gas Cylinder Number: <u>12</u>	Expiration Date: <u>12/23/11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>080</u> %		
Test 2: 0. <u>080</u> %		
Test 3: 0. <u>079</u> %		
(ATTACH TEST RECORD)		

Danielle Roscraven
Chemical Test Operator Signature

02/03/11
Date

Deb Kashner
Reviewed By

15 Feb 2011
Date

80-004944
p. 59 of 60
sam

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = CARR 8164.13.00 06/09
02/03/2011 19:36

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	19:37
02 Std. Gas	0.080	19:37
03 Room Air	0.000	19:38
04 Std. Gas	0.080	19:38
05 Room Air	0.000	19:39
06 Std. Gas	0.079	19:39
07 Room Air	0.000	19:40

Lot No = 709352
Cyl No = 12
Exp Date = 12/23/2011
County = 16 Oper No. = 100754

Danielle Rosewren

Operator Signature
DANIELLE M. ROSEWAREN

Remarks:

Form 106-I8000

80-004944
P. 60 of 60
JRM