



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Pete Nelson</u>		
Location: <u>West Fargo PD</u>		
Intoxilyzer® Serial Number: <u>80-00305-9</u>		
Gas Lot Number: <u>659357</u>	Gas Cylinder Number: <u>10</u>	Expiration Date: <u>9-3-11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>082</u> %		
Test 2: 0. <u>082</u> %		
Test 3: 0. <u>081</u> %		
(ATTACH TEST RECORD)		

[Signature]
Chemical Test Operator Signature

2-4-11
Date

Deb Kashur
Reviewed By

15 Feb 2011
Date

80-003059
p. 101 of 102
JFM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003059
Location = WFPD 8164.13.00 06/09
02/04/2011 14:14

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:15
02 Std. Gas	0.082	14:16
03 Room Air	0.000	14:16
04 Std. Gas	0.082	14:16
05 Room Air	0.000	14:17
06 Std. Gas	0.081	14:17
07 Room Air	0.000	14:18

Lot No = 659357
Cyl No = 10
Exp Date = 09/03/2011
County = 09 Oper No. = 039617



Operator Signature
PETER D. NIELSEN

Remarks:

Form 106-I8000

80-003059
P. 102 of 102
YFM