

**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**

Serial No.: 80-004189 Instrument Location: Williston LEC

Reason for Install/Repair: \_\_\_\_\_ Install After Lab Repair/Inspection  
X Other (Specify) New

Check When Done:

- X A. Surge Protector Installed/Properly Grounded.
- X B. Telephone Line Connected to Intoxilyzer®.
- X C. Breath Tube Heated.
- X D. Review/Enter Preliminary Data Entry (Level 2, Function E).
- X E. Review/Enter Gas Setup (Level 1, Function S).
- X F. Print Test (Level 1, Function P). (Attach Test Record.)
- X G. ACA Test (Level 1, Function C). (Attach Test Record.)
- X H. RFI Test (Use CMS Mode-First Room Air). (Attach Test Record.)
- I. Repair and/or Maintenance Performed: \_\_\_\_\_

02-05-2010  
Date

Darryn W. Abelle  
Field Inspector's Signature

10 Feb 2010  
Date

Deb Kashner  
Reviewed By

80-004189  
P. 1 of 4  
JAN



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004189  
Location = WILL      8164.13.00 06/09  
02/05/2010      13:09

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:10
02 Std. Gas	0.077	13:10
03 Room Air	0.000	13:11
04 Std. Gas	0.078	13:11
05 Room Air	0.000	13:12
06 Std. Gas	0.077	13:12
07 Room Air	0.000	13:13

Lot No = 659357  
Cyl No = 45  
Exp Date = 09/03/2011  
County = 53      Oper No. = 010101

*Darcy W Aberle*

Operator Signature  
DARCY W. ABERLE

Remarks:

ACA test

Form 106-I8000

80-004189  
P. 3 of 4  
JAN

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004189  
Location = WILL      8164.13.00 06/09  
02/05/2010      13:14

Test	AC	Time
01 Diagnostic	OK	13:15
02 Room Air	RFI*	13:16
03 Room Air	0.000	13:16

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, SUBJECT NA  
Sub DOB = 08/08/1980  
Sub Sex = Male      Weight = 200  
Test = OTH      Cit = NA  
Dr. Lic. = ND/NA  
Lot No = 659357  
Cyl No = 45  
Expiration Date = 09/03/2011  
County = 53      Oper No. = 010101

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
DARCY W. ABERLE

Remarks:

RFI test

Form 106-I8000

80-004189  
P. 9/4/11