

ETHANOL BREATH STANDARD CYLINDER REPORT

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Darcy W Aberle

Location: Williston LEC

Intoxilyzer Serial No.: 80-004189

Lot Number: 659357

Cylinder Number: 45

Expiration Date: 09/03/2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.077 %

Test 2: 0.078 %

Test 3: 0.077 %

(ATTACH TEST RECORD)

Darcy W Aberle
Chemical Test Operator Signature

02-05-2010
Date

Deb Kasher
Reviewed By

10 Feb 2010
Date

80-004189
P. 1 of 2
FM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = WILL 8164.13.00 06/09
02/05/2010 13:00

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:00
02 Std. Gas	0.077	13:01
03 Room Air	0.000	13:01
04 Std. Gas	0.078	13:02
05 Room Air	0.000	13:02
06 Std. Gas	0.077	13:02
07 Room Air	0.000	13:03

Lot No = 659358
Cyl No = 34
Exp Date = 09/03/2011
County = 53 Oper No. = 010101

Darcy W Aberle

Operator Signature
DARCY W. ABERLE

Remarks:

For 105G

Form 106-I8000

80-004189
P. 2 of 2
sm