



**ETHANOL BREATH STANDARD CYLINDER REPORT**  
 OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION  
 Toxicology Section/Breath Alcohol Program  
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>RAY Weber</u>		
Location: <u>TRAIL CO. S.D.</u>		
Intoxilyzer® Serial Number: <u>80-004200</u>		
Gas Lot Number: <u>15410080A1</u>	Gas Cylinder Number: <u>#15</u>	Expiration Date: <u>07/01/12</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: <u>0.078</u> % Test 2: <u>0.077</u> % Test 3: <u>0.077</u> %		
(ATTACH TEST RECORD)		

Ray Weber  
 Chemical Test Operator Signature

2-5-11  
 Date

Deb Kashner  
 Reviewed By

15 Feb 2011  
 Date

80-004200  
 P. 99 of 100  
 FM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004200  
Location = HILL      8164.13.00 06/09  
02/05/2011      20:06

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	20:07
02 Std. Gas	0.078	20:07
03 Room Air	0.000	20:08
04 Std. Gas	0.077	20:08
05 Room Air	0.000	20:09
06 Std. Gas	0.077	20:09
07 Room Air	0.000	20:10

Lot No = 15410080A1

Cyl No = 15

Exp Date = 07/01/2012

County = 49

Oper No. = 058405



Operator Signature  
RAY G. WEBER

Remarks:

Form 106-I8000

80-004200  
P.100 of 100  
RM