



ETHANOL BREATH STANDARD CYLINDER REPORT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Will Vance</u>		
Location: <u>Dickinson LCC</u>		
Intoxilyzer® Serial Number: <u>80-004936</u>		
Gas Lot Number: <u>15410080A1</u>	Gas Cylinder Number: <u>20</u>	Expiration Date: <u>7-1-12</u>
<input checked="" type="checkbox"/> New Cylinder Installed	<input type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.079</u> %		
Test 2: <u>0079</u> %		
Test 3: <u>0.079</u> %		
(ATTACH TEST RECORD)		

Will Vance
Chemical Test Operator Signature

2-6-11
Date

Deb Kashner
Reviewed By

15 Feb 2011
Date

80-004936
P. 64 of 65
ym


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004936
Location = DKPD 8164.13.00 06/09
02/06/2011 17:00

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:00
02 Std. Gas	0.079	17:01
03 Room Air	0.000	17:01
04 Std. Gas	0.079	17:01
05 Room Air	0.000	17:02
06 Std. Gas	0.079	17:02
07 Room Air	0.000	17:03

Lot No = 15410080A1
Cyl No = 20
Exp Date = 07/01/2012
County = 45 Oper No. = 019620



Operator Signature
WILLIAM D. VANCE

Remarks: ACA

Form 106-I8000

80-004936
P. 65 of 65
FM