



**ETHANOL BREATH STANDARD CYLINDER REPORT**  
 OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION  
 Toxicology Section/Breath Alcohol Program  
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Sean Murphy</u>		
Location: <u>Bottineau County Sheriff's Dept.</u>		
Intoxilyzer® Serial Number: <u>80-003065</u>		
Gas Lot Number: <u>16512080A2</u>	Gas Cylinder Number: <u>06</u>	Expiration Date: <u>8-1-14</u>
<input checked="" type="checkbox"/> New Cylinder Installed <input type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>081</u> %		
Test 2: 0. <u>081</u> %		
Test 3: 0. <u>082</u> %		

(ATTACHES)

CHI, Inc. Intoxilyzer Alcohol Analyzer  
 North Dakota Model 8000 SN 80-003065  
 Location = BOTT 8164.13.00 06/09  
 02/06/2013 22:11

Chemical Test Operator Signature

Deb Kashur

Reviewed By

Date

Date

2-6-13

25 Feb 13

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	22:11
02 Std. Gas	0.082	22:12
03 Room Air	0.000	22:12
04 Std. Gas	0.081	22:12
05 Room Air	0.000	22:13
06 Std. Gas	0.082	22:13
07 Room Air	0.000	22:14

Lot No = 16512080A2  
 Cyl No = 6  
 Exp Date = 08/01/2014  
 County = 05 Oper No. = 039117

Operator Signature  
 SEAN J. MURPHY

Remarks: