



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION

Toxicology Section/Breath Alcohol Program

SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Brenda Janssen</u>		
Location: <u>Mohall, Renoille County Sheriff Office</u>		
Intoxilyzer® Serial Number: <u>80-003069</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>27</u>	Expiration Date: <u>09/03/2011</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: 0. <u>081</u> % Test 2: 0. <u>081</u> % Test 3: 0. <u>080</u> %		
(ATTACH TEST RECORD)		

Brenda Janssen
Chemical Test Operator Signature

01/17/11
Date

Deb Kashner
Reviewed By

15 Feb 2011
Date

80-003069
P. 89 of 90
AM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003069
Location = MOHL 8164.13.00 06/09
02/07/2011 18:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:41
02 Std. Gas	0.081	18:42
03 Room Air	0.000	18:42
04 Std. Gas	0.081	18:42
05 Room Air	0.000	18:43
06 Std. Gas	0.080	18:43
07 Room Air	0.000	18:44

Lot No = 659358
Cyl No = 27
Exp Date = 09/03/2011
County = 38 Oper No. = 011102



Operator Signature
BRENDA K. JANSSEN

Remarks:

Form 106-I8000

80-003069
P. 90 of 90
SEM