



**ETHANOL BREATH STANDARD CYLINDER REPORT**  
 OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION  
 Toxicology Section/Breath Alcohol Program  
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Matthew Anderson</u>		
Location: <u>Richmond County Jail</u>		
Intoxilyzer® Serial Number: <u>80-004206</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>52</u>	Expiration Date: <u>09-03-2011</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: 0. <u>084</u> % Test 2: 0. <u>084</u> % Test 3: 0. <u>084</u> %		
(ATTACH TEST RECORD)		

  
 Chemical Test Operator Signature

2-07-2011  
 Date

Deb Kashner  
 Reviewed By

15 Feb 2011  
 Date


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004206  
Location = WAHP      8164.13.00 06/09  
02/07/2011      14:36

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:37
02 Std. Gas	0.084	14:37
03 Room Air	0.000	14:38
04 Std. Gas	0.084	14:38
05 Room Air	0.000	14:39
06 Std. Gas	0.084	14:39
07 Room Air	0.000	14:40

Lot No = 659358  
Cyl No = 52  
Exp Date = 09/03/2011  
County = 39      Oper No. = 020401



Operator Signature  
MATTHEW W. ANDERSON

Remarks:

Form 106-I8000

80-004206  
P. 113 of 113  
JFM