

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program

Form: BrF-105-G

**ETHANOL BREATH STANDARD CYLINDER REPORT**

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Lt. Tracy L. Meidinger # 258

Location: UND - Police Dept.

Intoxilyzer Serial No.: 80-004206

Lot Number: 659358

Cylinder Number: 31

Expiration Date: 09/03/2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 084 %

Test 2: 0. 084 %

Test 3: 0. 084 %

(ATTACH TEST RECORD)

Lt. Tracy L. Meidinger  
Chemical Test Operator Signature

DK 02/08/2010  
03-14-2010  
Date

Deb Kashner  
Reviewed By

23 Mar 2010  
Date

80-004206  
P.12/3/10

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004206  
Location = UNDP      8164.13.00 06/09  
02/08/2010      18:12

DRY CAL CHECK

| Test        | AC    | Time  |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 18:13 |
| 02 Std. Gas | 0.084 | 18:13 |
| 03 Room Air | 0.000 | 18:13 |
| 04 Std. Gas | 0.084 | 18:14 |
| 05 Room Air | 0.000 | 18:14 |
| 06 Std. Gas | 0.084 | 18:15 |
| 07 Room Air | 0.000 | 18:15 |

Lot No = 659358

Cyl No = 31

Exp Date = 09/03/2011

County = 18

Oper No. = 119409

Tracy L. Meidinger #258  
Operator Signature  
TRACY L. MEIDINGER

Remarks:

Form 106-I8000

ACA  
02-08-2010

80-004206  
P. 2 of 2  
7/10