

# Intoxilyzer 8000 Initial Inspection

<b>Serial Number</b>	<b>80-004938</b>
----------------------	------------------

- A. Power On  5 min. countdown  1 min.
- B. Security Checks.  Level 1  Level 2  Level 3
- C. Gas Cylinder Setup
- D. Printer. Level 1 Func P  Paper Feed  Internal  External
- E. Display is working.
- F. Configure Start Test Switch. Level 3 Func S Sub T
- G. Setup Preliminary Data. Level 2 Func E
- H. Configure General Setup. Level 3 Func S Sub G.
- I. Configure Port. Level 3 Func S Sub N
- J. Check Camera. ABA Mode  Oper ID  ND Lic.  Hot Key "R"
- K. Diagnostic Test. Level 2 Func D
- L. Check Ser Num / Version Level 2 Func V  
80-004938 Serial No. 816413 Version 1362.14 Display 7 CPLD
- M. Detector Voltages. Level 3 Func D Sub D  
3 12661 s 0 C 47.00  
9 12948 s 1 B 44 0 RFI Ind.
- N. Hot Keys
  - F5 Tone
  - F6 Green LED
  - F7 Red LED
  - F9 Fan
  - F10 Pump
  - F11 Sol Output
- O. RFI Dectector
- P. Atmospheric Monitor. Level 3 Func D Sub A. P 1020 CF 0.99

08/09

80-004938  
6/24 KRL

OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION  
 Toxicology Section/Breath Alcohol Program  
 Instrument Certification and Maintenance Unit

BrF-005-8

Q. Flow Monitor. Level 3 Func D Sub F. R 253 F 0.000 L/s  
 R. Temperature Monitor. Level 3 Func D Sub T  
 C 47.0 A 47.5 I 0.0  
 B 45.1 V 0.0  
 S. Gas Tank Monitor. Level 3 Func D Sub G  
 T. Configure Wet Bath Sim. Level 1 Func S  
983 psi Intox 990 psi Reg  
 U. Zero Alcohol Test Level 1 Func C  
DR3409 Sim SN      Lot 0.000 Rep Value  
0.000 AC 0.000 AC 0.000 AC  
 V. Low AC Test. Level 1 Func C  
DR5144 Sim SN      Lot 0.019 Rep Value  
0.019 AC 0.019 AC 0.020 AC  
 W. Linearity Test. Level 1 Func C  
DR5187 Sim SN      Lot 0.285 Rep Value  
0.289 AC 0.289 AC 0.289 AC  
 X. Interferent Test.  
 Level 1 Func B  
 Acetone DR 3845 Sim SN      Lot 0.05% Conc  
 Display  Printout  
 Methanol 6-10573 Sim SN      Lot 0.10% Conc  
 Display  Printout  
 IPA DR5191 Sim SN      Lot 0.10% Conc  
 Display  Printout  
 Y. Ambient Test  
 Z. Invalid Sample Test  
 AA. RFI Detector Test. CMS Mode  
 Display  Printout  
 BB. Configure Dry Gas Standard. Level 1 Func S  
 CC. Stability Test at 0.080AC. Level 3 Func T Sub S  
0.0802 Ave. 0.0005 Std Dev 0.66% Rel Std Dev  
 DD. Maintenance/Repairs Recalibrated

08 Feb 2010  
 Date  
12 Feb 2010  
 Date

Deb Kashner  
 Signature  
Mary Pearson  
 Reviewed by

80-004938  
 7/24 KM

08/09

TOXL  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-004938  
 02/05/2010 14:39:53

Auto Calibration

pg 1 of 2

<<<<<		3um	>>>>>	<<<<<		9um	>>>>>
Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1							
Sample	% Abs	(% Abs Ref)		% Abs	(% Abs Ref)		
Sample #1	0.1370	(-0.0190)		0.2530	(-0.0220)		
Sample #2	0.1240	(-0.0060)		0.2400	(-0.0150)		
Sample #3	0.1450	(0.0050)		0.2370	(-0.0060)		
Sample #4	0.1100	(0.0240)		0.2320	(-0.0070)		
Avg % Abs	0.1263	(0.0077)		0.2363	(-0.0093)		
STD DEV	0.0176	(0.0152)		0.0040	(0.0049)		
REL STD DEV	13.944	(197.957)		1.710	(52.852)		

Solution = 0.040 g/210L or 0.1905 mg/l, Samples = 4, Discarded = 1							
Sample	% Abs	(% Abs Ref)		% Abs	(% Abs Ref)		
Sample #1	0.8480	(-0.0120)		1.6410	(-0.0230)		
Sample #2	0.8310	(0.0000)		1.6000	(0.0080)		
Sample #3	0.8300	(0.0070)		1.6150	(-0.0010)		
Sample #4	0.8540	(0.0000)		1.6100	(-0.0120)		
Avg % Abs	0.8383	(0.0023)		1.6083	(-0.0017)		
STD DEV	0.0136	(0.0040)		0.0076	(0.0100)		
REL STD DEV	1.620	(173.205)		0.475	(600.999)		

Solution = 0.107 g/210L or 0.5095 mg/l, Samples = 4, Discarded = 1							
Sample	% Abs	(% Abs Ref)		% Abs	(% Abs Ref)		
Sample #1	2.0560	(0.0120)		4.0210	(0.0170)		
Sample #2	2.0360	(0.0400)		4.0080	(0.0190)		
Sample #3	2.0850	(0.0310)		4.0090	(0.0210)		
Sample #4	2.0560	(0.0440)		3.9850	(0.0300)		
Avg % Abs	2.0590	(0.0383)		4.0007	(0.0233)		
STD DEV	0.0246	(0.0067)		0.0136	(0.0059)		
REL STD DEV	1.197	(17.370)		0.339	(25.112)		

Solution = 0.147 g/210L or 0.7000 mg/l, Samples = 4, Discarded = 1							
Sample	% Abs	(% Abs Ref)		% Abs	(% Abs Ref)		
Sample #1	2.7580	(-0.0180)		5.3240	(-0.0140)		
Sample #2	2.7540	(0.0040)		5.3180	(-0.0150)		
Sample #3	2.7770	(0.0050)		5.3200	(-0.0070)		
Sample #4	2.7720	(0.0160)		5.3320	(-0.0030)		
Avg % Abs	2.7677	(0.0083)		5.3233	(-0.0083)		
STD DEV	0.0121	(0.0067)		0.0076	(0.0061)		
REL STD DEV	0.437	(79.900)		0.142	(73.321)		

*Calibration  
 Deb Kashur  
 p. 1 of 2*

Solution = 0.285 g/210L or 1.3571 mg/l, Samples = 4, Discarded = 1							
Sample	% Abs	(% Abs Ref)		% Abs	(% Abs Ref)		
Sample #1	5.2820	(-0.0140)		9.9220	(-0.0060)		
Sample #2	5.2250	(0.0300)		9.9270	(0.0140)		
Sample #3	5.2530	(0.0310)		9.9680	(0.0250)		
Sample #4	5.2320	(0.0360)		9.9660	(0.0170)		
Avg % Abs	5.2367	(0.0323)		9.9537	(0.0187)		
STD DEV	0.0146	(0.0032)		0.0231	(0.0057)		
REL STD DEV	0.278	(9.942)		0.232	(30.462)		

*80-004938  
 8/24 KM*

TOXL  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-004938  
 PPPP 0>3A  
 Zero Order Coef -323.64 -298.37  
 PP

Act (g/210L)	Fit (g/210L)	Residual (g/210L)	Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	0.000	-0.0002	0.000	0.000	-0.0004
0.040	0.040	0.0004	0.040	0.039	0.0008
0.107	0.107	-0.0004	0.107	0.108	-0.0009
0.147	0.147	0.0002	0.147	0.147	0.0005
0.285	0.285	0.0000	0.285	0.285	0.0000

<<<<< 3um >>>>> <<<<< 9um >>>>>

Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1

Sample	3um	9um
Sample #1	2997.00	3119.00
Sample #2	2914.00	3112.00
Sample #3	2997.00	3152.00
Sample #4	2926.00	3128.00
Avg	2945.6667	3130.6667
STD DEV	44.8590	20.1329
REL STD DEV	1.523	0.643
H2O adjust (mg/l*10k)	864	679

Atmospheric Pressure = 1020

\*\*\*\*\*CALIBRATION SUCCESSFUL\*\*\*\*\*

*Calibration  
 Deb Keshan  
 9/24/22*

80-004938  
 9/24/22

Intoxilyzer Test Record and Checklist  
NDDAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer  
North Dakota Model 8000 SN 80-004938  
Location = TOXL 8164.13.00 06/09  
02/05/2010 16:37

\*\*\*\*\* Printer Test \*\*\*\*\*

12 char/line  
abcdefghijkl  
ABCDEFGHIJKL

16 char per line  
abcdefghijklmnop  
ABCDEFGHIJKLMN

21 chars per line  
abcdefghijklmnopqrstu  
ABCDEFGHIJKLMNQRSTU

24 characters per line  
abcdefghijklmnopqrstuwx  
ABCDEFGHIJKLMNQRSTUWX

32 characters per line  
abcdefghijklmnopqrstuwx123456  
ABCDEFGHIJKLMNQRSTUWXYZ123456

42 characters per line  
abcdefghijklmnopqrstuwx1234567890abcdef  
ABCDEFGHIJKLMNQRSTUWXYZ1234567890ABCDEF

Current Instrument Setup

Data Entry Mode: Enabled  
Start Test Sequence: DABACABA  
Display Prelim Rslt? Yes  
Display Third Digit? Yes  
Inhib Printer(Y/N)? No  
Display Volume? No  
Disable On MenFull? Yes  
# of Print Copies? 1  
Select Std (0/W/1)? Dry  
Standard Value? 0.080  
Standard Lot #? 668180  
Standard Cyl #? 54  
Standard Expiration? 09/21/2011  
Oper No? 777777

Flow Cal. Date: 01/25/2010  
Slope 712  
Intercept -604165

IR Calibration Date: 02/05/2010  
3um 9um

0th Coef(\*100): -32364 -29837

1st Coef(\*100): 263210 133607

2nd Coef(\*100): 406 576

H2O adj(mg/l\*10K): 864 679

\*\*\*\*\* Test End \*\*\*\*\*

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Print test*

Form 106-18000

80-004938  
10/24 KM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004938  
Location = TOXL                    8164.13.00 06/09  
02/05/2010                            16:38

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-=  
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-=  
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#%&^\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            668180  
Standard Cyl #?            54  
Standard Expiration?        09/21/2011  
Oper No?                    777777

Flow Cal. Date:            01/25/2010  
Slope                        712  
Intercept                    -604165

IR Calibration Date:        02/05/2010  
                                  3um                    9um

	3um	9um
0th Coef(*100):	-32364	-29837
1st Coef(*100):	263210	133607
2nd Coef(*100):	406	576
H2O adj(mg/l*10k):	864	679

\*\*\*\*\* Printer Test End \*\*\*\*\*

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Print Test*

Form 106-I8000

80-004938  
11/24 192

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004938  
Location = TOXL      8164.13.00 06/09  
02/05/2010      16:40

Test	AC	Time
01 Room Air	0.000	16:41
02 *Subject Test	REF*	16:42
03 Room Air	0.000	16:43

\*Subject Test Refused

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 668180  
Cyl No = 54  
Expiration Date = 09/21/2011  
County = 08      Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Camera / Refusal check*

Form 106-I8000

80-004938  
12/24 FM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004938  
Location = TOXL      8164.13.00 06/09  
02/05/2010      16:43

DIAGNOSTIC

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temp Regulation Test	Pass

Oper No. = 77777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Diagnostic test*

Form 106-I8000

*80-004938  
13/24 KM*



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004938  
Location = TOXL      8164.13.00 06/09  
02/08/2010      13:03

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:03
02 Std. Sol.	0.000	13:04
03 Room Air	0.000	13:05
04 Std. Sol.	0.000	13:05
05 Room Air	0.000	13:06
06 Std. Sol.	0.000	13:07
07 Room Air	0.000	13:07

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*zero alcohol      0.000 AC*

Form 106-I8000

*80-004938  
14/24 KH*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004938  
Location = TOXL      8164.13.00 06/09  
02/08/2010      13:31

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:31
02 Std. Sol.	0.289	13:32
03 Room Air	0.000	13:33
04 Std. Sol.	0.289	13:33
05 Room Air	0.000	13:34
06 Std. Sol.	0.289	13:35
07 Room Air	0.000	13:35

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Linearity test 0.285 AC*

Form 106-I8000

80-004938  
15/24 1A2

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004938  
Location = TOXL      8164.13.00 06/09  
02/08/2010      14:00

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:01
02 Std. Sol.	0.019	14:02
03 Room Air	0.000	14:02
04 Std. Sol.	0.019	14:03
05 Room Air	0.000	14:03
06 Std. Sol.	0.020	14:04
07 Room Air	0.000	14:05

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature  
DEB KASHUR

Remarks:

LOW AC      0.019 AC

Form 106-I8000

80-004938  
14/24 VAK

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004938  
Location = TOXL      8164.13.00 06/09  
02/08/2010      14:18

Test	AC	Time
01 Room Air	0.000	14:19
02 *Subject Test	INT*	14:20
03 Room Air	0.000	14:20

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 668180  
Cyl No = 54  
Expiration Date = 09/21/2011  
County = 08      Oper No. = 777777

Deb Kashur  
Operator Signature  
DEB KASHUR

Remarks:

*acetone interferent*

Form 106-18000

80-004938  
12/24/10

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004938  
Location = TOXL      8164.13.00 06/09  
02/08/2010      14:29

Test	AC	Time
01 Room Air	0.000	14:30
02 *Subject Test	INT*	14:30
03 Room Air	0.000	14:31

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 668180  
Cyl No = 54  
Expiration Date = 09/21/2011  
County = 08      Oper No. = 777777

Deb Kashur

Operator Signature  
DEB KASHUR

Remarks:

*Methanol interferent*

Form 106-18000

80-004938  
12/24 FR

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004938  
Location = TOXL      8164.13.00 06/09  
02/08/2010      14:35

Test	AC	Time
01 Room Air	0.000	14:35
02 *Subject Test	INT*	14:36
03 Room Air	0.000	14:36

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 668180  
Cyl No = 54  
Expiration Date = 09/21/2011  
County = 08      Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*IPA interferent*

Form 106-I8000

80-004938  
19/24 RAL

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004938  
Location = TOXL                    8164.13.00 06/09  
02/08/2010                            15:56

Test	AC	Time
01 Room Air	AMB*	15:57
02 Room Air	0.000	15:58

\*Invalid Test  
Check Ambient Conditions

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 668180  
Cyl No = 54  
Expiration Date = 09/21/2011  
County = 08                            Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Ambient test*

Form 106-I8000

80-004938  
2/24/11

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004938  
Location = TOXL                    8164.13.00 06/09  
02/08/2010                            16:02

Test	AC	Time
01 Room Air	0.000	16:03
02 Invalid Sample	X.XXX	16:03
03 Room Air	PUR*	16:04

\*Purge Fail  
\*\*Invalid Test - Mouth Alcohol

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 668180  
Cyl No = 54  
Expiration Date = 09/21/2011  
County = 08                            Oper No. = 777777

Deb Kashur  
Operator Signature  
DEB KASHUR

Remarks:

*Invalid sample test*

Form 106-I8000

80-004938  
2/24 KME



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004938  
Location = TOXL                    8164.13.00 06/09  
02/08/2010                            14:59

Test	AC	Time
01 Diagnostic	OK	14:59
02 Room Air	RFI*	15:00
03 Room Air	0.000	15:00

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 668180  
Cyl No = 54  
Expiration Date = 09/21/2011  
County = 08                            Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*RFI check*

Form 106-I8000

*80004938  
2/24/11*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004938  
Location = TOXL      8164.13.00 06/09  
02/08/2010      15:06

Test	AC	Time
01 Room Air	0.000	15:06
02 Std. Gas	0.080	15:07
03 Room Air	0.000	15:07
04 Std. Gas	0.080	15:08
05 Room Air	0.000	15:08
06 Std. Gas	0.080	15:08
07 Room Air	0.000	15:09
08 Std. Gas	0.080	15:09
09 Room Air	0.000	15:10
10 Std. Gas	0.080	15:10
11 Room Air	0.000	15:11
12 Std. Gas	0.079	15:11
13 Room Air	0.000	15:12
14 Std. Gas	0.081	15:12
15 Room Air	0.000	15:12
16 Std. Gas	0.081	15:13
17 Room Air	0.000	15:13
18 Std. Gas	0.081	15:14
19 Room Air	0.000	15:14
20 Std. Gas	0.080	15:15
21 Room Air	0.000	15:15
22 Std. Gas	0.080	15:15
23 Room Air	0.000	15:16
24 Std. Gas	0.080	15:16
25 Room Air	0.000	15:17
26 Std. Gas	0.080	15:17
27 Room Air	0.000	15:18
28 Std. Gas	0.080	15:18
29 Room Air	0.000	15:18
30 Std. Gas	0.081	15:19
31 Room Air	0.000	15:19
32 Std. Gas	0.081	15:20
33 Room Air	0.000	15:20
34 Std. Gas	0.080	15:21
35 Room Air	0.000	15:21
36 Std. Gas	0.080	15:21
37 Room Air	0.000	15:22
38 Std. Gas	0.080	15:22
39 Room Air	0.000	15:23
40 Std. Gas	0.080	15:23
41 Room Air	0.000	15:24
42 Cal Check Stats:		
43 Average	0.0802	
44 Std Dev	0.0005	
45 Rel Std Dev	0.66%	

*Stability test  
Deb Kashner  
p. 1 of 2*

*80-004938  
23/24 KKL*

Operator Signature  
DEB KASHUR

Remarks:

Lot #: 668180 Cyl. #: 54 Exp. date: 09/21/2011

Form 106-18000

Stability test  
Deb Kashur  
p-2092

80-004938  
29/24 KAK