

INTOXILYZER 5000 G and 8000 INSPECTION

Serial No.: 80-004938

Location: TOXL

Check When Done

- A. Is the warm-up time less than 20 minutes? ✓
- B. Is a three-pronged grounded outlet used? ✓
- C. Is the breath tube heated? ✓
- D. Is the diagnostic check complete? ✓
- E. Is the time, date, and year correct (re-set if necessary)? ✓
- F. Print test. (Attach test record.) ✓
- G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5144 Lot No. Rep AC 0.019 ✓
- H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR5187 Lot No. Rep AC 0.285 ✓
- I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR5114 Lot No. Rep AC 0.100 AC ✓
- J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓

K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. ⁶⁶⁸¹⁸⁰DK 659358 Cyl No. DK 46 54 Exp. Date ^{DK}659358 21 Sept 2011

Test 1 0.080 Test 4 0.079 Test 7 0.079

Test 2 0.080 Test 5 0.080 Test 8 0.080

Test 3 0.080 Test 6 0.080 Test 9 0.080

Average 0.080

L. Remarks/Maintenance record: Annual inspection

08 Feb 2010
Date

Deb Kashur
Field Inspector's Signature

12 Feb 2010
Date

Maryann A. Bean
Reviewed by

80-004938
25/33 KM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.13.00 06/09
02/05/2010 16:39

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 668180
Standard Cyl #? 54
Standard Expiration? 09/21/2011
Oper No? 777777

Flow Cal. Date: 01/25/2010
Slope 712
Intercept -604165

IR Calibration Date: 02/05/2010
 3um 9um

0th Coef(*100): -32364 -29837
1st Coef(*100): 263210 133607
2nd Coef(*100): 406 576
H2O adj(mg/l*10k): 864 679

***** Printer Test End *****

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Print test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.13.00 06/09
02/08/2010 13:55

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:55
02 Std. Sol.	0.018	13:56
03 Room Air	0.000	13:57
04 Std. Sol.	0.018	13:57
05 Room Air	0.000	13:58
06 Std. Sol.	0.019	13:59
07 Room Air	0.000	13:59

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Low AC 0.019 AC

Form 106-18000

*80-004938
2/33 KMR*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.13.00 06/09
02/08/2010 13:37

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:37
02 Std. Sol.	0.288	13:38
03 Room Air	0.000	13:39
04 Std. Sol.	0.288	13:39
05 Room Air	0.000	13:40
06 Std. Sol.	0.287	13:41
07 Room Air	0.000	13:41

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Linearity test

Form 106-I8000

80-604938
28/33 KLR

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.13.00 06/09
02/08/2010 14:37

Test	AC	Time
01 Room Air	0.000	14:37
02 *Subject Test	INT*	14:38
03 Room Air	0.000	14:39

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 668180
Cyl No = 54
Expiration Date = 09/21/2011
County = 08 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:

Interferent check

Form 106-I8000

80-004938
29/33 FH

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.13.00 06/09
02/08/2010 15:01

Test	AC	Time
01 Diagnostic	OK	15:02
02 Room Air	RFI*	15:02
03 Room Air	0.000	15:02

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 668180
Cyl No = 54
Expiration Date = 09/21/2011
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

RFI check

Form 106-I8000

80-004938
30/33 1/4

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.13.00 06/09
02/05/2010 16:28

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:29
02 Std. Gas	0.080	16:29
03 Room Air	0.000	16:30
04 Std. Gas	0.080	16:30
05 Room Air	0.000	16:31
06 Std. Gas	0.080	16:31
07 Room Air	0.000	16:32

Lot No = 668180
Cyl No = 54
Exp Date = 09/21/2011
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check

Form 106-I8000

80-004938
3/13 KH

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.13.00 06/09
02/08/2010 14:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:41
02 Std. Gas	0.079	14:41
03 Room Air	0.000	14:42
04 Std. Gas	0.080	14:42
05 Room Air	0.000	14:43
06 Std. Gas	0.080	14:43
07 Room Air	0.000	14:44

Lot No = 668180
Cyl No = 54
Exp Date = 09/21/2011
County = 08 Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

80-004938
32/33 FL

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.13.00 06/09
02/08/2010 14:44

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:45
02 Std. Gas	0.079	14:45
03 Room Air	0.000	14:46
04 Std. Gas	0.080	14:46
05 Room Air	0.000	14:47
06 Std. Gas	0.080	14:47
07 Room Air	0.000	14:48

Lot No = 668180
Cyl No = 54
Exp Date = 09/21/2011
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check

Form 106-I8000

80-004938
33/33 FA