

## Intoxilyzer 8000 Initial Inspection

<b>Serial Number</b>	<b>80-004936</b>
----------------------	------------------

A. Power On                       5 min. countdown     1 min.

B. Security Checks.                       Level 1     Level 2     Level 3

C. Gas Cylinder Setup

D. Printer. Level 1 Func P                       Paper Feed     Internal     External

E. Display is working.

F. Configure Start Test Switch. Level 3 Func S Sub T

G. Setup Preliminary Data. Level 2 Func E

H. Configure General Setup. Level 3 Func S Sub G.

I. Configure Port. Level 3 Func S Sub N

J. Check Camera. ABA Mode     Oper ID     ND Lic.     Hot Key "R"

K. Diagnostic Test. Level 2 Func D

L. Check Ser Num / Version Level 2 Func V  
80-004936 Serial No. 816413 Version 1362.14 Display 7 CPLD

M. Detector Voltages. Level 3 Func D Sub D  

 3 12856                      s 2                      C 47.00  
 9 12974                      s 2                      B 44 0 RFI Ind.

N. Hot Keys  

 F5 Tone  
 F6 Green LED  
 F7 Red LED  
 F9 Fan  
 F10 Pump  
 F11 Sol Output

O. RFI Dectector

P. Atmospheric Monitor. Level 3 Func D Sub A.                      P 1028                      CF 0.98

80-004936  
5/23 FM

08/09

**OFFICE OF ATTORNEY GENERAL**  
**CRIME LABORATORY DIVISION**  
 Toxicology Section/Breath Alcohol Program  
 Instrument Certification and Maintenance Unit

BrF-005-8

Q. Flow Monitor. Level 3 Func D Sub F. R 215 F 0.000 L/s

R. Temperature Monitor. Level 3 Func D Sub T  
 C 47.0 A 48.0 I 0.0  
 B 44.9 V 0.0

S. Gas Tank Monitor. Level 3 Func D Sub G  
 T. Configure Wet Bath Sim. Level 1 Func S  
814 psi Intox 800 psi Reg

U. Zero Alcohol Test Level 1 Func C  
DR349 Sim SN — Lot 0.000 Rep Value  
0.000 AC 0.000 AC 0.000 AC

V. Low AC Test. Level 1 Func C  
DR5144 Sim SN — Lot 0.019 Rep Value  
0.018 AC 0.018 AC 0.018 AC

W. Linearity Test. Level 1 Func C  
DR5187 Sim SN — Lot 0.285 Rep Value  
0.281 AC 0.280 AC 0.281 AC

X. Interferent Test.  
 Level 1 Func B  
 Acetone DR3845 Sim SN — Lot 0.05% Conc  
 Display  Printout  
 Methanol DR10573 Sim SN — Lot 0.10% Conc  
 Display  Printout  
 IPA DR5191 Sim SN — Lot 0.10% Conc  
 Display  Printout

Y. Ambient Test  
 Display  Printout

Z. Invalid Sample Test  
 Display  Printout

AA. RFI Detector Test. CMS Mode  
 Display  Printout

BB. Configure Dry Gas Standard. Level 1 Func S

CC. Stability Test at 0.080AC. Level 3 Func T Sub S  
0.0795 Ave. 0.0005 Std Dev 0.65% Rel Std Dev

DD. Maintenance/Repairs Recalibration.

09 Feb 2010  
 Date  
12 Feb 2010  
 Date

Deb Kashur  
 Signature  
Mary Pearson  
 Reviewed by

80-004936  
 4/23/10

TOXL  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-004936  
 02/09/2010 09:16:32

Auto Calibration

pg 1 of 2

	<<<<<	3um	>>>>>	<<<<<	9um	>>>>>
-----						
Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1						
Sample	% Abs	(% Abs Ref)		% Abs	(% Abs Ref)	
Sample #1	0.1010	(0.0070)		0.2770	(-0.0060)	
Sample #2	0.0790	(0.0310)		0.2870	(0.0180)	
Sample #3	0.1320	(0.0310)		0.2810	(0.0190)	
Sample #4	0.1490	(0.0290)		0.2960	(0.0060)	
Avg % Abs	0.1200	(0.0303)		0.2880	(0.0143)	
STD DEV	0.0365	(0.0012)		0.0075	(0.0072)	
REL STD DEV	30.425	(3.807)		2.621	(50.471)	
-----						
Solution = 0.040 g/210L or 0.1905 mg/l, Samples = 4, Discarded = 1						
Sample	% Abs	(% Abs Ref)		% Abs	(% Abs Ref)	
Sample #1	0.8710	(0.0000)		1.7370	(0.0010)	
Sample #2	0.8770	(0.0090)		1.7110	(0.0120)	
Sample #3	0.8680	(0.0200)		1.7120	(0.0160)	
Sample #4	0.8370	(0.0310)		1.6830	(0.0370)	
Avg % Abs	0.8607	(0.0200)		1.7020	(0.0217)	
STD DEV	0.0210	(0.0110)		0.0165	(0.0134)	
REL STD DEV	2.438	(55.000)		0.967	(61.979)	
-----						
Solution = 0.107 g/210L or 0.5095 mg/l, Samples = 4, Discarded = 1						
Sample	% Abs	(% Abs Ref)		% Abs	(% Abs Ref)	
Sample #1	2.1120	(-0.0090)		4.1540	(-0.0190)	
Sample #2	2.1020	(0.0300)		4.0850	(0.0470)	
Sample #3	2.0750	(0.0310)		4.0840	(0.0300)	
Sample #4	2.1070	(0.0250)		4.0750	(0.0470)	
Avg % Abs	2.0947	(0.0287)		4.0813	(0.0413)	
STD DEV	0.0172	(0.0032)		0.0055	(0.0098)	
REL STD DEV	0.822	(11.214)		0.135	(23.746)	
-----						
Solution = 0.147 g/210L or 0.7000 mg/l, Samples = 4, Discarded = 1						
Sample	% Abs	(% Abs Ref)		% Abs	(% Abs Ref)	
Sample #1	2.8300	(-0.0130)		5.4970	(0.0040)	
Sample #2	2.8260	(0.0150)		5.4800	(0.0570)	
Sample #3	2.7860	(0.0390)		5.4840	(0.0330)	
Sample #4	2.8030	(0.0240)		5.4690	(0.0650)	
Avg % Abs	2.8050	(0.0260)		5.4777	(0.0517)	
STD DEV	0.0201	(0.0121)		0.0078	(0.0167)	
REL STD DEV	0.716	(46.632)		0.142	(32.232)	
-----						
Solution = 0.285 g/210L or 1.3571 mg/l, Samples = 4, Discarded = 1						
Sample	% Abs	(% Abs Ref)		% Abs	(% Abs Ref)	
Sample #1	5.4440	(-0.0260)		10.3460	(-0.0060)	
Sample #2	5.4260	(0.0080)		10.2860	(0.0420)	
Sample #3	5.4160	(0.0260)		10.3000	(0.0430)	
Sample #4	5.3950	(0.0310)		10.2630	(0.0490)	
Avg % Abs	5.4123	(0.0217)		10.2830	(0.0447)	
STD DEV	0.0158	(0.0121)		0.0187	(0.0038)	
REL STD DEV	0.292	(55.832)		0.182	(8.476)	
-----						

*Calibration  
 Deb Kashner  
 p. 1 of 2*

80-004936  
 2/9/10  
 2/23 KM

TOXL  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-004936  
 PPPP 0>3A  
 Zero Order Coef -332.94 -381.46  
 PP

Act (g/210L)	Fit (g/210L)	Residual (g/210L)	Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	-0.000	0.0004	0.000	0.000	-0.0001
0.040	0.040	-0.0004	0.040	0.040	0.0002
0.107	0.108	-0.0007	0.107	0.107	-0.0002
0.147	0.146	0.0009	0.147	0.147	0.0001
0.285	0.285	-0.0001	0.285	0.285	0.0000

<<<<< 3um >>>>> <<<<< 9um >>>>>

Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1

Sample	3um	9um
Sample #1	3142.00	3212.00
Sample #2	3126.00	3202.00
Sample #3	3237.00	3206.00
Sample #4	3166.00	3219.00
Avg	3176.3333	3209.0000
STD DEV	56.2168	8.8882
REL STD DEV	1.770	0.277
H2O adjust (mg/l*10k)	633	600

Atmospheric Pressure = 1028

\*\*\*\*\*CALIBRATION SUCCESSFUL\*\*\*\*\*

*Calibration  
 Del Keshner  
 p. 2 of 2*

80-004936  
 8/23 YAL

CMI, Inc. Intoxilyzer Alcohol Analyzer  
North Dakota Model 8000 SN 80-004936  
Location = TOXL 8164.13.00 06/09  
02/09/2010 10:38

\*\*\*\*\* Printer Test \*\*\*\*\*

12 char/line  
abcdefghijkl  
ABCDEFGHIJKL

16 char per line  
abcdefghijklmnop  
ABCDEFGHIJKLMN

21 chars per line  
abcdefghijklmnopqrstu  
ABCDEFGHIJKLMNQRSTU

24 characters per line  
abcdefghijklmnopqrstuwx  
ABCDEFGHIJKLMNQRSTUWX

32 characters per line  
abcdefghijklmnopqrstuwxz123456  
ABCDEFGHIJKLMNQRSTUWXYZ123456

42 characters per line  
abcdefghijklmnopqrstuwxz1234567890abcdef  
ABCDEFGHIJKLMNQRSTUWXYZ1234567890ABCDEF

Current Instrument Setup  
Data Entry Mode: Enabled  
Start Test Sequence: DABACABA  
Display Prelim Rslt? Yes  
Display Third Digit? Yes  
Inhib Printer(Y/N)? No  
Display Volume? No  
Disable On MenFull? Yes  
# of Print Copies? 1  
Select Std (D/W/I)? Dry  
Standard Value? 0.000  
Standard Lot #? 668180  
Standard Cyl #? 54  
Standard Expiration? 09/21/2011  
Oper No? 77777

Flow Cal. Date: 01/18/2010  
Slope 732  
Intercept -767769

IR Calibration Date: 02/09/2010  
3um 9um  
-----  
0th Coef(\*100): -33294 -38146  
1st Coef(\*100): 263151 133532  
2nd Coef(\*100): -1136 209  
H2O adj(mg/l\*10k): 633 600

\*\*\*\*\* Test End \*\*\*\*\*

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Print test*

Form 106-18000

80-004936  
9/23/10

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004936  
Location = TOXL                    8164.13.00 06/09  
02/09/2010                            10:42

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-=  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-=  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            668180  
Standard Cyl #?            54  
Standard Expiration?       09/21/2011  
Oper No?                    777777

Flow Cal. Date:            01/18/2010  
Slope                        732  
Intercept                   -767769

IR Calibration Date:	3um	9um
0th Coef(*100):	-33294	-38146
1st Coef(*100):	263151	133532
2nd Coef(*100):	-1136	209
H2O adj(mg/l*10k):	633	600

\*\*\*\*\* Printer Test End \*\*\*\*\*

Deb Kasher  
Operator Signature  
DEB KASHUR

Remarks:

*Print test*

80-004936  
10/23 sm


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004936  
Location = TOXL      8164.13.00 06/09  
02/09/2010      10:55

Test	AC	Time
01 Room Air	0.000	10:55
02 *Subject Test	REF*	10:56
03 Room Air	0.000	10:56

\*Subject Test Refused

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 668180  
Cyl No = 54  
Expiration Date = 09/21/2011  
County = 08      Oper No. = 777777

  
\_\_\_\_\_  
Operator Signature  
DEB KASHUR

Remarks:

*Camera/Refusal check*

Form 106-I8000

80-004936  
11/23 KM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004936  
Location = TOXL      8164.13.00 06/09  
02/09/2010      10:58

DIAGNOSTIC

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temp Regulation Test	Pass

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Diagnostic test*

Form 106-I8000

80-004936  
12/23 FM



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004936  
Location = TOXL      8164.13.00 06/09  
02/09/2010      11:07

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:08
02 Std. Sol.	0.000	11:08
03 Room Air	0.000	11:09
04 Std. Sol.	0.000	11:10
05 Room Air	0.000	11:10
06 Std. Sol.	0.000	11:11
07 Room Air	0.000	11:11

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*zero alcohol      0.000 AC*

Form 106-I8000

80-004936  
13/23 *LDZ*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004936  
Location = TOXL      8164.13.00 06/09  
02/09/2010      11:12

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:13
02 Std. Sol.	0.018	11:14
03 Room Air	0.000	11:14
04 Std. Sol.	0.018	11:15
05 Room Air	0.000	11:15
06 Std. Sol.	0.018	11:16
07 Room Air	0.000	11:17

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Low AC 0.019 AC*

Form 106-I8000

*80-004936  
14/23 km*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004936  
Location = TOXL      8164.13.00 06/09  
02/09/2010      11:24

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:25
02 Std. Sol.	0.281	11:26
03 Room Air	0.000	11:26
04 Std. Sol.	0.280	11:27
05 Room Air	0.000	11:27
06 Std. Sol.	0.281	11:28
07 Room Air	0.000	11:29

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Linearity test 0.285 AC*

Form 106-I8000

80-004936  
15/23 *WKL*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004936  
Location = TOXL      8164.13.00 06/09  
02/09/2010      13:20

Test	AC	Time
01 Room Air	0.000	13:21
02 *Subject Test	INT*	13:21
03 Room Air	0.000	13:22

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Simul Ser No = 00000  
Std Sol No = 000  
County = 08      Oper No. = 777777

Deb Kashur

Operator Signature  
DEB KASHUR

Remarks:

*acetone interferent*

Form 106-I8000

80-004936  
10/23 JFM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004936  
Location = TOXL      8164.13.00 06/09  
02/09/2010      13:31

Test	AC	Time
01 Room Air	0.000	13:32
02 *Subject Test	INT*	13:32
03 Room Air	0.000	13:33

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Simul Ser No = 00000  
Std Sol No = 000  
County = 08      Oper No. = 777777

Deb Kasher  
Operator Signature  
DEB KASHUR

Remarks:

*Methanol interferent*

Form 106-I8000

80-004936  
1/23 KAL

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004936  
Location = TOXL      8164.13.00 06/09  
02/09/2010      13:39

Test	AC	Time
01 Room Air	0.000	13:46
02 *Subject Test	INT*	13:46
03 Room Air	0.000	13:47

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Simul Ser No = 00000  
Std Sol No = 000  
County = 08      Oper No. = 777777

Deb Kasher  
Operator Signature  
DEB KASHUR

Remarks:

*IPA interferent*

Form 106-I8000

80-004936  
12/23 KAL

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004936  
Location = TOXL      8164.13.00 06/09  
02/09/2010      12:58

Test	AC	Time
01 Room Air	AMB*	12:58
02 Room Air	0.000	12:59

\*Invalid Test  
Check Ambient Conditions

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Simul Ser No = 00000  
Std Sol No = 000  
County = 08      Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*ambient test*

Form 106-I8000

80-004936  
19/23 *FW*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004936  
Location = TOXL                    8164.13.00 06/09  
02/09/2010                            13:00

Test	AC	Time
01 Room Air	0.000	13:01
02 Invalid Sample	X.XXX	13:01
03 Room Air	0.000	13:02

\*Invalid Test - Mouth Alcohol

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Simul Ser No = 00000  
Std Sol No = 000  
County = 08                            Oper No. = 777777

Deb Kashur

Operator Signature  
DEB KASHUR

Remarks:

*Invalid sample test*

Form 106-18000

80-004936  
20/23 FK



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004936  
Location = TOXL      8164.13.00 06/09  
02/09/2010      13:02

Test	AC	Time
01 Diagnostic	OK	13:03
02 Room Air	RFI*	13:03
03 Room Air	0.000	13:03

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Simul Ser No = 00000  
Std Sol No = 000  
County = 08      Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*RFI check*

Form 106-18000

80-004936  
2/13 KH

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004936  
Location = TOXL      8164.13.00 06/09  
02/09/2010      11:38

Test	AC	Time
01 Room Air	0.000	11:39
02 Std. Gas	0.079	11:39
03 Room Air	0.000	11:40
04 Std. Gas	0.079	11:40
05 Room Air	0.000	11:41
06 Std. Gas	0.079	11:41
07 Room Air	0.000	11:42
08 Std. Gas	0.079	11:42
09 Room Air	0.000	11:43
10 Std. Gas	0.079	11:43
11 Room Air	0.000	11:43
12 Std. Gas	0.080	11:44
13 Room Air	0.000	11:44
14 Std. Gas	0.079	11:45
15 Room Air	0.000	11:45
16 Std. Gas	0.080	11:46
17 Room Air	0.000	11:46
18 Std. Gas	0.079	11:46
19 Room Air	0.000	11:47
20 Std. Gas	0.080	11:47
21 Room Air	0.000	11:48
22 Std. Gas	0.079	11:48
23 Room Air	0.000	11:49
24 Std. Gas	0.079	11:49
25 Room Air	0.000	11:50
26 Std. Gas	0.080	11:50
27 Room Air	0.000	11:50
28 Std. Gas	0.079	11:51
29 Room Air	0.000	11:51
30 Std. Gas	0.080	11:52
31 Room Air	0.000	11:52
32 Std. Gas	0.080	11:53
33 Room Air	0.000	11:53
34 Std. Gas	0.080	11:53
35 Room Air	0.000	11:54
36 Std. Gas	0.080	11:54
37 Room Air	0.000	11:55
38 Std. Gas	0.080	11:55
39 Room Air	0.000	11:56
40 Std. Gas	0.079	11:56
41 Room Air	0.000	11:57
42 Cal Check Stats:		
43 Average	0.0795	
44 Std Dev	0.0005	
45 Rel Std Dev	0.65%	

*Stability check  
Deb Kashner  
p. 1 of 2*

*80-004936  
22/23 PA*

Operator Signature  
DEB KASHUR

Remarks:

~~DK  
Lot #~~

Lot #: 668180

Cyl. #: 54

Exp. date: 09/21/2011

Form 106-I8000

Stability check  
Deb Kashur  
p. 2 of 2

80-004936  
23/23 FH