

INTOXILYZER 5000 G and 8000 INSPECTION

Serial No.: 80-004936

Location: TOXL

- | | Check When Done |
|--|-----------------|
| A. Is the warm-up time less than 20 minutes? | <u>✓</u> |
| B. Is a three-pronged grounded outlet used? | <u>✓</u> |
| C. Is the breath tube heated? | <u>✓</u> |
| D. Is the diagnostic check complete? | <u>✓</u> |
| E. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| F. Print test. (Attach test record.) | <u>✓</u> |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR5144</u> Lot No. <u> </u> Rep AC <u>0.019</u> | <u>✓</u> |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR5187</u> Lot No. <u> </u> Rep AC <u>0.285</u> | <u>✓</u> |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN <u>DR5114</u> Lot No. <u> </u> Rep AC <u>0.100</u> | <u>✓</u> |
| J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) | <u>✓</u> |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) | |

Lot No. 668180 Cyl No. 54 Exp. Date 09/21/2011

Test 1 0.079 Test 4 0.079 Test 7 0.079

Test 2 0.079 Test 5 0.079 Test 8 0.080

Test 3 0.079 Test 6 0.079 Test 9 0.079

Average 0.079

L. Remarks/Maintenance record: Annual inspection

09 Feb 2010
Date

Db Kashur
Field Inspector's Signature

12 Feb 2010
Date

Maryn Besor
Reviewed by

80-004936
24/32 KM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004936
Location = TOXL 8164.13.00 06/09
02/09/2010 10:42

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 668180
Standard Cyl #? 54
Standard Expiration? 09/21/2011
Oper No? 777777

Flow Cal. Date: 01/18/2010
Slope 732
Intercept -767769

IR Calibration Date: 02/09/2010
 3um 9um

0th Coef(*100): -33294 -38146
1st Coef(*100): 263151 133532
2nd Coef(*100): -1136 209
H2O adj(mg/l*10k): 633 600

***** Printer Test End *****



Operator Signature
DEB KASHUR

Remarks:

Print test

Form 106-I8000

80-004936
25/32 1514

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004936
Location = TOXL 8164.13.00 06/09
02/09/2010 11:18

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:19
02 Std. Sol.	0.019	11:20
03 Room Air	0.000	11:20
04 Std. Sol.	0.019	11:21
05 Room Air	0.000	11:22
06 Std. Sol.	0.018	11:22
07 Room Air	0.000	11:23

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Low AC 0.019 AC

Form 106-I8000

80-004936
26/32 F12

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004936
Location = TOXL 8164.13.00 06/09
02/09/2010 11:29

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:30
02 Std. Sol.	0.282	11:30
03 Room Air	0.000	11:31
04 Std. Sol.	0.281	11:32
05 Room Air	0.000	11:32
06 Std. Sol.	0.281	11:33
07 Room Air	0.000	11:34

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Linearity test 0.285 AC

Form 106-18000

80-004936
2/27/32 KKH

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004936
Location = TOXL 8164.13.00 06/09
02/09/2010 13:50

Test	AC	Time
01 Room Air	0.000	13:50
02 *Subject Test	INT*	13:51
03 Room Air	0.000	13:52

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Simul Ser No = 00000
Std Sol No = 000
County = 08 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:

Interferent check

Form 106-I8000

80-004936
2/3/2 FH

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004936
Location = TOXL 8164.13.00 06/09
02/09/2010 13:04

Test	AC	Time
01 Diagnostic	OK	13:05
02 Room Air	RFI*	13:05
03 Room Air	0.000	13:05

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Simul Ser No = 00000
Std Sol No = 000
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

RFI test

Form 106-I8000

80-004936
29/32 KM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004936
Location = TOXL 8164.13.00 06/09
02/09/2010 10:43

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:44
02 Std. Gas	0.079	10:44
03 Room Air	0.000	10:44
04 Std. Gas	0.079	10:45
05 Room Air	0.000	10:45
06 Std. Gas	0.079	10:46
07 Room Air	0.000	10:46

Lot No = 668180

Cyl No = 54

Exp Date = 09/21/2011

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

calibration check 0.080 AC

Form 106-I8000

*80-004936
3/32 FH*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004936
Location = TOXL 8164.13.00 06/09
02/09/2010 10:47

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:47
02 Std. Gas	0.079	10:48
03 Room Air	0.000	10:48
04 Std. Gas	0.079	10:49
05 Room Air	0.000	10:49
06 Std. Gas	0.079	10:49
07 Room Air	0.000	10:50

Lot No = 668180
Cyl No = 54
Exp Date = 09/21/2011
County = 08 Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

calibration check 0.080 AC

Form 106-I8000

80-004936
3/32 YK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004936
Location = TOXL 8164.13.00 06/09
02/09/2010 10:50

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:51
02 Std. Gas	0.079	10:51
03 Room Air	0.000	10:52
04 Std. Gas	0.080	10:52
05 Room Air	0.000	10:53
06 Std. Gas	0.079	10:53
07 Room Air	0.000	10:53

Lot No = 668180

Cyl No = 54

Exp Date = 09/21/2011

County = 08

Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

80-004936
32/32 KLR