



# ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Hinton, Rick W</u>		
Location: <u>Cass County Jail</u>		
Intoxilyzer® Serial Number: <u>80-004946</u>		
Gas Lot Number: <u>15412080A1</u>	Gas Cylinder Number: <u>10</u>	Expiration Date: <u>7-1-12</u>
<input checked="" type="checkbox"/> New Cylinder Installed	<input type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.077</u> %		
Test 2: <u>0.077</u> %		
Test 3: <u>0.077</u> %		
(ATTACH TEST RECORD)		

*Rick W Hinton*  
Chemical Test Operator Signature

2/9/11  
Date

Deb Kashur  
Reviewed By

22 Feb 2011  
Date

80-004946  
P. 69 of 70  
JAM


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004946  
Location = CASS      8164.13.00 06/09  
02/09/2011      19:31

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	19:32
02 Std. Gas	0.077	19:32
03 Room Air	0.000	19:33
04 Std. Gas	0.077	19:33
05 Room Air	0.000	19:34
06 Std. Gas	0.077	19:34
07 Room Air	0.000	19:35

Lot No = 668180  
Cyl No = 76  
Exp Date = 09/21/2011  
County = 09      Oper No. = 029920

  
\_\_\_\_\_  
Operator Signature  
RICK W. HINTON

Remarks:

Form 106-I8000

80-004946  
P. 70 of 70  
RM