


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = CASS 8164.13.00 06/09
02/09/2011 19:31

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	19:32
02 Std. Gas	0.077	19:32
03 Room Air	0.000	19:33
04 Std. Gas	0.077	19:33
05 Room Air	0.000	19:34
06 Std. Gas	0.077	19:34
07 Room Air	0.000	19:35

Lot No = ~~668180~~ 15410080 A1 DK 29 Apr 11
Cyl No = 76 10
Exp Date = ~~09/21/2011~~ 07/01/2012
County = 09 Oper No. = 029920


Operator Signature
RICK W. HINTON

Remarks:

Form 106-I8000

80-004946
P. 79 of 79
FM

80-004946
P. 70 of 70
FM