

INTOXILYZER 5000 G and 8000 INSPECTION

Serial No.: 80-004937

Location: TOXL

- | | Check When Done |
|--|-----------------|
| A. Is the warm-up time less than 20 minutes? | <u>✓</u> |
| B. Is a three-pronged grounded outlet used? | <u>✓</u> |
| C. Is the breath tube heated? | <u>✓</u> |
| D. Is the diagnostic check complete? | <u>✓</u> |
| E. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| F. Print test. (Attach test record.) | <u>✓</u> |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR5144</u> Lot No. <u> </u> Rep AC <u>0.019</u> | <u>✓</u> |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR5 DK</u> Lot No. <u> </u> Rep AC <u>0.285</u>
<u>DR5187</u> | <u>✓</u> |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN <u>DR5114</u> Lot No. <u> </u> Rep AC <u>0.100</u> | <u>✓</u> |
| J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) | <u>✓</u> |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) | |

Lot No. 668180 Cyl No. 54 Exp. Date 09/21/2011

Test 1 0.080 Test 4 0.080 Test 7 0.080

Test 2 0.080 Test 5 0.079 Test 8 0.079

Test 3 0.080 Test 6 0.080 Test 9 0.080

Average 0.080

L. Remarks/Maintenance record: Annual inspection

10 Feb 2010
Date

Deb Kashur
Field Inspector's Signature

12 Feb 2010
Date

Mary Pluss
Reviewed by

80-004937
25/33 KPL

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004937
Location = TOXL 8164.13.00 06/09
02/10/2010 14:26

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:27
02 Std. Sol.	0.019	14:29
03 Room Air	0.000	14:29
04 Std. Sol.	0.018	14:30
05 Room Air	0.000	14:30
06 Std. Sol.	0.019	14:31
07 Room Air	0.000	14:31

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Low AC 0.019 AC

Form 106-18000

80-004937
27/33 *KAL*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004937
Location = TOXL 8164.13.00 06/09
02/10/2010 14:44

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:44
02 Std. Sol.	0.283	14:45
03 Room Air	0.000	14:46
04 Std. Sol.	0.283	14:46
05 Room Air	0.000	14:47
06 Std. Sol.	0.284	14:48
07 Room Air	0.000	14:48

08 Sim Temp = 34.0°C

Simul Ser No = 00000
Std Sol No = 000
County = 08 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

Linearity test 0.285 AC

Form 106-I8000

80-004937
28/33 KZ

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004937
Location = TOXL 8164.13.00 06/09
02/10/2010 14:50

Test	AC	Time
01 Room Air	0.000	14:50
02 *Subject Test	INT*	14:51
03 Room Air	0.000	14:51

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 668180
Cyl No = 54
Expiration Date = 09/21/2011
County = 08 Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Interferent check

Form 106-I8000

80-004937
29/33 1/42

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004937
Location = TOXL 8164.13.00 06/09
02/10/2010 14:52

Test	AC	Time
01 Diagnostic	OK	14:53
02 Room Air	RFI*	14:53
03 Room Air	0.000	14:53

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 668180
Cyl No = 54
Expiration Date = 09/21/2011
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

RFI check

Form 106-I8000

80004937
30/33 *FK*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004937
Location = TOXL 8164.13.00 06/09
02/10/2010 16:00

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:01
02 Std. Gas	0.080	16:01
03 Room Air	0.000	16:02
04 Std. Gas	0.080	16:02
05 Room Air	0.000	16:03
06 Std. Gas	0.080	16:03
07 Room Air	0.000	16:03

Lot No = 668180
Cyl No = 54
Exp Date = 09/21/2011
County = 08 Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-18000

80-004937
3/13/10

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004937
Location = TOXL 8164.13.00 06/09
02/10/2010 16:12

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:13
02 Std. Gas	0.080	16:13
03 Room Air	0.000	16:14
04 Std. Gas	0.079	16:14
05 Room Air	0.000	16:15
06 Std. Gas	0.080	16:15
07 Room Air	0.000	16:15

Lot No = 668180

Cyl No = 54

Exp Date = 09/21/2011

County = 08

Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080AC

Form 106-I8000

80-004937
32/33 KML

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004937
Location = TOXL 8164.13.00 06/09
02/10/2010 16:16

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:17
02 Std. Gas	0.080	16:17
03 Room Air	0.000	16:18
04 Std. Gas	0.079	16:18
05 Room Air	0.000	16:19
06 Std. Gas	0.080	16:19
07 Room Air	0.000	16:19

Lot No = 668180

Cyl No = 54

Exp Date = 09/21/2011

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

80-004937
33/33 *JK*