

**INTOXILYZER 5000 G and 8000 INSPECTION**

Serial No.: 80-004949

Location: TOXL

- |  | Check When Done |
|--|-----------------|
| A. Is the warm-up time less than 20 minutes?   | <u>✓</u>        |
| B. Is a three-pronged grounded outlet used?  | <u>✓</u>        |
| C. Is the breath tube heated?  | <u>✓</u>        |
| D. Is the diagnostic check complete?   | <u>✓</u>        |
| E. Is the time, date, and year correct (re-set if necessary)?  | <u>✓</u>        |
| F. Print test. (Attach test record.)   | <u>✓</u>        |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5144</u> Lot No. <u>    </u> Rep AC <u>0.019</u>   | <u>✓</u>        |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5187</u> Lot No. <u>    </u> Rep AC <u>0.285</u>                                | <u>✓</u>        |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.<br>(Attach test record.)<br>Sim SN <u>DR5114</u> Lot No. <u>    </u> Rep AC <u>0.100</u> | <u>✓</u>        |
| J. RFI check. Run CMS mode. Key radio on first room air.<br>(Attach test record.)  | <u>✓</u>        |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.<br>(Attach test record.)  |                 |

Lot No. <u>668180</u>	Cyl No. <u>54</u>	Exp. Date <u>09/21/2011</u>
Test 1 <u>0.079</u>	Test 4 <u>0.079</u>	Test 7 <u>0.080</u>
Test 2 <u>0.080</u>	Test 5 <u>0.080</u>	Test 8 <u>0.080</u>
Test 3 <u>0.080</u>	Test 6 <u>0.080</u>	Test 9 <u>0.080</u>
Average <u>0.080</u>		

L. Remarks/Maintenance record: Annual inspection.

10 Feb 2010  
Date  
15 Feb 2010  
Date

Deb Kashur  
Field Inspector's Signature  
Mary Dean  
Reviewed by

80-004949  
25/33 KA



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004949  
Location = TOXL      8164.13.00 06/09  
02/09/2010      16:17

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:18
02 Std. Sol.	0.017	16:18
03 Room Air	0.000	16:19
04 Std. Sol.	0.018	16:20
05 Room Air	0.000	16:20
06 Std. Sol.	0.019	16:21
07 Room Air	0.000	16:21

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Low AC*

*0.019 AC*

Form 106-I8000

80-004949  
27/33 VFL

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004949  
Location = TOXL      8164.13.00 06/09  
02/09/2010      16:34

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:35
02 Std. Sol.	0.285	16:36
03 Room Air	0.000	16:36
04 Std. Sol.	0.285	16:37
05 Room Air	0.000	16:38
06 Std. Sol.	0.286	16:38
07 Room Air	0.000	16:39

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

*Deb Keshur*

Operator Signature  
DEB KASHUR

Remarks:

*Linearity test 0.285 AC*

Form 106-18000

80-004949  
2/3/33 AK

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer                      Alcohol Analyzer  
North Dakota Model 8000                      SN 80-004949  
Location = TOXL                                      8164.13.00 06/09  
02/10/2010    08:39

Test	AC	Time
01 Room Air	0.000	08:40
02 *Subject Test	INT*	08:40
03 Room Air	0.000	08:41

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                                      Weight = 150  
Test = DUI    Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 668180  
Cyl No = 54  
Expiration Date = 09/21/2011  
County = 08    Oper No. = 777777

  
\_\_\_\_\_  
Operator Signature  
DEB KASHUR

Remarks:  


Form 106-18000

80-004949  
29/33 SK

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004949  
Location = TOXL      8164.13.00 06/09  
02/10/2010      08:37

Test	AC	Time
01 Room Air	RFI*	08:37
02 Room Air	0.000	08:38

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 668180  
Cyl No = 54  
Expiration Date = 09/21/2011  
County = 08      Oper No. = 777777

Deb Kashur

Operator Signature  
DEB KASHUR

Remarks:

*RFI test*

Form 106-I8000

80-004949  
30/33 KPL

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004949  
Location = TOXL      8164.13.00 06/09  
02/09/2010      15:06

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:06
02 Std. Gas	0.079	15:06
03 Room Air	0.000	15:07
04 Std. Gas	0.080	15:07
05 Room Air	0.000	15:08
06 Std. Gas	0.080	15:08
07 Room Air	0.000	15:09

Lot No = 668180  
Cyl No = 54  
Exp Date = 09/21/2011  
County = 08      Oper No. = 777777

Deb Kashur

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*

Form 106-I8000

80-004949  
3/33 JFL

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004949  
Location = TOXL      8164.13.00 06/09  
02/09/2010      15:46

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:47
02 Std. Gas	0.079	15:47
03 Room Air	0.000	15:48
04 Std. Gas	0.080	15:48
05 Room Air	0.000	15:49
06 Std. Gas	0.080	15:49
07 Room Air	0.000	15:50

Lot No = 668180  
Cyl No = 54  
Exp Date = 09/21/2011  
County = 08      Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*

Form 106-I8000

80-004949  
32/33 VFL

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004949  
Location = TOXL      8164.13.00 06/09  
02/09/2010      15:50

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:50
02 Std. Gas	0.080	15:51
03 Room Air	0.000	15:51
04 Std. Gas	0.080	15:52
05 Room Air	0.000	15:52
06 Std. Gas	0.080	15:52
07 Room Air	0.000	15:53

Lot No = 668180

Cyl No = 54

Exp Date = 09/21/2011

County = 08

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*

Form 106-I8000

80-004949  
33/33 *FK*