

Intoxilyzer 8000 Initial Inspection

| | |
|----------------------|------------------|
| Serial Number | 80-004946 |
|----------------------|------------------|

| | | | | | | | | |
|--|---|---|----------------|------------|----------------|----------------|------------|-------------------------------|
| <input checked="" type="checkbox"/> A. Power On | <input checked="" type="checkbox"/> 5 min. countdown | <input checked="" type="checkbox"/> 1 min. | | | | | | |
| <input checked="" type="checkbox"/> B. Security Checks. | <input checked="" type="checkbox"/> Level 1 | <input checked="" type="checkbox"/> Level 2 <input checked="" type="checkbox"/> Level 3 | | | | | | |
| <input checked="" type="checkbox"/> C. Gas Cylinder Setup | | | | | | | | |
| <input checked="" type="checkbox"/> D. Printer. Level 1 Func P | <input checked="" type="checkbox"/> Paper Feed | <input checked="" type="checkbox"/> Internal <input checked="" type="checkbox"/> External | | | | | | |
| <input checked="" type="checkbox"/> E. Display is working. | | | | | | | | |
| <input checked="" type="checkbox"/> F. Configure Start Test Switch. Level 3 Func S Sub T | | | | | | | | |
| <input checked="" type="checkbox"/> G. Setup Preliminary Data. Level 2 Func E | | | | | | | | |
| <input checked="" type="checkbox"/> H. Configure General Setup. Level 3 Func S Sub G. | | | | | | | | |
| <input checked="" type="checkbox"/> I. Configure Port. Level 3 Func S Sub N P U | | | | | | | | |
| <input checked="" type="checkbox"/> J. Check Camera. ABA Mode | <input checked="" type="checkbox"/> Oper ID | <input checked="" type="checkbox"/> ND Lic. <input checked="" type="checkbox"/> Hot Key "R" | | | | | | |
| <input checked="" type="checkbox"/> K. Diagnostic Test. Level 2 Func D | | | | | | | | |
| <input checked="" type="checkbox"/> L. Check Ser Num / Version Level 2 Func V | 80-004946 Serial No. 8164.13 Version 1362.14 Display 7 CPLD | | | | | | | |
| <input checked="" type="checkbox"/> M. Detector Voltages. Level 3 Func D Sub D | <table style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 33%; text-align: center;">3 <u>12905</u></td> <td style="width: 33%; text-align: center;">s <u>1</u></td> <td style="width: 33%; text-align: center;">C <u>47.06</u></td> </tr> <tr> <td style="text-align: center;">9 <u>12906</u></td> <td style="text-align: center;">s <u>1</u></td> <td style="text-align: center;">B <u>44</u> <u>0</u> RFI Ind.</td> </tr> </table> | | 3 <u>12905</u> | s <u>1</u> | C <u>47.06</u> | 9 <u>12906</u> | s <u>1</u> | B <u>44</u> <u>0</u> RFI Ind. |
| 3 <u>12905</u> | s <u>1</u> | C <u>47.06</u> | | | | | | |
| 9 <u>12906</u> | s <u>1</u> | B <u>44</u> <u>0</u> RFI Ind. | | | | | | |
| <input checked="" type="checkbox"/> N. Hot Keys | <input checked="" type="checkbox"/> F5 Tone <input checked="" type="checkbox"/> F6 Green LED <input checked="" type="checkbox"/> F7 Red LED <input checked="" type="checkbox"/> F9 Fan <input checked="" type="checkbox"/> F10 Pump <input checked="" type="checkbox"/> F11 Sol Output | | | | | | | |
| <input checked="" type="checkbox"/> O. RFI Dectector | | | | | | | | |
| <input checked="" type="checkbox"/> P. Atmospheric Monitor. Level 3 Func D Sub A. | P <u>1019</u> | CF <u>0.99</u> | | | | | | |

80-004946
6/24 FM

08/09

OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 Instrument Certification and Maintenance Unit

BrF-005-8

✓ Q. Flow Monitor. Level 3 Func D Sub F. R 221 F 0.00 L/s

✓ R. Temperature Monitor. Level 3 Func D Sub T
 C 47.0 A 45.0 I 0.00
 B 45.0 V 0

✓ S. Gas Tank Monitor. Level 3 Func D Sub G 181 psi Intox 180 psi Reg

✓ T. Configure Wet Bath Sim. Level 1 Func S

✓ U. Zero Alcohol Test Level 1 Func C
010329 Sim SN — Lot 0.00 Rep Value
0.000 AC 0.000 AC 0.000 AC

✓ V. Low AC Test. Level 1 Func C
DR5190 Sim SN — Lot 0.021 Rep Value
0.020 AC 0.019 AC 0.019 AC

✓ W. Linearity Test. Level 1 Func C
DR5113 Sim SN — Lot 0.285 Rep Value
0.267 AC 0.280 AC 0.281 AC
 M.P. 0.276

✓ X. Interferent Test. Level 1 Func B
 Acetone DR5186 Sim SN — Lot 0.05% Conc
✓ Display ✓ Printout
 Methanol DR3846 Sim SN — Lot 0.10% Conc
✓ Display ✓ Printout
 IPA DR3840 Sim SN — Lot 0.10% Conc
✓ Display ✓ Printout

✓ Y. Ambient Test 1000 Display 1000 Printout

✓ Z. Invalid Sample Test ✓ Display ✓ Printout

✓ AA. RFI Detector Test. CMS Mode ✓ Display ✓ Printout

✓ BB. Configure Dry Gas Standard. Level 1 Func S

✓ CC. Stability Test at 0.080AC. Level 3 Func T Sub S
0.0784 Ave. .0006 Std Dev 0.75 Rel Std Dev

✓ DD. Maintenance/Repairs Calibrate Optical Bench

11 February 2010
 Date
10 Feb 2010 ^{DK} 12 Feb 2010
 Date

Mary Pearson
 Signature
Deb Kasher
 Reviewed by

80-004946
 7/24 FM

08/09

TOXL
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-004946
 02/10/2010 10:49:27

Auto Calibration

pg 1 of 2

<<<<< 3um >>>>> <<<<< 9um >>>>>

Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1

| Sample | % Abs | (% Abs Ref) | % Abs | (% Abs Ref) |
|-------------|---------|-------------|--------|-------------|
| Sample #1 | 0.0160 | (-0.0070) | 0.0740 | (-0.0310) |
| Sample #2 | 0.0460 | (-0.0150) | 0.0760 | (-0.0220) |
| Sample #3 | -0.0050 | (0.0120) | 0.0740 | (-0.0260) |
| Sample #4 | 0.0070 | (-0.0050) | 0.0720 | (-0.0340) |
| Avg % Abs | 0.0160 | (-0.0027) | 0.0740 | (-0.0273) |
| STD DEV | 0.0267 | (0.0137) | 0.0020 | (0.0061) |
| REL STD DEV | 166.654 | (511.890) | 2.703 | (22.354) |

Solution = 0.021 g/210L or 0.1000 mg/l, Samples = 4, Discarded = 1

| Sample | % Abs | (% Abs Ref) | % Abs | (% Abs Ref) |
|-------------|--------|-------------|--------|-------------|
| Sample #1 | 0.3290 | (0.0120) | 0.7070 | (0.0050) |
| Sample #2 | 0.3580 | (0.0060) | 0.7060 | (0.0220) |
| Sample #3 | 0.3360 | (0.0200) | 0.7340 | (0.0140) |
| Sample #4 | 0.3590 | (0.0100) | 0.7220 | (0.0160) |
| Avg % Abs | 0.3510 | (0.0120) | 0.7207 | (0.0173) |
| STD DEV | 0.0130 | (0.0072) | 0.0140 | (0.0042) |
| REL STD DEV | 3.704 | (60.093) | 1.949 | (24.019) |

Solution = 0.041 g/210L or 0.1952 mg/l, Samples = 4, Discarded = 1

| Sample | % Abs | (% Abs Ref) | % Abs | (% Abs Ref) |
|-------------|--------|-------------|--------|-------------|
| Sample #1 | 0.7000 | (-0.0020) | 1.5150 | (-0.0070) |
| Sample #2 | 0.7130 | (0.0000) | 1.5130 | (0.0000) |
| Sample #3 | 0.7090 | (0.0040) | 1.4820 | (0.0180) |
| Sample #4 | 0.7340 | (0.0100) | 1.5120 | (-0.0010) |
| Avg % Abs | 0.7187 | (0.0047) | 1.5023 | (0.0057) |
| STD DEV | 0.0134 | (0.0050) | 0.0176 | (0.0107) |
| REL STD DEV | 1.869 | (107.855) | 1.173 | (188.694) |

Solution = 0.106 g/210L or 0.5048 mg/l, Samples = 4, Discarded = 1

| Sample | % Abs | (% Abs Ref) | % Abs | (% Abs Ref) |
|-------------|--------|-------------|--------|-------------|
| Sample #1 | 1.9720 | (-0.0250) | 3.7980 | (-0.0190) |
| Sample #2 | 1.9620 | (-0.0150) | 3.8130 | (-0.0230) |
| Sample #3 | 1.9460 | (0.0010) | 3.7870 | (-0.0110) |
| Sample #4 | 1.9680 | (-0.0060) | 3.8490 | (-0.0310) |
| Avg % Abs | 1.9587 | (-0.0067) | 3.8163 | (-0.0217) |
| STD DEV | 0.0114 | (0.0080) | 0.0311 | (0.0101) |
| REL STD DEV | 0.581 | (120.312) | 0.816 | (46.461) |

Solution = 0.285 g/210L or 1.3571 mg/l, Samples = 4, Discarded = 1

| Sample | % Abs | (% Abs Ref) | % Abs | (% Abs Ref) |
|-------------|--------|-------------|--------|-------------|
| Sample #1 | 5.1480 | (-0.0260) | 9.6530 | (-0.0160) |
| Sample #2 | 5.1520 | (-0.0090) | 9.6780 | (-0.0120) |
| Sample #3 | 5.1380 | (0.0070) | 9.6810 | (-0.0070) |
| Sample #4 | 5.1340 | (-0.0060) | 9.6300 | (-0.0100) |
| Avg % Abs | 5.1413 | (-0.0027) | 9.6630 | (-0.0097) |
| STD DEV | 0.0095 | (0.0085) | 0.0286 | (0.0025) |
| REL STD DEV | 0.184 | (318.934) | 0.296 | (26.034) |

*Calibration
 p1 of 2
 MAP*

80-004946
 8/24 15M

TOXL
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-004946
 02/10/2010 10:49:27

Auto Calibration

pg 2 of 2

<<<<< 3um >>>>>

 Zero Order Coef 43.21
 First Order Coef 2549.96
 Second Order Coef 15.60

<<<<< 9um >>>>>

 -23.10
 1290.19
 12.03

| Act (g/210L) | Fit (g/210L) | Residual (g/210L) |
|-----------------|-----------------|----------------------|
| 0.000 | 0.002 | -0.0018 |
| 0.021 | 0.020 | 0.0013 |
| 0.041 | 0.040 | 0.0014 |
| 0.106 | 0.107 | -0.0010 |
| 0.285 | 0.285 | 0.0001 |

| Act (g/210L) | Fit (g/210L) | Residual (g/210L) |
|-----------------|-----------------|----------------------|
| 0.000 | 0.002 | -0.0015 |
| 0.021 | 0.019 | 0.0018 |
| 0.041 | 0.041 | 0.0002 |
| 0.106 | 0.107 | -0.0006 |
| 0.285 | 0.285 | 0.0001 |

<<<<< 3um >>>>> <<<<< 9um >>>>>

 Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1
 Sample

| | | |
|-----------------------|-----------|-----------|
| Sample #1 | 3439.00 | 3338.00 |
| Sample #2 | 3412.00 | 3348.00 |
| Sample #3 | 3445.00 | 3348.00 |
| Sample #4 | 3342.00 | 3234.00 |
| Avg | 3399.6667 | 3310.0000 |
| STD DEV | 52.5959 | 65.8179 |
| REL STD DEV | 1.547 | 1.988 |
| H2O adjust (mg/l*10k) | 410 | 499 |

Atmospheric Pressure = 1018

*****CALIBRATION SUCCESSFUL*****

*Calibration
 prog
 MAP*

80-004946
 9/24 KM

Intoxilyzer Test Record and Checklist
NDQAG Crime Lab. Div., Bismarck, ND 58501

CMi, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = TOXL 8164.13.00 06/05
02/10/2010 10:21

***** Printer Test *****

12 char/line
abcdefghijkl
ABCDEFGHIJKL

16 char per line
abcdefghijklmnop
ABCDEFGHIJKLMN

21 chars per line
abcdefghijklmnopqrstu
ABCDEFGHIJKLMNQRSTU

24 characters per line
abcdefghijklmnopqrstuwx
ABCDEFGHIJKLMNQRSTUWX

32 characters per line
abcdefghijklmnopqrstuwx123456
ABCDEFGHIJKLMNQRSTUWXYZ123456

42 characters per line
abcdefghijklmnopqrstuwx1234567890abcdef
ABCDEFGHIJKLMNQRSTUWXYZ1234567890ABCDEF

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Menfull? No
of Print Copies? 1
Select Std (D/W/L)? Dry
Standard Value? 0.000
Standard Lot #? 12345678
Standard Cyl #? 1234567
Standard Expiration? 11/11/2011
Oper No? 107501

Flow Cal. Date: 01/18/2010
Slope 717
Intercept -728279

IR Calibration Date: 01/19/2010

| | 3Ln | 9Ln |
|--------------------|--------|--------|
| 0th Coef(*100): | -14511 | -21944 |
| 1st Coef(*100): | 272115 | 136613 |
| 2nd Coef(*100): | 1283 | 1052 |
| H2O adj(ng/l*10K): | 361 | 409 |

***** Test End *****

Margy Pearson

Operator Signature
MARGY PEARSON

Remarks:

Print

Form 106-18000

80-004946
10/24 KM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = TOXL 8164.13.00 06/09
02/10/2010 10:22

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 12345678
Standard Cyl #? 1234567
Standard Expiration? 11/11/2011
Oper No? 107501

Flow Cal. Date: 01/18/2010
Slope 717
Intercept -728279

IR Calibration Date: 01/19/2010

| | 3um | 9um |
|--------------------|--------|--------|
| 0th Coef(*100): | -14511 | -21944 |
| 1st Coef(*100): | 272115 | 138613 |
| 2nd Coef(*100): | 1283 | 1052 |
| H2O adj(mg/l*10k): | 361 | 409 |

***** Printer Test End *****



Operator Signature
MARGY PEARSON

Remarks:

Print


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = TOXL 8164.13.00 06/09
02/10/2010 10:42

| Test | AC | Time |
|------------------|-------|-------|
| 01 Room Air | 0.000 | 10:43 |
| 02 *Subject Test | REF* | 10:43 |
| 03 Room Air | 0.000 | 10:44 |

*Subject Test Refused

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 11/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS821456
Lot No = 659358
Cyl No = 38
Expiration Date = 09/03/2011
County = 08 Oper No. = 107501



Operator Signature
MARGY PEARSON

Remarks:

Camera / Refused Hot "R"

Form 106-I8000

80004946
12/24/10


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = TOXL 8164.13.00 06/09
02/10/2010 10:44

DIAGNOSTIC

| | |
|---------------------------|------|
| Voltage/Current Test | Pass |
| RAM Test | Pass |
| EEPROM Checksum Test | Pass |
| Real Time Clock Test | Pass |
| DSP Test | Pass |
| Analytical Stability Test | Pass |
| Modem Test | Pass |
| Temp Regulation Test | Pass |

Oper No. = 107501



Operator Signature
MARGY PEARSON

Remarks:

DIAGNOSTIC

Form 106-I8000

80-004946
13/24 KM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = TOXL 8164.13.00 06/09
02/10/2010 11:49

WET CAL CHECK

| Test | AC | Time |
|--------------|-------|-------|
| 01 Room Air | 0.000 | 11:50 |
| 02 Std. Sol. | 0.000 | 11:50 |
| 03 Room Air | 0.000 | 11:51 |
| 04 Std. Sol. | 0.000 | 11:52 |
| 05 Room Air | 0.000 | 11:52 |
| 06 Std. Sol. | 0.000 | 11:53 |
| 07 Room Air | 0.000 | 11:53 |


08 Sim Temp = 34.0°C

Simul Ser No = 0000000

Std Sol No = 000000

County = 08

Oper No. = 107501



Operator Signature
MARGY PEARSON

Remarks:

Zero 0.00AC

Form 106-18000

80-004946
1/24 VM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = TOXL 8164.13.00 06/09
02/10/2010 11:54

WET CAL CHECK

| Test | AC | Time |
|--------------|-------|-------|
| 01 Room Air | 0.000 | 11:55 |
| 02 Std. Sol. | 0.020 | 11:55 |
| 03 Room Air | 0.000 | 11:56 |
| 04 Std. Sol. | 0.019 | 11:57 |
| 05 Room Air | 0.000 | 11:57 |
| 06 Std. Sol. | 0.019 | 11:58 |
| 07 Room Air | 0.000 | 11:58 |


08 Sim Temp = 34.0°C

Simul Ser No = 0000000

Std Sol No = 000000

County = 08

Oper No. = 107501



Operator Signature
MARGY PEARSON

Remarks:

Low 0.021 AC

Form 106-18000

80-004946
1/24 PM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = TOXL 8164.13.00 06/09
02/10/2010 12:05

WET CAL CHECK

| Test | AC | Time |
|--------------|-------|-------|
| 01 Room Air | 0.000 | 12:06 |
| 02 Std. Sol. | 0.276 | 12:07 |
| 03 Room Air | 0.000 | 12:07 |
| 04 Std. Sol. | 0.280 | 12:08 |
| 05 Room Air | 0.000 | 12:09 |
| 06 Std. Sol. | 0.281 | 12:09 |
| 07 Room Air | 0.000 | 12:10 |

08 Sim Temp = 34.0°C

Simul Ser No = 0000000

Std Sol No = 000000

County = 08

Oper No. = 107501

Margy Pearson

Operator Signature
MARGY PEARSON

Remarks:

Linearity 0.285 AC

Form 106-I8000

*80-004946
10/24/10*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = TOXL 8164.13.00 06/09
02/11/2010 15:30

| Test | AC | Time |
|------------------|-------|-------|
| 01 Room Air | 0.000 | 15:31 |
| 02 *Subject Test | INT* | 15:32 |
| 03 Room Air | 0.000 | 15:32 |

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 11/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS821456
Lot No = 659358
Cyl No = 15
Expiration Date = 09/03/2011
County = 08 Oper No. = 107501



Operator Signature
MARGY PEARSON

Remarks:

Interfer Acetone

Form 106-I8000

80004946
17/24 PM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = TOXL 8164.13.00 06/09
02/11/2010 15:32

| Test | AC | Time |
|------------------|-------|-------|
| 01 Room Air | 0.000 | 15:33 |
| 02 *Subject Test | INT* | 15:34 |
| 03 Room Air | 0.000 | 15:34 |

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 11/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS821456
Lot No = 659358
Cyl No = 15
Expiration Date = 09/03/2011
County = 08 Oper No. = 107501



Operator Signature
MARGY PEARSON

Remarks:

MeOH Interfer

Form 106-I8000

80-004946
1/24/11


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = TOXL 8164.13.00 06/09
02/11/2010 15:57

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 15:58 |
| 02 Std. Gas | 0.079 | 15:58 |
| 03 Room Air | 0.000 | 15:59 |
| 04 Std. Gas | 0.078 | 15:59 |
| 05 Room Air | 0.000 | 16:00 |
| 06 Std. Gas | 0.079 | 16:00 |
| 07 Room Air | 0.000 | 16:01 |

Lot No = 659358
Cyl No = 15
Exp Date = 09/03/2011
County = 08 Oper No. = 107501



Operator Signature
MARGY PEARSON

Remarks:

IPA Interferent

Form 106-I8000

*80-004946
19/24 PM*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = TOXL 8164.13.00 06/09
02/11/2010 16:14

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | AMB* | 16:15 |
| 02 Room Air | 0.000 | 16:15 |

*Invalid Test
Check Ambient Conditions

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 11/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS821456
Lot No = 659358
Cyl No = 15
Expiration Date = 09/03/2011
County = 08 Oper No. = 107501



Operator Signature
MARGY PEARSON

Remarks:

Ambient Fail
Form 106-I8000

80-004946
2/24/12


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = TOXL 8164.13.00 06/09
02/11/2010 16:16

| Test | AC | Time |
|-------------------|-------|-------|
| 01 Room Air | 0.000 | 16:17 |
| 02 Invalid Sample | X.XXX | 16:17 |
| 03 Room Air | 0.000 | 16:18 |

*Invalid Test - Mouth Alcohol

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 11/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS821456
Lot No = 659358
Cyl No = 15
Expiration Date = 09/03/2011
County = 08 Oper No. = 107501



Operator Signature
MARGY PEARSON

Remarks:

Invalid Sample

Form 106-I8000

80-004946
2/24 FM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = TOXL 8164.13.00 06/09
02/11/2010 16:10

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | RFI* | 16:11 |
| 02 Room Air | 0.000 | 16:11 |

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 11/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS821456
Lot No = 659358
Cyl No = 15
Expiration Date = 09/03/2011
County = 08 Oper No. = 107501



Operator Signature
MARGY PEARSON

Remarks:

RFI

Form 106-18000

80-004946
22/24 FM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = TOXL 8164.13.00 06/09
02/11/2010 14:25

| Test | AC | Time |
|---------------------|--------|-------|
| 01 Room Air | 0.000 | 14:25 |
| 02 Std. Gas | 0.079 | 14:26 |
| 03 Room Air | 0.000 | 14:26 |
| 04 Std. Gas | 0.079 | 14:27 |
| 05 Room Air | 0.000 | 14:27 |
| 06 Std. Gas | 0.079 | 14:28 |
| 07 Room Air | 0.000 | 14:28 |
| 08 Std. Gas | 0.078 | 14:28 |
| 09 Room Air | 0.000 | 14:29 |
| 10 Std. Gas | 0.078 | 14:29 |
| 11 Room Air | 0.000 | 14:30 |
| 12 Std. Gas | 0.078 | 14:30 |
| 13 Room Air | 0.000 | 14:31 |
| 14 Std. Gas | 0.077 | 14:31 |
| 15 Room Air | 0.000 | 14:32 |
| 16 Std. Gas | 0.078 | 14:32 |
| 17 Room Air | 0.000 | 14:32 |
| 18 Std. Gas | 0.078 | 14:33 |
| 19 Room Air | 0.000 | 14:33 |
| 20 Std. Gas | 0.078 | 14:34 |
| 21 Room Air | 0.000 | 14:34 |
| 22 Std. Gas | 0.079 | 14:34 |
| 23 Room Air | 0.000 | 14:35 |
| 24 Std. Gas | 0.079 | 14:35 |
| 25 Room Air | 0.000 | 14:36 |
| 26 Std. Gas | 0.078 | 14:36 |
| 27 Room Air | 0.000 | 14:37 |
| 28 Std. Gas | 0.079 | 14:37 |
| 29 Room Air | 0.000 | 14:38 |
| 30 Std. Gas | 0.079 | 14:38 |
| 31 Room Air | 0.000 | 14:38 |
| 32 Std. Gas | 0.078 | 14:39 |
| 33 Room Air | 0.000 | 14:39 |
| 34 Std. Gas | 0.078 | 14:40 |
| 35 Room Air | 0.000 | 14:40 |
| 36 Std. Gas | 0.078 | 14:41 |
| 37 Room Air | 0.000 | 14:41 |
| 38 Std. Gas | 0.078 | 14:41 |
| 39 Room Air | 0.000 | 14:42 |
| 40 Std. Gas | 0.079 | 14:42 |
| 41 Room Air | 0.000 | 14:43 |
| 42 Cal Check Stats: | | |
| 43 Average | 0.0784 | |
| 44 Std Dev | 0.0006 | |
| 45 Rel Std Dev | 0.75% | |

Stability Test
P1 06 2
MGP

80-004946
23/24 174

Operator Signature
MARGY PEARSON

Remarks:

Form 106-I8000

LOT NO 659359D
Cyl #15
9/3/2011

Stability Test
P 2010
MAP

80-004946
29/24 FM