



OFFICE OF ATTORNEY GENERAL
 Crime Laboratory Division
 2641 East Main Avenue
 P.O. Box 937
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Mark W. McDonald

Location: Valley City Police Dept.

Intoxilyzer Serial No.: 80-004201

Lot Number: ~~24~~ 659358

Cylinder Number: 24

Expiration Date: 09/03/11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 079 %

Test 2: 0. 080 %

Test 3: 0. 079 %

(ATTACH TEST RECORD)

Deb Kashur
 Chemical Test Operator Signature
 Deb Kashur

02/11/10
 Date Cylinder Changed
 18 Feb 2010

Form 105-G
 Toxicology Laboratory
 Crime Laboratory Division

80-004201
 6/8/09 KM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004201
Location = VALL 8164.13.00 06/09
02/11/2010 16:39

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:40
02 Std. Gas	0.079	16:40
03 Room Air	0.000	16:41
04 Std. Gas	0.080	16:41
05 Room Air	0.000	16:42
06 Std. Gas	0.079	16:42
07 Room Air	0.000	16:43

Lot No = 659358
Cyl No = 24
Exp Date = 09/03/2011
County = 02 Oper No. = 119502



Operator Signature
MARK W. MCDONALD

Remarks:

Form 106-I8000

80-004201
09/09 KAL