

### BREATH ALCOHOL INSTRUMENT CHAIN OF CUSTODY WORKSHEET

Instrument: 8000 Serial Number: 80-004189

Transported Out of the Crime Laboratory:

Agency Being Assigned Instrument: \_\_\_\_\_

Name/Agency of Person Transporting Instrument:  
\_\_\_\_\_  
(Print Name) \_\_\_\_\_ (Print Agency)

Signature of Person Receiving Instrument: \_\_\_\_\_

Date Receiving Instrument: \_\_\_\_/\_\_\_\_/\_\_\_\_ From: \_\_\_\_\_

Or Shipped Via:  
USPS \_\_\_\_ UPS \_\_\_\_ FedEx \_\_\_\_ Other \_\_\_\_\_

Date Sending Instrument: \_\_\_\_/\_\_\_\_/\_\_\_\_ Air Bill No. \_\_\_\_\_

Complete Air Bill (Identify Instrument/Serial Number) and attach a copy to this worksheet.

Transported to the Crime Laboratory:

Agency Returning Instrument: CMI

Name/Agency of Person Returning Instrument:  
\_\_\_\_\_  
(Print Name) \_\_\_\_\_ (Print Agency)

Signature of Person Returning Instrument: Deb Kashur<sup>DK</sup>

Date Returning Instrument: 11, Feb, 2011 Received By: Deb Kashur

Or Received Via:  
USPS \_\_\_\_ UPS \_\_\_\_ FedEx  Other \_\_\_\_\_

Date Receiving Instrument: \_\_\_\_/\_\_\_\_/\_\_\_\_ Air Bill No. \_\_\_\_\_

Attach shipping paperwork (if any) to this worksheet.

Instrument Transfer Information Entered Into the Equipment Database:

Date of Action 11, Feb, 2011 By: Deb Kashur

80-004189  
P. 95 of 97  
FM



# SERVICE WORK ORDER

DATE 1-18-11  
 RCVD VIA FEDX

316 E. 9<sup>th</sup> St., Owensboro KY 42303  
 Phone 866-835-0690  
 Fax: 270-685-6268

325772

**BILL TO:**

ATTN:

CITY STATE ZIP

CUSTOMER NO. 585013

PHONE

**SHIP TO:** Attorney Generals Office

ATTN: Deb Kashur

Crime Lab Division

2641 E. Main Ave.

CITY STATE ZIP  
Bismarck ND 58501

PHONE 701-328-6159

FAX 701-328-6185

MODEL # 8000 SERIAL # 80-004189 BILL CODE ES

EXTRA PARTS RCVD Bt. gas box, battery EW

DESCRIPTION OF PROBLEM

**Infrared / Fuel Cell**  
 PRELIMINARY TEST INFORMATION

BATTERY- EXT 0.00 V; INT 3.02 V

VERIFIED PROBLEM? YES

WILL NOT AUTORANGE

WORK PERFORMED:  
Reset gain, checked out, sent to cal - final.

SEAL # 03784

**Miscellaneous**  
 PRELIMINARY TEST INFORMATION

WORK PERFORMED:

**PARTS USED**

QTY	PART NO.	DESCRIPTION

REPAIR <u>L.H.</u>	CAL. ADJ. <u>KR</u>	CALIBRATION <u>ICR</u>	FINAL <u>KR</u>
1/2 HOURS <u>12</u>	HOURS <u>.7</u>	HOURS <u>1</u>	HOURS <u>1</u>
DATE <u>01/18/11</u>	DATE <u>01/21/11</u>	DATE <u>01/31/11</u>	DATE <u>02/04/11</u>

SHIPPING STICKER HERE

SPECIAL SHIPPING INSTRUCTIONS

80-004189  
P. 96 of 97  
KFM

# Certificate of Calibration

## **Intoxilyzer**® 8000

*This is to certify that the calibration of **Intoxilyzer**® serial number 80-004189, manufactured by CMI Inc., a subsidiary of MPD, Inc. of Owensboro, Kentucky, was tested and found to be in compliance with the National Highway Traffic Safety Administration Standard for Devices to Measure Breath Alcohol (F.R., No. 179 48705-48710 Sept. 17, 1993) and with International Recommendation OIML R126 Edition 1998 (E). Calibration solutions are traceable to NIST standard material 1828.*

Date Feb 4 2011 Signed Kristi Robinson  
Technician

**CMI** INC.

80-004189  
P. 97 of 97  
FM

316 East 9<sup>th</sup> Street  
Owensboro, KY 42303  
USA

Part No. 650519 CMI 7/9/09