

BREATH ALCOHOL INSTRUMENT CHAIN OF CUSTODY WORKSHEET

Instrument: 8000 Serial Number: 80-004198

Transported Out of the Crime Laboratory:

Agency Being Assigned Instrument: _____

Name/Agency of Person Transporting Instrument:

(Print Name) _____ (Print Agency)

Signature of Person Receiving Instrument: _____

Date Receiving Instrument: ____/____/____ From: _____

Or Shipped Via:
USPS ____ UPS ____ FedEx ____ Other _____

Date Sending Instrument: ____/____/____ Air Bill No. _____

Complete Air Bill (Identify Instrument/Serial Number) and attach a copy to this worksheet.

Transported to the Crime Laboratory:

Agency Returning Instrument: CM1

Name/Agency of Person Returning Instrument:

(Print Name) _____ (Print Agency)

Signature of Person Returning Instrument: _____

Date Returning Instrument: 11 Feb 2011 Received By: Deb Kashur

Or Received Via:
USPS ____ UPS ____ FedEx Other _____

Date Receiving Instrument: ____/____/____ Air Bill No. _____

Attach shipping paperwork (if any) to this worksheet.

Instrument Transfer Information Entered Into the Equipment Database:

Date of Action 11 Feb 2011 By: Deb Kashur

80-004198
P. 113 of 115
jm



SERVICE WORK ORDER

DATE	1-18-11
RCVD VIA	FEDX

316 E. 9th St., Owensboro KY 42303
 Phone 866-835-0690
 Fax: 270-685-6268

325771

BILL TO:
ATTN:
CITY STATE ZIP
CUSTOMER NO. 585013
PHONE

SHIP TO: Attorney Generals Office
ATTN: Deb Kashur
Crime Lab Division
2041 E. Main Ave.
CITY STATE ZIP Bismarck ND 58501
PHONE 701-328-6159
FAX 701-328-6185

MODEL # 8000	SERIAL # 50-004198	BILL CODE E
--------------	--------------------	-------------

EXTRA PARTS RCVD: BT, gas box, battery

DESCRIPTION OF PROBLEM: STUCK IN NOT READY

Infrared / Fuel Cell
 PRELIMINARY TEST INFORMATION
 BATTERY-EXT 00 V; INT 3.03 V

STUCK IN NOT READY. VERIFIED PROBLEM? YES

WORK PERFORMED: Repaired bad connection on the cell temp sensor plug. Changed battery. Checked out - sent to cal final.

not
SEAL# 23787

Miscellaneous
 PRELIMINARY TEST INFORMATION

WORK PERFORMED:

PARTS USED		
QTY	PART NO.	DESCRIPTION
/		

REPAIR L.H.	CAL. ADJ. KR	CALIBRATION KR	FINAL KR
1/2 HOURS 13	HOURS .7	HOURS 1	HOURS 1
DATE 04/20/11	DATE 2/3/11	DATE 2/3/11	DATE 2/3/11

SHIPPING STICKER HERE

SPECIAL SHIPPING INSTRUCTIONS

80-004198
P. 114 of 115
KRN

Certificate of Calibration

Intoxilyzer® 8000

*This is to certify that the calibration of **Intoxilyzer**® serial number 80-004198, manufactured by CMI Inc., a subsidiary of MPD, Inc. of Owensboro, Kentucky, was tested and found to be in compliance with the National Highway Traffic Safety Administration Standard for Devices to Measure Breath Alcohol (F.R., No. 179 48705-48710 Sept. 17, 1993) and with International Recommendation OIML R126 Edition 1998 (E). Calibration solutions are traceable to NIST standard material 1828.*

Date Feb 4 2011

Signed Kristi Robinson

Technician

CMI INC.

80-004198
p. 115 of 115
FM

316 East 9th Street
Owensboro, KY 42303
USA

Part No. 650519 CMI 7/9/09