

INTOXILYZER 5000 G and 8000 INSPECTION

Serial No.: 80-004950

Location: TOXL

Check When Done

- A. Is the warm-up time less than 20 minutes? ✓
- B. Is a three-pronged grounded outlet used? ✓
- C. Is the breath tube heated? ✓
- D. Is the diagnostic check complete? ✓
- E. Is the time, date, and year correct (re-set if necessary)? ✓
- F. Print test. (Attach test record.) ✓
- G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5144 Lot No. — Rep AC 0.019 ✓
- H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR5187 Lot No. — Rep AC 0.285 ✓
- I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR5114 Lot No. — Rep AC 0.100 ✓
- J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
- K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 6608180 Cyl No. 54 Exp. Date 09/21/2011

Test 1 0.079 Test 4 0.080 Test 7 0.081

Test 2 0.080 Test 5 0.080 Test 8 0.079

Test 3 0.080 Test 6 0.080 Test 9 0.080

Average 0.080

L. Remarks/Maintenance record: Annual inspection.

12 Feb 2010

Date

15 Feb 2010

Date

Deb Kashner
Field Inspector's Signature

Mary Plann
Reviewed by

80-004950
25/33 km

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004950
Location = TOXL 8164.13.00 06/09
02/12/2010 09:05

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:06
02 Std. Sol.	0.018	09:07
03 Room Air	0.000	09:07
04 Std. Sol.	0.019	09:08
05 Room Air	0.000	09:08
06 Std. Sol.	0.018	09:09
07 Room Air	0.000	09:10

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Low AC 0.018 AC

Form 106-I8000

80-004950
27/33 JAK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004950
Location = TOXL 8164.13.00 06/09
02/12/2010 09:30

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:31
02 Std. Sol.	0.285	09:31
03 Room Air	0.000	09:32
04 Std. Sol.	0.286	09:33
05 Room Air	0.000	09:33
06 Std. Sol.	0.287	09:34
07 Room Air	0.000	09:34

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Linearity test 0.285 AC

Form 106-I8000

80-004950
2/13/14

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004950
Location = TOXL 8164.13.00 06/09
02/12/2010 09:58

Test	AC	Time
01 Room Air	0.000	09:59
02 *Subject Test	INT*	10:00
03 Room Air	0.000	10:00

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 54
Cyl No = 668180
Expiration Date = 09/21/2011
County = 08 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

Interferent check

Form 106-I8000

80-004950
29/33 KAH

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004950
Location = TOXL 8164.13.00 06/09
02/12/2010 10:14

Test	AC	Time
01 Diagnostic	OK	10:15
02 Room Air	RFI*	10:15
03 Room Air	0.000	10:15

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 54
Cyl No = 668180
Expiration Date = 09/21/2011
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

RFI test

Form 106-I8000

80-004950
30/33 KTH

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004950
Location = TOXL 8164.13.00 06/09
02/12/2010 08:31

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:32
02 Std. Gas	0.079	08:32
03 Room Air	0.000	08:33
04 Std. Gas	0.080	08:33
05 Room Air	0.000	08:34
06 Std. Gas	0.080	08:34
07 Room Air	0.000	08:34

Lot No = 668180
Cyl No = 54
Exp Date = 09/21/2011
County = 08 Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

80-004950
31/33 KH

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004950
Location = TOXL 8164.13.00 06/09
02/12/2010 10:22

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:23
02 Std. Gas	0.080	10:23
03 Room Air	0.000	10:23
04 Std. Gas	0.080	10:24
05 Room Air	0.000	10:24
06 Std. Gas	0.080	10:25
07 Room Air	0.000	10:25

Lot No = 54
Cyl No = 668180
Exp Date = 09/21/2011
County = 08 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

calibration check 0.080 AC

Form 106-I8000

80-004950
32/33 YAK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004950
Location = TOXL 8164.13.00 06/09
02/12/2010 10:42

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:42
02 Std. Gas	0.081	10:43
03 Room Air	0.000	10:43
04 Std. Gas	0.079	10:43
05 Room Air	0.000	10:44
06 Std. Gas	0.080	10:44
07 Room Air	0.000	10:45

Lot No = 54
Cyl No = 668180
Exp Date = 09/21/2011
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

80-004950
33/33 1/2