



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59281 (1-2010)

Serial Number: 80-004949 Instrument Location: TRAILL County S.O.

Reason for Install/Repair:

Install After Lab Repair/Inspection

Other (Specify) Replace Intoxilyzer 8000

Check When Done:

- A. Surge Protector Installed/Properly Grounded.
- B. Telephone Line Connected to Intoxilyzer® 8000.
- C. Breath Tube Heated.
- D. Review/Enter Preliminary Data Entry (Level 2, Function E).
- E. Review/Enter Gas Setup (Level 1, Function S).
- F. Print Test (Level 1, Function P). (Attach Test Record.)
- G. ACA Test (Level 1, Function C). (Attach Test Record.)
- H. RFI Test (Use CMS Mode-First Room Air). (Attach Test Record.)
- I. Repair and/or Maintenance Performed: _____

Ray Weber
 Field Inspector Signature

2/16/2011
 Date

Deb Kashur
 Reviewed By

01 Mar 2011
 Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004949
Location = HILL 8164.13.00 06/09
02/16/2011 17:59

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNQPQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNQPQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 15410080A1
Standard Cyl #? 15
Standard Expiration? 07/01/2012
Oper No? 058405

Flow Cal. Date: 09/24/2010
Slope 784
Intercept -862571

IR Calibration Date: 09/24/2010
 3um 9um

0th Coef(*100): -23764 -22121
1st Coef(*100): 267615 132337
2nd Coef(*100): 1069 992
H2O adj(mg/l*10k): 531 480

***** Printer Test End *****


Operator Signature
RAY G. WEBER

Remarks:


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004949
Location = HILL 8164.13.00 06/09
02/16/2011 18:01

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:02
02 Std. Gas	0.079	18:02
03 Room Air	0.000	18:03
04 Std. Gas	0.079	18:03
05 Room Air	0.000	18:03
06 Std. Gas	0.079	18:04
07 Room Air	0.000	18:04

Lot No = 15410080A1
Cyl No = 15
Exp Date = 07/01/2012
County = 49 Oper No. = 058405



Operator Signature
RAY G. WEBER

Remarks:

Form 106-I8000

80-004949
p. 92 of 93
km

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004949
Location = HILL 8164.13.00 06/09
02/16/2011 18:23

Test	AC	Time
01 Room Air	RFI*	18:29
02 Room Air	0.000	18:29

*Invalid Test
Inhibited - RFI

Sub Name = TEST, TEST G
Sub DOB = 02/15/2011
Sub Sex = Unknown Weight = NA
Test = OTH Cit = NA
Dr. Lic. = ND/NA
Lot No = 15410080A1
Cyl No = 15
Expiration Date = 07/01/2012
County = 49 Oper No. = 058405


Operator Signature
RAY G. WEBER

Remarks:

Form 106-I8000

80-004949
p. 93 of 93
RM