



ETHANOL BREATH STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Ray Weher</u>		
Location: <u>Trail County S.O.</u>		
Intoxilyzer® Serial Number: <u>80-004949</u>		
Gas Lot Number: <u>15410080A1</u>	Gas Cylinder Number: <u>#15</u>	Expiration Date: <u>07/01/2012</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: 0. <u>080</u> % Test 2: 0. <u>080</u> % Test 3: 0. <u>080</u> %		
(ATTACH TEST RECORD)		

Ray Weher
 Chemical Test Operator Signature

2/16/2011
 Date

Deb Kashur
 Reviewed By

01 Mar 2011
 Date

80-004949
 P. 88 of 89
 KM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004949
Location = HILL 8164.13.00 06/09
02/16/2011 18:41

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:42
02 Std. Gas	0.080	18:43
03 Room Air	0.000	18:43
04 Std. Gas	0.080	18:44
05 Room Air	0.000	18:44
06 Std. Gas	0.080	18:45
07 Room Air	0.000	18:45


Lot No = 15410080A1

Cyl No = 15

Exp Date = 07/01/2012

County = 49

Oper No. = 058405



Operator Signature
RAY WEBER

Remarks:

Form 106-I8000

80-004949
P. 89/89
JW