



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>William Clark</u>		
Location: <u>MCRA7 AFB</u>		
Intoxilyzer® Serial Number: <u>80-004936</u>		
Lot Number: <u>659358</u>	Cylinder Number: <u>49</u>	Expiration Date: <u>9-3-11</u>
<input checked="" type="checkbox"/> New Cylinder Installed	<input type="checkbox"/> Cylinder not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.077</u> %		
Test 2: <u>0.078</u> %		
Test 3: <u>0.079</u> %		
(ATTACH TEST RECORD)		

William Clark

Chemical Test Operator Signature

17 FEB 2010

Date

Deb Kashner

Reviewed By

10 May 2010

Date


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004936
Location = MAFB 8164.13.00 06/09
02/17/2010 08:24

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:25
02 Std. Gas	0.077	08:25
03 Room Air	0.000	08:26
04 Std. Gas	0.078	08:26
05 Room Air	0.000	08:27
06 Std. Gas	0.079	08:27
07 Room Air	0.000	08:28

Lot No = 659358
Cyl No = 49
Exp Date = 09/03/2011
County = 51 Oper No. = 020809



Operator Signature
WILLIAM CLARK

Remarks:

Form 106-I8000

80-004936
P. 34/36 KA