

### INTOXILYZER® 8000 INITIAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006497

#### A. GENERAL Setup and Checks:

- Power On  
 5 min. countdown during initial warm-up  
 1 min. countdown to begin test after standby mode
- Display Intensity (Ctrl + ↑ or ↓)
3. Password Security:  
 Level 1 (Operator)     Level 2 (Field Inspector)     Level 3 (Lab Personnel)
- Printer Setup (Level 3,S,P)
- Forms Load (CMI, Inc., CoBrA Software)

#### B. LEVEL 1 Setup and Checks:

- Internal Printer Test and Paper Feed (Level 1,P). Attach record.
- External Printer Test (Level 1,P). Attach record.
3. Barcode Reader (Level 1,B):  
 Operator ID     ND License     R key: Refusal. Attach record.

#### C. LEVEL 2 Setup and Checks:

- Continuous Air Blank (Level 2,A)
- Diagnostic (Level 2,D). Attach record.
- Prelim Data Entry (Level 2,E)
4. Ser Number/Version Info (Level 2,V):  
Inst Ser#: 80-006497    DSP Ver: 1362.14  
Core Ver: 8164.13    CPLD Ver: 7

#### D. LEVEL 3 Setup and Checks:

- DVM Monitor (Level 3,D,D): 3: 12454 s: 1 C: 47.06  
9: 13028 s: 2 B: 45 0  
(RFI indicator with keyed radio): 1
2. Function (F) Keys: :  F5 – Tone                       F9 - Fan  
 F6 – Green LED                       F10 – Pump  
 F7 – Red LED                          F11 – Solenoid
3. Atmospheric Monitor (Level 3,D,A): Atmo Sensor, P: 955 CF: 1.05
4. Flow Monitor (Level 3,D,F): Flow Sensor, R: 215 F: 0.000 L/s

5. Temperature Monitor (Level 3,D,T): C: 47.0 A: 48.5 I: 0.0  
B: 44.9 V: 0.0
6. Tank Monitor (Level 3,D,G): P (display): 418 psi P (gauge): ≈420 psi
7.  General Setup (Level 3,S,G)
8.  Configure Start Test (Level 3,S,T)
9.  Network (Level 3,S,N)

**E. TESTS (attach all test records):**

**Configure wet bath simulator standards before running tests (Level 1,S):**

1. Zero AC (Level 1,C):  
Sim ser #: DR3375 Lot #: N/A Milli-Q Exp date: N/A  
Known value: 0.000 AC  
Results: 0.000 AC 0.000 AC 0.000 AC
2. Low AC (Level 1,C):  
Sim ser #: DR3374 Lot #: 14240 Exp date: 10/20/16  
Known value: 0.020 AC  
Results: 0.019 AC 0.019 AC 0.019 AC
3. High AC (Level 1,C):  
Sim ser #: DR711 Lot #: 201408C Exp date: 08/12/16  
Known value: 0.300 AC  
Results: 0.297 AC 0.295 AC 0.296 AC

**Configure dry gas standard before running tests (Level 1,S):**

4. Ethanol + 0.05% Acetone (Level 1,B):  
Sim ser #: DR5144 Lot #: ICS4 Exp date: 09/09/17  
Known value: 0.10 AC + 0.05% acetone  
 Display reads "Interferent Detect"
5. Acetone (Level 1,B):  
Sim ser #: DR3455 Lot #: Acet Exp date: 09/29/17  
Known value: 0.05% acetone  
 Display reads "Interferent Detect"
6. Methanol (Level 1,B):  
Sim ser #: DR5189 Lot #: Met 2 Exp date: 09/29/17  
Known value: 0.10% methanol  
 Display reads "Interferent Detect"
7. Isopropanol (Level 1,B):  
Sim ser #: DR3454 Lot #: IPA 2 Exp date: 09/29/17  
Known value: 0.10% IPA  
 Display reads "Interferent Detect"

- 8.  Ambient Fail (Level 1,B): Display and test record read correct.
- 9.  Invalid Sample (Level 1,B): Display and test record read correct.
- 10.  Sequence Aborted (Level 1,B): Display and test record read correct.
- 11.  ABA (Level 1,B): Test record reads correct.
- 12.  Calibration Check Out of Tolerance (CMS mode): Display and test record read correct.
- 13.  RFI Detected (B,C, or CMS mode): Display and test record read correct.
- 14.  Tank Pressure Below Minimum (CMS mode): Display reads correct.
- 15.  Deficient Samples (CMS mode): Display and test record read correct.
- 16.  Improper Sample (CMS mode): Display and test record read correct.
- 17.  Difference Too Great (CMS mode): Display and test record read correct.
- 18.  CMS (CMS mode): Test record reads correct.
- 19. Stability Test (Level 3,T,S):

Ave: 0.0815 Std Dev: 0.0005 Rel Std Dev: 0.6370

Maintenance/Repairs/Notes: \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

Instrument is acceptable to be used in the field.

Deb Shanaver  
Field Inspector's Signature

18 Feb 2016  
Date

Charles E. Eden  
Reviewed by

18 FEB. 16  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer  
North Dakota Model 8000 SN 80-006497  
Location = TOXL 8164.13.00 06/09  
02/08/2016 09:10

\*\*\*\*\* Printer Test \*\*\*\*\*

12 char/line  
abcdefghijkl  
ABCDEFGHIJKL

16 char per line  
abcdefghijklmnop  
ABCDEFGHIJKLMN

21 chars per line  
abcdefghijklmnopqrstu  
ABCDEFGHIJKLMN

24 characters per line  
abcdefghijklmnopqrstuwx  
ABCDEFGHIJKLMN

32 characters per line  
abcdefghijklmnopqrstuwx123456  
ABCDEFGHIJKLMN

42 characters per line  
abcdefghijklmnopqrstuwx1234567890abcdef  
ABCDEFGHIJKLMN

Current Instrument Setup

Data Entry Mode: Enabled  
Start Test Sequence: DABACABA  
Display Prelim Rslt? Yes  
Display Third Digit? Yes  
Inhib Printer(Y/N)? No  
Display Volume? No  
Disable On MemFull? No  
# of Print Copies? 1  
Select Std (D/W/I)? Dry  
Standard Value? 0.000  
Standard Lot #? 80  
Standard Cyl #? 123456  
Standard Expiration? 11/11/2020  
Oper No? 777777

Flow Cal. Date: 08/19/2015  
Slope 657  
Intercept -642041

IR Calibration Date: 08/19/2015  
3um 9um

0th Coef(*100):	-16255	-17993
1st Coef(*100):	263339	136695
2nd Coef(*100):	5198	1666
H2O adj(mg/l*10k):	766	522

\*\*\*\*\* Test End \*\*\*\*\*

*Deb Shanauer*

Operator Signature  
DEB SHANAUER

Remarks:

*Print Test*

Form 106-18000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006497  
Location = TOXL                8164.13.00 06/09  
02/17/2016                      09:23

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

Current Instrument Setup


Data Entry Mode:                Enabled  
Start Test Sequence:            DABACABA  
Display Prelim Rslt?            Yes  
Display Third Digit?            Yes  
Inhib Printer(Y/N)?            No  
Display Volume?                No  
Disable On Memfull?            No  
# of Print Copies?              1  
Select Std (D/W/I)?            Dry  
Standard Value?                0.080  
Standard Lot #?                 80  
Standard Cyl #?                123456  
Standard Expiration?            11/11/2020  
Oper No?                        777777

Flow Cal. Date:                08/19/2015  
Slope                            657  
Intercept                       -642041

IR Calibration Date:            08/19/2015

	3um	9um
0th Coef(*100):	-16255	-17993
1st Coef(*100):	263339	136695
2nd Coef(*100):	5198	1666
H2O adj(mg/l*10k):	766	522

\*\*\*\*\* Printer Test End \*\*\*\*\*

  
\_\_\_\_\_  
Operator Signature  
DEB SHANAVER

Remarks:

*Print test*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006497  
Location = TOXL                 8164.13.00 06/09  
02/17/2016                        09:29

Test	AC	Time
01 Room Air	0.000	09:29
02 *Subject Test	REF*	09:29
03 Room Air	0.000	09:30

\*Subject Test Refused

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 80  
Cyl No = 123456  
Expiration Date = 11/11/2020  
County = 08                          Oper No. = 777777

  
\_\_\_\_\_  
Operator Signature  
DEB SHANAVER

Remarks:  


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006497  
Location = TOXL      8164.13.00 06/09  
02/17/2016      09:34

DIAGNOSTIC

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temp Regulation Test	Pass

Oper No. = 777777

*Deb Shanaver*

---

Operator Signature  
DEB SHANAVER

Remarks:

*Diagnostic*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006497  
Location = TOXL      8164.13.00 06/09  
02/17/2016      09:41

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:41
02 Std. Sol.	0.000	09:42
03 Room Air	0.000	09:43
04 Std. Sol.	0.000	09:43
05 Room Air	0.000	09:44
06 Std. Sol.	0.000	09:45
07 Room Air	0.000	09:45

08 Sim Temp = 34.0°C

Simul Ser No = DR3375  
Std Sol No = MILLIQ  
County = 08      Oper No. = 777777

*Deb Shanaver*

Operator Signature  
DEB SHANAVER

Remarks:

*zero alcohol      0.000 AC*

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006497  
Location = TOXL      8164.13.00 06/09  
02/17/2016      13:06

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:07
02 Std. Sol.	0.019	13:07
03 Room Air	0.000	13:08
04 Std. Sol.	0.019	13:09
05 Room Air	0.000	13:10
06 Std. Sol.	0.019	13:10
07 Room Air	0.000	13:11


08 Sim Temp = 34.0°C

Simul Ser No = DR3374

Std Sol No = 14240

County = 08

Oper No. = 777777

  
\_\_\_\_\_  
Operator Signature  
DEB SHANAVER

Remarks:

*Low alcohol      0.020 AC*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006497  
Location = TOXL                    8164.13.00 06/09  
02/17/2016                            13:13

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:14
02 Std. Sol.	0.297	13:14
03 Room Air	0.000	13:15
04 Std. Sol.	0.295	13:16
05 Room Air	0.000	13:17
06 Std. Sol.	0.296	13:18
07 Room Air	0.000	13:18

08 Sim Temp = 34.0°C

Simul Ser No = DR7111  
Std Sol No = 201408C  
County = 08                            Oper No. = 777777

  
\_\_\_\_\_  
Operator Signature  
DEB SHANAVER

Remarks:

*High alcohol      0.300 AC*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006497  
Location = TOXL                8164.13.00 06/09  
02/17/2016                        13:23

Test	AC	Time
01 Room Air	0.000	13:26
02 *Subject Test	INT*	13:26
03 Room Air	0.000	13:27

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 29614080A2  
Cyl No = 25  
Expiration Date = 12/05/2016  
County = 08                            Oper No. = 777777

Deb Shanaver  
Operator Signature  
DEB SHANAVER

Remarks:

*Interferent detected            0.10 AC + 0.05% acetone*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006497  
Location = TOXL                8164.13.00 06/09  
02/17/2016                        13:28

Test	AC	Time
01 Room Air	0.000	13:29
02 *Subject Test	INT*	13:29
03 Room Air	0.000	13:30

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                Weight = 150  
Test = DUI                        Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 29614080A2  
Cyl No = 25  
Expiration Date = 12/05/2016  
County = 08                        Oper No. = 777777

*Deb Shanaver*

Operator Signature  
DEB SHANAVER

Remarks:

*Interferent detected 0.05% acetone*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006497  
Location = TOXL                8164.13.00 06/09  
02/17/2016                        13:31

Test	AC	Time
01 Room Air	0.000	13:32
02 *Subject Test	INT*	13:32
03 Room Air	0.000	13:33

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                Weight = 150  
Test = DUI                        Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 29614080A2  
Cyl No = 25  
Expiration Date = 12/05/2016  
County = 08                        Oper No. = 777777

Deb Shanaver  
Operator Signature  
DEB SHANAVER

Remarks:  
*Interferent detected 0.10% Methanol*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006497  
Location = TOXL                    8164.13.00 06/09  
02/17/2016                            13:34

Test	AC	Time
01 Room Air	0.000	13:35
02 *Subject Test	INT*	13:36
03 Room Air	0.000	13:36

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 29614080A2  
Cyl No = 25  
Expiration Date = 12/05/2016  
County = 08                            Oper No. = 777777

*Deb Shanaver*

Operator Signature  
DEB SHANAVER

Remarks:

*Interferent detected 0.10% isopropanol*

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006497  
Location = TOXL                8164.13.00 06/09  
02/17/2016                      13:52

Test	AC	Time
01 Room Air	0.000	13:52
02 Invalid Sample	X.XXX	13:53
03 Room Air	0.000	13:54

\*Invalid Test - Mouth Alcohol

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                Weight = 150  
Test = DUI                        Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 29614080A2  
Cyl No = 25  
Expiration Date = 12/05/2016  
County = 08                      Oper No. = 777777

*Deb Shanaver*

Operator Signature  
DEB SHANAVER

Remarks:

*DS Invalid sample*

Form 106-I8000





Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006497  
Location = TOXL                    8164.13.00 06/09  
02/17/2016                            14:43

Test	AC	Time
01 Room Air	0.000	14:43
02 Subject Test 1	0.000	14:44
03 Room Air	0.000	14:44
04 Reported AC	0.000	14:44

No RFI Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 29614080A2  
Cyl No = 25  
Expiration Date = 12/05/2016  
County = 08                            Oper No. = 777777

Deb Shanaver  
Operator Signature  
DEB SHANAVER

Remarks:

ABA test

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006497  
Location = TOXL                    8164.13.00 06/09  
02/17/2016                            16:10

Test	AC	Time
01 Diagnostic	OK	16:10
02 Room Air	0.000	16:11
03 Subject Test 1	0.000	16:11
04 Room Air	0.000	16:13
05 Std. Gas	0.000*	16:14
06 Room Air	0.000	16:15

\*Cal Check Out of Tolerance

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 29614080A2  
Cyl No = 25  
Expiration Date = 12/05/2016  
County = 08                            Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Shanaver  
Operator Signature  
DEB SHANAVER

Remarks:

*Calibration check out of tolerance*



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006497  
Location = TOXL                8164.13.00 06/09  
02/17/2016                        16:17

Test	AC	Time
01 Diagnostic	OK	16:18
02 Room Air	0.000	16:18
03 *Subject Test	0.000*	16:22
04 Room Air	0.000	16:24
05 Std. Gas	0.082	16:25
06 Room Air	0.000	16:26
07 *Subject Test	0.000*	16:29
08 Room Air	0.000	16:30


No RFI Detected

\*Deficient Sample - Value Printed was  
Highest Obtained

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 29614080A2  
Cyl No = 25  
Expiration Date = 12/05/2016  
County = 08                          Oper No. = 777777

20 minute waiting period ascertained? Y  
Clean Mouthpiece used and disposed? Y  
Clean Mouthpiece used and disposed? Y

I followed the Approved Method and the  
instructions displayed by the Intoxilyzer  
in conducting this test.



Operator Signature  
DEB SHANAVER

Remarks:

*Deficient samples*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006497  
Location = TOXL                    8164.13.00 06/09  
02/18/2016                            08:38

Test	AC	Time
01 Diagnostic	OK	08:39
02 Room Air	IPS*	08:39
03 Room Air	0.000	08:39

\*Improper Sample

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 29614080A2  
Cyl No = 25  
Expiration Date = 12/05/2016  
County = 08                            Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

*Deb Shanaver*

Operator Signature  
DEB SHANAVER

Remarks:

*Improper sample*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006497  
Location = TOXL                    8164.13.00 06/09  
02/18/2016                            08:39

Test	AC	Time
01 Diagnostic	OK	08:40
02 Room Air	0.000	08:41
03 Subject Test 1	0.038*	08:41
04 Room Air	0.000	08:43
05 Std. Gas	0.080	08:45
06 Room Air	0.000	08:46
07 Subject Test 2	0.000*	08:46
08 Room Air	0.000	08:47

\*Difference Too Great

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 29614080A2  
Cyl No = 25  
Expiration Date = 12/05/2016  
County = 08                          Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

*Deb Shanaver*

Operator Signature  
DEB SHANAVER

Remarks:

*Difference too great*

Form 106-I8000





Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006497  
Location = TOXL                8164.13.00 06/09  
02/17/2016                      14:44

Test	AC	Time
01 Room Air	0.000	14:46
02 Std. Gas	0.081	14:46
03 Room Air	0.000	14:47
04 Std. Gas	0.081	14:47
05 Room Air	0.000	14:48
06 Std. Gas	0.082	14:48
07 Room Air	0.000	14:48
08 Std. Gas	0.081	14:49
09 Room Air	0.000	14:49
10 Std. Gas	0.082	14:50
11 Room Air	0.000	14:50
12 Std. Gas	0.081	14:51
13 Room Air	0.000	14:51
14 Std. Gas	0.081	14:52
15 Room Air	0.000	14:52
16 Std. Gas	0.082	14:53
17 Room Air	0.000	14:53
18 Std. Gas	0.081	14:53
19 Room Air	0.000	14:54
20 Std. Gas	0.081	14:54
21 Room Air	0.000	14:55
22 Std. Gas	0.081	14:55
23 Room Air	0.000	14:56
24 Std. Gas	0.081	14:56
25 Room Air	0.000	14:57
26 Std. Gas	0.081	14:57
27 Room Air	0.000	14:58
28 Std. Gas	0.082	14:58
29 Room Air	0.000	14:59
30 Std. Gas	0.082	14:59
31 Room Air	0.000	15:00
32 Std. Gas	0.082	15:00
33 Room Air	0.000	15:01
34 Std. Gas	0.082	15:01
35 Room Air	0.000	15:01
36 Std. Gas	0.082	15:02
37 Room Air	0.000	15:02
38 Std. Gas	0.081	15:03
39 Room Air	0.000	15:03
40 Std. Gas	0.082	15:04
41 Room Air	0.000	15:04
42 Cal Check Stats:		
43 Average	0.0815	
44 Std Dev	0.0005	
45 Rel Std Dev	0.63%	

*Stability check  
Deb Shanaver*

Operator Signature  
DEB SHANAVER

Remarks:

lot #: 29614080A2

cyl #: 25

exp. date: <sup>DS</sup>~~12/00~~  
12/05/2016

Form 106-I8000

Stability check  
Deb Shanaver