



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Phil Wolf</u>		
Location: <u>Cavalier, ND</u>		
Intoxilyzer® Serial Number: <u>80-002669</u>		
Gas Lot Number: <u>659357</u>	Gas Cylinder Number: <u>9</u>	Expiration Date: <u>4/3/11</u>
<input type="checkbox"/> New Cylinder Installed	<input type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>077</u> %		
Test 2: 0. <u>078</u> %		
Test 3: 0. <u>077</u> %		
(ATTACH TEST RECORD)		

P. Wolf
Chemical Test Operator Signature

2/20/11
Date

Deb Keshur
Reviewed By

01 Mar 2011
Date

80-00 2669
P. 105/105
VCM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = CAVA 8164.13.00 06/09
02/20/2011 02:27

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	02:28
02 Std. Gas	0.077	02:28
03 Room Air	0.000	02:29
04 Std. Gas	0.078	02:29
05 Room Air	0.000	02:30
06 Std. Gas	0.077	02:30
07 Room Air	0.000	02:30

Lot No = 659357
Cyl No = 9
Exp Date = 09/03/2011
County = 34 Oper No. = 120547



Operator Signature
PHIL WOLF

Remarks:

Form 106-I8000

80-002669
P. 105 of 105
JSM